


Data Path : Z:\HPCHEM1\BNA_F\Data\BF030316\
 Data File : BF085154.D
 Acq On : 3 Mar 2016 13:46
 Operator : SJ/IZ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF030316.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Mar 04 00:54:58 2016



AutoFind: Scans 827, 828, 829; Background Corrected with Scan 824

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	43.3	24400	PASS
68	69	0.00	2	1.4	335	PASS
69	198	0.00	100	41.1	23158	PASS
70	69	0.00	2	0.5	109	PASS
127	198	10	80	54.4	30680	PASS
197	198	0.00	2	0.6	332	PASS
198	198	100	100	100.0	56363	PASS
199	198	5	9	6.8	3844	PASS
275	198	10	60	22.1	12483	PASS
365	198	1	100	2.7	1520	PASS
441	442	0.01	24	15.5	6137	PASS
442	198	50	100	70.4	39684	PASS
443	442	15	24	19.6	7771	PASS