

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF030317\
 Data File : BF093374.D
 Acq On : 3 Mar 2017 22:24
 Operator : SJ/MA
 Sample : I2086-06
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SRI-MW-5-20170302

Integration Parameters: rteint.p
 Integrator: RTE
 Smoothing : ON
 Sampling : 1
 Start Thrs: 0.2
 Stop Thrs : 0
 Filtering: 5
 Min Area: 3 % of largest Peak
 Max Peaks: 100
 Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >
 Peak separation: 5

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF013017.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

Signal : TIC

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	4.933	247	249	256	rBV	206226	261718	4.54%	0.667%
2	5.379	286	288	305	rBV	1647338	2478357	42.97%	6.316%
3	6.430	378	380	388	rBV	1601053	1671755	28.99%	4.260%
4	6.556	388	391	393	rBV	3654844	4144720	71.86%	10.562%
5	6.784	409	411	413	rBV	1227205	1262955	21.90%	3.218%
6	6.933	422	424	427	rBV	2558838	3492563	60.56%	8.900%
7	7.356	458	461	464	rBV	2820581	3083337	53.46%	7.858%
8	8.076	521	524	526	rBV	1528443	1729066	29.98%	4.406%
9	9.150	615	618	620	rBV	5230346	5767570	100.00%	14.698%
10	9.539	650	652	654	rBV	134370	154432	2.68%	0.394%
11	9.756	668	671	674	rBV	49985	100740	1.75%	0.257%
12	9.825	674	677	679	rVV	1551115	1896102	32.88%	4.832%
13	10.248	713	714	716	rVB	113841	85401	1.48%	0.218%
14	10.613	743	746	749	rBV	2802284	3110815	53.94%	7.928%
15	10.716	754	755	760	rVB2	121488	171885	2.98%	0.438%
16	11.162	792	794	799	rBV2	83455	133071	2.31%	0.339%
17	11.299	804	806	809	rBV	1514607	1833286	31.79%	4.672%
18	12.888	942	945	948	rBV	3952519	5049017	87.54%	12.867%
19	13.779	1021	1023	1028	rBV	101832	130867	2.27%	0.333%
20	13.939	1035	1037	1040	rBV	1520387	1516262	26.29%	3.864%
21	15.357	1158	1161	1165	rBV	939151	1166683	20.23%	2.973%


Sum of corrected areas: 39240602

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF030317\
Data File : BF093374.D
Acq On : 3 Mar 2017 22:24
Operator : SJ/MA
Sample : I2086-06
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SRI-MW-5-20170302

Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF013017.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

TIC Library : C:\DATABASE\NIST02.L
TIC Integration Parameters: LSCINT.P



Data Path : Z:\HPCHEM1\BNA_F\DATA\BF030317\
Data File : BF093374.D
Acq On : 3 Mar 2017 22:24
Operator : SJ/MA
Sample : I2086-06
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SRI-MW-5-20170302










Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF013017.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

TIC Library : C:\DATABASE\NIST02.L
TIC Integration Parameters: LSCINT.P

Peak Number 1 2-Pentanone, 4-hydroxy-4-me... Concentration Rank 3

R.T.	EstConc	Area	Relative to ISTD	R.T.
4.93	4.14 ng	261718	1,4-Dichlorobenzene-d4	6.78

Hit#	of	5	Tentative ID	MW	MolForm	CAS#	Qual
1			2-Pentanone, 4-hydroxy-4-methyl-	116	C6H12O2	000123-42-2	50
2			2-Hexanol, 2-methyl-	116	C7H16O	000625-23-0	40
3			Propane, 2-methyl-2-(1-methylethoxy)-	116	C7H16O	017348-59-3	33
4			3-Octanol	130	C8H18O	000589-98-0	33
5			3-Hexanol, 4-methyl-	116	C7H16O	000615-29-2	33



Data Path : Z:\HPCHEM1\BNA_F\DATA\BF030317\
Data File : BF093374.D
Acq On : 3 Mar 2017 22:24
Operator : SJ/MA
Sample : I2086-06
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SRI-MW-5-20170302










Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF013017.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

TIC Library : C:\DATABASE\NIST02.L
TIC Integration Parameters: LSCINT.P

Peak Number 2 unknown6.56 Concentration Rank 1

R.T.	EstConc	Area	Relative to ISTD	R.T.
6.56	65.64 ng	4144720	1,4-Dichlorobenzene-d4	6.78

Hit#	of	Tentative ID	MW	MolForm	CAS#	Qual
1	5	3-Chloro-6-fluoro-pyrazine	132	C4H2ClFN2	1000146-10-7	27
2		Tranylcypromine-propionyl	189	C12H15NO	1000123-86-3	12
3		(E)-3-Chloro-2-methyl-2-pentenal	132	C6H9ClO	031357-76-3	12
4		5-Aminoindole	132	C8H8N2	005192-03-0	12
5		Benzene, 1,3,5-trifluoro-	132	C6H3F3	000372-38-3	9



Data Path : Z:\HPCHEM1\BNA_F\DATA\BF030317\
Data File : BF093374.D
Acq On : 3 Mar 2017 22:24
Operator : SJ/MA
Sample : I2086-06
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SRI-MW-5-20170302

Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF013017.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

TIC Library : C:\DATABASE\NIST02.L
TIC Integration Parameters: LSCINT.P

TIC Top Hit name	RT	EstConc	Units	Response	---Internal Standard---			
					#	RT	Resp	Conc
2-Pentanone, 4-hy...	4.93	4.1	ng	261718	1	6.78	1262960	20.0
unknown6.56	6.56	65.6	ng	4144720	1	6.78	1262960	20.0