

DDT Breakdown

Date	Instrument Name	DFTPP Data File
3/3/2021	BNA_F	<u>BF123352.D</u>
Compound Name	Response	Retention Time
DDT	479978	13.54
DDD	3935	13.24
DDE	969	12.9
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
4904	484882	1.01