




Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF030321\
Data File : BF123352.D
Acq On : 3 Mar 2021 12:07
Operator : JU/CG
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Mar 03 12:45:25 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF021921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 19 16:08:53 2021
Response via : Initial Calibration



TIC: BF123352.D\data.ms

(70) Pentachlorophenol (C)




11.186min (-0.012) 25538.44 ng

response 168153

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.10	63.05
264.00	62.30	60.97
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF030321\
Data File : BF123352.D
Acq On : 3 Mar 2021 12:07
Operator : JU/CG
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Mar 03 12:45:25 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF021921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 19 16:08:53 2021
Response via : Initial Calibration



TIC: BF123352.D\data.ms

(77) Benzidine

12.721min (-0.012) 0.00 ng

response 1150518

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.70	14.55
183.00	12.10	12.22
0.00	0.00	0.00