



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF030321\
 Data File : BF123352.D
 Acq On : 3 Mar 2021 12:07
 Operator : JU/CG
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF021921.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Mar 03 13:18:24 2021






AutoFind: Scans 1622, 1623, 1624; Background Corrected with Scan 1616

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	43.2	88021	PASS
68	69	0.00	2	1.8	1579	PASS
69	198	0.00	100	44.0	89725	PASS
70	69	0.00	2	0.5	488	PASS
127	198	10	80	52.0	105907	PASS
197	198	0.00	2	0.2	343	PASS
198	198	100	100	100.0	203733	PASS
199	198	5	9	6.7	13553	PASS
275	198	10	60	26.5	53891	PASS
365	198	1	100	2.9	5878	PASS
441	198	0.01	100	13.5	27415	PASS
442	442	50	100	100.0	174939	PASS
443	442	15	24	18.9	33000	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF030321\
Data File : BF123352.D
Acq On : 3 Mar 2021 12:07
Operator : JU/CG
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Mar 03 12:45:25 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF021921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 19 16:08:53 2021
Response via : Initial Calibration



TIC: BF123352.D\data.ms

(70) Pentachlorophenol (C)

11.186min (-0.012) 25538.44 ng




response 168153

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.10	63.05
264.00	62.30	60.97
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF030321\
Data File : BF123352.D
Acq On : 3 Mar 2021 12:07
Operator : JU/CG
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Mar 03 12:45:25 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF021921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 19 16:08:53 2021
Response via : Initial Calibration



TIC: BF123352.D\data.ms

(77) Benzidine

12.721min (-0.012) 0.00 ng

response 1150518

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.70	14.55
183.00	12.10	12.22
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
3/3/2021	BNA_F	<u>BF123352.D</u>
Compound Name	Response	Retention Time
DDT	479978	13.54
DDD	3935	13.24
DDE	969	12.9
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
4904	484882	1.01

Instrument :
BNA_F
ClientSampleId :
DFTPP