

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF030322\
 Data File : BF127188.D
 Acq On : 03 Mar 2022 23:16
 Operator : CG\JU
 Sample : N1689-11DL 2X
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DUP022522DL

Quant Time: Mar 04 02:15:36 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF021622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 01 23:57:09 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Christian Giraldo 03/04/2022
 Supervised By :Jagrut Upadhyay 03/04/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.892	152	99033	20.000	ng	0.00
21) Naphthalene-d8	8.175	136	366906	20.000	ng	# 0.00
39) Acenaphthene-d10	9.933	164	165068	20.000	ng	0.00
64) Phenanthrene-d10	11.422	188	240948	20.000	ng	# 0.00
76) Chrysene-d12	14.068	240	171865	20.000	ng	# 0.00
86) Perylene-d12	15.562	264	144053	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.516	112	155497	24.268	ng	0.00
7) Phenol-d6	6.522	99	146456	16.939	ng	0.00
23) Nitrobenzene-d5	7.457	82	301783	37.126	ng	0.00
42) 2,4,6-Tribromophenol	10.722	330	105953	55.749	ng	0.00
45) 2-Fluorobiphenyl	9.251	172	483441	43.116	ng	0.00
79) Terphenyl-d14	13.010	244	391334	33.473	ng	0.00
Target Compounds						
37) 2-Methylnaphthalene	8.886	142	435917	35.075	ng	98
38) 1-Methylnaphthalene	8.986	142	327303	27.502	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF030322\
Data File : BF127188.D
Acq On : 03 Mar 2022 23:16
Operator : CG\JU
Sample : N1689-11DL 2X
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DUP022522DL

Manual Integrations
APPROVED

Reviewed By : Christian Giraldo 03/04/2022
Supervised By : Jagrut Upadhyay 03/04/2022

Quant Time: Mar 04 02:15:36 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF021622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 01 23:57:09 2022
Response via : Initial Calibration

