

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF030322\
 Data File : BF127162.D
 Acq On : 03 Mar 2022 10:17
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 03 16:57:02 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF021622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 01 23:57:09 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	74	0.00
2	1,4-Dioxane	0.592	0.530	10.5	68	-0.02
3	Pyridine	1.553	1.546	0.5	74	-0.01
4	n-Nitrosodimethylamine	0.761	0.762	-0.1	73	-0.01
5 S	2-Fluorophenol	1.294	1.320	-2.0	76	0.00
6	Aniline	2.069	2.002	3.2	70	0.00
7 S	Phenol-d6	1.746	1.746	0.0	74	0.00
8	2-Chlorophenol	1.388	1.429	-3.0	76	0.00
9	Benzaldehyde	1.063	0.922	13.3	72	0.00
10 C	Phenol	1.764	1.888	-7.0	81	0.00
11	bis(2-Chloroethyl)ether	1.384	1.314	5.1	71	0.00
12	1,3-Dichlorobenzene	1.498	1.539	-2.7	78	0.00
13 C	1,4-Dichlorobenzene	1.518	1.572	-3.6	78	0.00
14	1,2-Dichlorobenzene	1.448	1.482	-2.3	77	0.00
15	Benzyl Alcohol	1.471	1.357	7.7	67	0.00
16	2,2'-oxybis(1-Chloropropane	2.179	1.780	18.3	60	0.00
17	2-Methylphenol	1.201	1.227	-2.2	75	0.00
18	Hexachloroethane	0.582	0.585	-0.5	75	0.00
19 P	n-Nitroso-di-n-propylamine	1.125	1.045	7.1	69	0.00
20	3+4-Methylphenols	1.560	1.542	1.2	72	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	71	0.00
22	Acetophenone	0.527	0.519	1.5	73	0.00
23 S	Nitrobenzene-d5	0.443	0.439	0.9	73	0.00
24	Nitrobenzene	0.419	0.405	3.3	71	0.00
25	Isophorone	0.733	0.680	7.2	68	0.00
26 C	2-Nitrophenol	0.178	0.187	-5.1	75	0.00
27	2,4-Dimethylphenol	0.293	0.265	9.6	67	0.00
28	bis(2-Chloroethoxy)methane	0.445	0.416	6.5	69	0.00
29 C	2,4-Dichlorophenol	0.275	0.282	-2.5	75	0.00
30	1,2,4-Trichlorobenzene	0.308	0.315	-2.3	76	0.00
31	Naphthalene	1.044	1.025	1.8	73	0.00
32	Benzoic acid	0.208	0.175	15.9	57	0.00
33	4-Chloroaniline	0.411	0.411	0.0	73	0.00
34 C	Hexachlorobutadiene	0.180	0.189	-5.0	77	0.00
35	Caprolactam	0.096	0.095	1.0	71	0.00
36 C	4-Chloro-3-methylphenol	0.352	0.353	-0.3	72	0.00
37	2-Methylnaphthalene	0.677	0.663	2.1	72	0.00
38	1-Methylnaphthalene	0.649	0.633	2.5	72	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	71	0.00
40	1,2,4,5-Tetrachlorobenzene	0.526	0.534	-1.5	77	0.00
41 P	Hexachlorocyclopentadiene	0.286	0.229	19.9	57	0.00
42 S	2,4,6-Tribromophenol	0.230	0.257	-11.7	81	0.00
43 C	2,4,6-Trichlorophenol	0.386	0.378	2.1	72	0.00
44	2,4,5-Trichlorophenol	0.406	0.392	3.4	71	0.00
45 S	2-Fluorobiphenyl	1.359	1.327	2.4	75	0.00
46	1,1'-Biphenyl	1.585	1.486	6.2	71	0.00
47	2-Chloronaphthalene	1.175	1.125	4.3	71	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF030322\
 Data File : BF127162.D
 Acq On : 03 Mar 2022 10:17
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 03 16:57:02 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF021622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 01 23:57:09 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.464	0.443	4.5	71	0.00
49	Acenaphthylene	1.909	1.853	2.9	72	0.00
50	Dimethylphthalate	1.430	1.403	1.9	73	0.00
51	2,6-Dinitrotoluene	0.291	0.298	-2.4	75	0.00
52 C	Acenaphthene	1.241	1.232	0.7	73	0.00
53	3-Nitroaniline	0.356	0.353	0.8	72	0.00
54 P	2,4-Dinitrophenol	0.154	0.165	-7.1	73	0.00
55	Dibenzofuran	1.707	1.704	0.2	75	0.00
56 P	4-Nitrophenol	0.263	0.254	3.4	69	0.00
57	2,4-Dinitrotoluene	0.393	0.405	-3.1	75	0.00
58	Fluorene	1.330	1.328	0.2	75	0.00
59	2,3,4,6-Tetrachlorophenol	0.322	0.330	-2.5	75	0.00
60	Diethylphthalate	1.494	1.455	2.6	72	0.00
61	4-Chlorophenyl-phenylether	0.627	0.642	-2.4	76	0.00
62	4-Nitroaniline	0.351	0.365	-4.0	75	0.00
63	Azobenzene	1.663	1.525	8.3	68	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	74	0.00
65	4,6-Dinitro-2-methylphenol	0.118	0.127	-7.6	76	0.00
66 c	n-Nitrosodiphenylamine	0.656	0.632	3.7	74	0.00
67	4-Bromophenyl-phenylether	0.223	0.226	-1.3	78	0.00
68	Hexachlorobenzene	0.241	0.248	-2.9	80	0.00
69	Atrazine	0.203	0.203	0.0	78	0.00
70 C	Pentachlorophenol	0.131	0.120	8.4	65	0.00
71	Phenanthrene	1.119	1.107	1.1	78	0.00
72	Anthracene	1.114	1.109	0.4	77	0.00
73	Carbazole	1.022	1.037	-1.5	79	0.00
74	Di-n-butylphthalate	1.314	1.244	5.3	73	0.00
75 C	Fluoranthene	1.063	1.090	-2.5	80	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	82	0.00
77	Benzydine	0.544	0.514	5.5	75	0.00
78	Pyrene	1.738	1.666	4.1	78	0.00
79 S	Terphenyl-d14	1.361	1.360	0.1	84	0.00
80	Butylbenzylphthalate	0.823	0.781	5.1	77	0.00
81	Benzo(a)anthracene	1.348	1.324	1.8	82	0.00
82	3,3'-Dichlorobenzidine	0.425	0.436	-2.6	84	0.00
83	Chrysene	1.268	1.341	-5.8	88	0.00
84	Bis(2-ethylhexyl)phthalate	1.080	0.953	11.8	73	0.00
85 c	Di-n-octyl phthalate	1.555	1.460	6.1	79	0.00
86 I	Perylene-d12	1.000	1.000	0.0	84	0.00
87	Indeno(1,2,3-cd)pyrene	1.311	1.262	3.7	78	0.00
88	Benzo(b)fluoranthene	1.343	1.348	-0.4	87	0.00
89	Benzo(k)fluoranthene	1.296	1.338	-3.2	87	0.00
90 C	Benzo(a)pyrene	1.049	1.054	-0.5	85	0.00
91	Dibenzo(a,h)anthracene	1.085	1.015	6.5	76	0.00
92	Benzo(g,h,i)perylene	1.069	1.010	5.5	76	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF030322\
Data File : BF127162.D
Acq On : 03 Mar 2022 10:17
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Mar 03 16:57:02 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF021622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 01 23:57:09 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0