

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF030322\
 Data File : BF127171.D
 Acq On : 03 Mar 2022 15:05
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 03 17:01:46 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF021622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 01 23:57:09 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	89	0.00
2	1,4-Dioxane	40.000	34.448	13.9	79	-0.02
3	Pyridine	40.000	40.648	-1.6	91	-0.02
4	n-Nitrosodimethylamine	40.000	37.387	6.5	82	-0.02
5 S	2-Fluorophenol	80.000	81.058	-1.3	92	0.00
6	Aniline	40.000	39.471	1.3	87	0.00
7 S	Phenol-d6	80.000	80.599	-0.7	91	0.00
8	2-Chlorophenol	40.000	40.888	-2.2	92	0.00
9	Benzaldehyde	40.000	35.714	10.7	89	0.00
10 C	Phenol	40.000	43.414	-8.5	99	0.00
11	bis(2-Chloroethyl)ether	40.000	39.350	1.6	89	0.00
12	1,3-Dichlorobenzene	40.000	41.037	-2.6	94	0.00
13 C	1,4-Dichlorobenzene	40.000	40.577	-1.4	93	0.00
14	1,2-Dichlorobenzene	40.000	39.972	0.1	91	0.00
15	Benzyl Alcohol	40.000	35.592	11.0	79	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	33.980	15.1	76	0.00
17	2-Methylphenol	40.000	40.636	-1.6	91	0.00
18	Hexachloroethane	40.000	39.844	0.4	91	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	36.460	8.8	82	0.00
20	3+4-Methylphenols	40.000	39.203	2.0	87	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	87	0.00
22	Acetophenone	40.000	37.357	6.6	85	0.00
23 S	Nitrobenzene-d5	80.000	76.293	4.6	86	0.00
24	Nitrobenzene	40.000	38.561	3.6	87	0.00
25	Isophorone	40.000	37.247	6.9	84	0.00
26 C	2-Nitrophenol	40.000	41.842	-4.6	92	0.00
27	2,4-Dimethylphenol	40.000	35.739	10.7	81	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.530	3.7	87	0.00
29 C	2,4-Dichlorophenol	40.000	40.596	-1.5	91	0.00
30	1,2,4-Trichlorobenzene	40.000	39.642	0.9	90	0.00
31	Naphthalene	40.000	39.089	2.3	90	0.00
32	Benzoic acid	40.000	36.087	9.8	75	0.00
33	4-Chloroaniline	40.000	39.885	0.3	90	0.00
34 C	Hexachlorobutadiene	40.000	39.007	2.5	88	0.00
35	Caprolactam	40.000	40.854	-2.1	90	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.514	1.2	87	0.00
37	2-Methylnaphthalene	40.000	38.768	3.1	88	0.00
38	1-Methylnaphthalene	40.000	39.529	1.2	90	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	86	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.919	-2.3	94	0.00
41 P	Hexachlorocyclopentadiene	40.000	35.640	10.9	76	0.00
42 S	2,4,6-Tribromophenol	80.000	84.243	-5.3	92	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.089	-0.2	90	0.00
44	2,4,5-Trichlorophenol	40.000	40.519	-1.3	90	0.00
45 S	2-Fluorobiphenyl	80.000	78.054	2.4	90	0.00
46	1,1'-Biphenyl	40.000	39.093	2.3	89	0.00
47	2-Chloronaphthalene	40.000	39.371	1.6	89	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF030322\
 Data File : BF127171.D
 Acq On : 03 Mar 2022 15:05
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 03 17:01:46 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF021622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 01 23:57:09 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.346	1.6	88	0.00
49	Acenaphthylene	40.000	39.389	1.5	89	0.00
50	Dimethylphthalate	40.000	39.312	1.7	89	0.00
51	2,6-Dinitrotoluene	40.000	41.555	-3.9	92	0.00
52 C	Acenaphthene	40.000	39.372	1.6	88	0.00
53	3-Nitroaniline	40.000	41.064	-2.7	90	0.00
54 P	2,4-Dinitrophenol	40.000	39.766	0.6	82	0.00
55	Dibenzofuran	40.000	39.496	1.3	89	0.00
56 P	4-Nitrophenol	40.000	38.556	3.6	83	0.00
57	2,4-Dinitrotoluene	40.000	41.694	-4.2	91	0.00
58	Fluorene	40.000	39.577	1.1	90	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.454	-1.1	90	0.00
60	Diethylphthalate	40.000	38.209	4.5	86	0.00
61	4-Chlorophenyl-phenylether	40.000	39.319	1.7	89	0.00
62	4-Nitroaniline	40.000	42.431	-6.1	92	0.00
63	Azobenzene	40.000	37.663	5.8	85	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	90	0.00
65	4,6-Dinitro-2-methylphenol	40.000	42.692	-6.7	91	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.286	4.3	89	0.00
67	4-Bromophenyl-phenylether	40.000	38.993	2.5	91	0.00
68	Hexachlorobenzene	40.000	39.768	0.6	94	0.00
69	Atrazine	40.000	39.031	2.4	92	0.00
70 C	Pentachlorophenol	40.000	34.712	13.2	75	0.00
71	Phenanthrene	40.000	38.916	2.7	93	0.00
72	Anthracene	40.000	38.806	3.0	91	0.00
73	Carbazole	40.000	39.634	0.9	93	0.00
74	Di-n-butylphthalate	40.000	37.042	7.4	86	0.00
75 C	Fluoranthene	40.000	40.652	-1.6	96	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	103	0.00
77	Benzydine	40.000	35.086	12.3	88	0.00
78	Pyrene	40.000	37.769	5.6	97	0.00
79 S	Terphenyl-d14	80.000	74.076	7.4	97	0.00
80	Butylbenzylphthalate	40.000	36.559	8.6	93	0.00
81	Benzo(a)anthracene	40.000	39.325	1.7	103	0.00
82	3,3'-Dichlorobenzidine	40.000	40.680	-1.7	104	0.00
83	Chrysene	40.000	39.154	2.1	103	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	34.755	13.1	90	0.00
85 c	Di-n-octyl phthalate	40.000	34.455	13.9	91	0.00
86 I	Perylene-d12	20.000	20.000	0.0	99	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.113	-0.3	95	-0.01
88	Benzo(b)fluoranthene	40.000	38.996	2.5	99	0.00
89	Benzo(k)fluoranthene	40.000	41.993	-5.0	104	0.00
90 C	Benzo(a)pyrene	40.000	40.318	-0.8	100	0.00
91	Dibenzo(a,h)anthracene	40.000	39.624	0.9	94	-0.01
92	Benzo(g,h,i)perylene	40.000	40.971	-2.4	96	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF030322\
Data File : BF127171.D
Acq On : 03 Mar 2022 15:05
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Mar 03 17:01:46 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF021622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 01 23:57:09 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0