

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF030415\
 Data File : BF077588.D
 Acq On : 4 Mar 2015 18:13
 Operator : TP/IZ
 Sample : G1412-15
 Misc : W/TCLP
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 WC030215

Quant Time: Mar 05 00:01:41 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF021915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 04 21:51:23 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.22	152	62027	20.00	ng	0.00
21) Naphthalene-d8	8.80	136	254618	20.00	ng	0.00
38) Acenaphthene-d10	10.97	164	123725	20.00	ng	0.00
63) Phenanthrene-d10	12.81	188	241450	20.00	ng	0.00
75) Chrysene-d12	16.10	240	251788	20.00	ng	-0.01
86) Perylene-d12	17.83	264	245479	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.51	112	460782	124.02	ng	0.00
7) Phenol-d6	6.78	99	490160	102.61	ng	0.00
23) Nitrobenzene-d5	7.91	82	408520	99.77	ng	0.00
41) 2,4,6-Tribromophenol	11.95	330	173198	145.38	ng	0.00
44) 2-Fluorobiphenyl	10.15	172	796923	100.43	ng	0.00
78) Terphenyl-d14	14.81	244	956555	88.81	ng	0.00
Target Compounds						
31) Naphthalene	8.82	128	199647	15.68	ng	99
37) 2-Methylnaphthalene	9.68	142	85279	9.43	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF030415\
Data File : BF077588.D
Acq On : 4 Mar 2015 18:13
Operator : TP/IZ
Sample : G1412-15
Misc : W/TCLP
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
WC030215

Quant Time: Mar 05 00:01:41 2015
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF021915.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 04 21:51:23 2015
Response via : Initial Calibration

