

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF030415\
 Data File : BF077590.D
 Acq On : 4 Mar 2015 19:10
 Operator : TP/IZ
 Sample : G1426-04
 Misc : W
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DCW-WW-067-3315

Manual Integrations
 APPROVED

mohammad
 3/5/2015 12:47:51 PM

Quant Time: Mar 05 11:07:10 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF021915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 04 21:51:23 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.22	152	60526	20.00	ng	0.00
21) Naphthalene-d8	8.80	136	243840	20.00	ng	0.00
38) Acenaphthene-d10	10.97	164	122953	20.00	ng	0.00
63) Phenanthrene-d10	12.81	188	228940	20.00	ng	0.00
75) Chrysene-d12	16.10	240	253577	20.00	ng	0.00
86) Perylene-d12	17.86	264	249839	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.51	112	234582	64.70	ng	0.00
7) Phenol-d6	6.78	99	189643	40.68	ng	0.00
23) Nitrobenzene-d5	7.91	82	408702	104.23	ng	0.00
41) 2,4,6-Tribromophenol	11.95	330	166340	140.50	ng	0.00
44) 2-Fluorobiphenyl	10.15	172	802551	101.78	ng	0.00
78) Terphenyl-d14	14.81	244	892612	82.29	ng	0.00
Target Compounds						Qvalue
3) Pyridine	3.01	79	13829m	3.14	ng	
6) Aniline	6.81	93	21860	3.39	ng	90
10) Phenol	6.80	94	11334	2.05	ng	75
13) 1,4-Dichlorobenzene	7.24	146	43991	9.29	ng	98
14) 1,2-Dichlorobenzene	7.43	146	811457	180.05	ng	95
30) 1,2,4-Trichlorobenzene	8.73	180	362729	92.91	ng	97
31) Naphthalene	8.82	128	249450	20.46	ng	99
33) 4-Chloroaniline	8.88	127	335518	61.88	ng	# 92
37) 2-Methylnaphthalene	9.68	142	199661	23.06	ng	98
45) 1,1'-Biphenyl	10.27	154	24744	2.66	ng	98
49) Dimethylphthalate	10.65	163	22729	2.63	ng	98
51) Acenaphthene	11.01	154	69725	10.10	ng	99
54) Dibenzofuran	11.23	168	32513	3.05	ng	97
56) 2,4-Dinitrotoluene	11.22	165	61153	26.11	ng	90
57) Fluorene	11.65	166	74056	8.69	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF030415\
Data File : BF077590.D
Acq On : 4 Mar 2015 19:10
Operator : TP/IZ
Sample : G1426-04
Misc : W
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DCW-WW-067-3315

Manual Integrations
APPROVED

mohammad
3/5/2015 12:47:51 PM

Quant Time: Mar 05 11:07:10 2015
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF021915.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 04 21:51:23 2015
Response via : Initial Calibration

