

Data Path : Z:\HPCHEM1\BNA F\Data\BF030415\
 Data File : BF077580.D
 Acq On : 4 Mar 2015 13:24
 Operator : TP/IZ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 04 21:57:39 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF021915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 04 13:27:09 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
2	1,4-Dioxane	0.508	0.515	-1.4	97	0.00
3	Pyridine	1.454	1.550	-6.6	95	0.00
4	n-Nitrosodimethylamine	0.632	0.681	-7.8	105	0.00
5 S	2-Fluorophenol	1.198	1.250	-4.3	97	0.00
6	Aniline	2.129	2.059	3.3	88	0.00
7 S	Phenol-d6	1.540	1.512	1.8	90	0.00
8	2-Chlorophenol	1.413	1.400	0.9	92	0.00
9	Benzaldehyde	0.935	1.020	-9.1	105	0.00
10 C	Phenol	1.828	1.829	-0.1	89	0.00
11	bis(2-Chloroethyl)ether	1.313	1.191	9.3	86	0.00
12	1,3-Dichlorobenzene	1.539	1.471	4.4	88	0.00
13 C	1,4-Dichlorobenzene	1.566	1.482	5.4	87	0.00
14	1,2-Dichlorobenzene	1.489	1.390	6.6	85	0.00
15	Benzyl Alcohol	1.171	1.074	8.3	85	0.00
16	2,2'-oxybis(1-Chloropropane	1.877	1.672	10.9	81	0.00
17	2-Methylphenol	1.105	1.058	4.3	84	0.00
18	Hexachloroethane	0.523	0.505	3.4	89	0.00
19 P	n-Nitroso-di-n-propylamine	0.978	0.949	3.0	87	0.00
20	3+4-Methylphenols	1.455	1.412	3.0	88	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	89	0.00
22	Acetophenone	0.470	0.486	-3.4	94	0.00
23 S	Nitrobenzene-d5	0.322	0.331	-2.8	88	0.00
24	Nitrobenzene	0.338	0.348	-3.0	90	0.00
25	Isophorone	0.638	0.596	6.6	78	0.00
26 C	2-Nitrophenol	0.139	0.160	-15.1	92	0.00
27	2,4-Dimethylphenol	0.310	0.306	1.3	86	0.00
28	bis(2-Chloroethoxy)methane	0.403	0.396	1.7	85	0.00
29 C	2,4-Dichlorophenol	0.291	0.289	0.7	85	0.00
30	1,2,4-Trichlorobenzene	0.320	0.297	7.2	81	0.00
31	Naphthalene	1.000	0.982	1.8	87	0.00
32	Benzoic acid	0.151	0.180	-19.2	97	0.00
33	4-Chloroaniline	0.445	0.423	4.9	81	0.00
34 C	Hexachlorobutadiene	0.183	0.161	12.0	77	0.00
35	Caprolactam	0.089	0.092	-3.4	87	0.00
36 C	4-Chloro-3-methylphenol	0.293	0.294	-0.3	84	0.00
37	2-Methylnaphthalene	0.710	0.681	4.1	81	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	77	0.00
39	1,2,4,5-Tetrachlorobenzene	0.574	0.604	-5.2	78	0.00
40 P	Hexachlorocyclopentadiene	0.308	0.273	11.4	61	0.00
41 S	2,4,6-Tribromophenol	0.193	0.199	-3.1	74	0.00
42 C	2,4,6-Trichlorophenol	0.380	0.405	-6.6	75	0.00
43	2,4,5-Trichlorophenol	0.406	0.451	-11.1	80	0.00
44 S	2-Fluorobiphenyl	1.283	1.288	-0.4	76	0.00

Data Path : Z:\HPCHEM1\BNA F\Data\BF030415\
 Data File : BF077580.D
 Acq On : 4 Mar 2015 13:24
 Operator : TP/IZ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 04 21:57:39 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF021915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 04 13:27:09 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.515	1.634	-7.9	80	0.00
46	2-Chloronaphthalene	1.188	1.280	-7.7	82	0.00
47	2-Nitroaniline	0.293	0.374	-27.6#	93	0.00
48	Acenaphthylene	1.881	1.965	-4.5	79	0.00
49	Dimethylphthalate	1.406	1.430	-1.7	76	0.00
50	2,6-Dinitrotoluene	0.282	0.314	-11.3	76	0.00
51 C	Acenaphthene	1.123	1.162	-3.5	76	0.00
52	3-Nitroaniline	0.325	0.410	-26.2#	89	0.00
53 P	2,4-Dinitrophenol	0.086	0.109	-26.7#	92	0.00
54	Dibenzofuran	1.733	1.823	-5.2	78	0.00
55 P	4-Nitrophenol	0.261	0.330	-26.4#	87	0.00
56	2,4-Dinitrotoluene	0.381	0.410	-7.6	73	0.00
57	Fluorene	1.386	1.442	-4.0	77	0.00
58	2,3,4,6-Tetrachlorophenol	0.327	0.351	-7.3	75	0.00
59	Diethylphthalate	1.386	1.393	-0.5	75	0.00
60	4-Chlorophenyl-phenylether	0.668	0.650	2.7	72	0.00
61	4-Nitroaniline	0.312	0.419	-34.3#	97	0.00
62	Azobenzene	1.315	1.378	-4.8	76	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	82	0.00
64	4,6-Dinitro-2-methylphenol	0.075	0.082	-9.3	84	0.00
65 c	n-Nitrosodiphenylamine	0.664	0.656	1.2	78	0.00
66	4-Bromophenyl-phenylether	0.234	0.204	12.8	71	0.00
67	Hexachlorobenzene	0.251	0.224	10.8	74	0.00
68	Atrazine	0.216	0.189	12.5	76	0.00
69 C	Pentachlorophenol	0.132	0.128	3.0	74	0.00
70	Phenanthrene	1.159	1.071	7.6	76	0.00
71	Anthracene	1.150	1.114	3.1	81	0.00
72	Carbazole	1.094	1.128	-3.1	81	0.00
73	Di-n-butylphthalate	1.284	1.088	15.3	65	0.00
74 C	Fluoranthene	1.266	1.225	3.2	78	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	80	0.00
76	Benzidine	0.676	0.595	12.0	77	0.00
77	Pyrene	1.223	1.204	1.6	80	0.00
78 S	Terphenyl-d14	0.856	0.733	14.4	69	0.00
79	Butylbenzylphthalate	0.503	0.464	7.8	70	0.00
80	Benzo(a)anthracene	1.172	1.127	3.8	79	0.00
81	3,3'-Dichlorobenzidine	0.430	0.433	-0.7	80	0.00
82	Chrysene	1.101	1.064	3.4	80	0.00
83	Bis(2-ethylhexyl)phthalate	0.807	0.635	21.3	62	0.00
84 c	Di-n-octyl phthalate	1.348	1.133	15.9	64	0.00
85	Indeno(1,2,3-cd)pyrene	1.255	1.334	-6.3	85	-0.01
86 I	Perylene-d12	1.000	1.000	0.0	91	-0.01
87	Benzo(b)fluoranthene	1.163	1.052	9.5	77	0.00

Data Path : Z:\HPCHEM1\BNA F\Data\BF030415\
Data File : BF077580.D
Acq On : 4 Mar 2015 13:24
Operator : TP/IZ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Mar 04 21:57:39 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF021915.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 04 13:27:09 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.140	1.070	6.1	91	0.00
89 C	Benzo(a)pyrene	1.108	1.066	3.8	84	-0.01
90	Dibenzo(a,h)anthracene	1.056	1.023	3.1	85	-0.01
91	Benzo(a,h,i)perylene	1.066	1.057	0.8	88	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0