

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA F\DATA\BF030518\
 Data File : BF103403.D
 Acq On : 5 Mar 2018 12:24
 Operator : SJ/JU
 Sample : J1724-15
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SP-25

Manual Integrations
 APPROVED

Sohil
 3/6/2018 4:22:41 PM

Quant Time: Mar 05 13:53:26 2018
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF022218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 05 13:40:47 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.87	152	142191	20.00	ng	0.00
21) Naphthalene-d8	8.15	136	592203	20.00	ng	0.00
38) Acenaphthene-d10	9.91	164	239395	20.00	ng	0.00
63) Phenanthrene-d10	11.40	188	374999	20.00	ng	0.00
75) Chrysene-d12	14.06	240	248217	20.00	ng	0.00
86) Perylene-d12	15.58	264	221706	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.50	112	994602	105.79	ng	0.00
7) Phenol-d6	6.52	99	1187537	103.23	ng	0.00
23) Nitrobenzene-d5	7.44	82	745460	71.89	ng	0.00
41) 2,4,6-Tribromophenol	10.71	330	178555	88.12	ng	0.00
44) 2-Fluorobiphenyl	9.23	172	995649	68.34	ng	0.00
78) Terphenyl-d14	12.99	244	700967	67.89	ng	0.00
Target Compounds						
10) Phenol	6.53	94	30473	2.282	ng	Qvalue # 59
49) Dimethylphthalate	9.62	163	73304	3.817	ng	99
70) Phenanthrene	11.43	178	212538	10.299	ng	99
71) Anthracene	11.48	178	63206	2.987	ng	98
74) Fluoranthene	12.63	202	390320	18.821	ng	99
77) Pyrene	12.86	202	391250	20.217	ng	99
80) Benzo(a)anthracene	14.05	228	188119	11.681	ng	98
82) Chrysene	14.09	228	182825	11.691	ng	98
85) Indeno(1,2,3-cd)pyrene	17.12	276	71430	5.770	ng	95
87) Benzo(b)fluoranthene	15.13	252	214790m	14.735	ng	
88) Benzo(k)fluoranthene	15.16	252	63105m	4.597	ng	
89) Benzo(a)pyrene	15.52	252	141226	10.849	ng	# 92
90) Dibenzo(a,h)anthracene	17.12	278	20903	2.078	ng	# 55
91) Benzo(a,h,i)perylene	17.59	276	72548	7.380	ng	# 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SVOASRV\HPCHEM1\BNA F\DATA\BF030518\
Data File : BF103403.D
Acq On : 5 Mar 2018 12:24
Operator : SJ/JU
Sample : J1724-15
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SP-25

Manual Integrations
APPROVED

Sohil
3/6/2018 4:22:41 PM

Quant Time: Mar 05 13:53:26 2018
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF022218.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 05 13:40:47 2018
Response via : Initial Calibration

