

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA F\DATA\BF030518\
 Data File : BF103426.D
 Acq On : 5 Mar 2018 23:08
 Operator : SJ/JU
 Sample : J1723-15
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SP-33

Manual Integrations
 APPROVED

Sohil
 3/6/2018 4:23:40 PM

Quant Time: Mar 06 01:16:46 2018
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF022218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 05 13:40:47 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.87	152	126025	20.00	ng	0.00
21) Naphthalene-d8	8.16	136	507657	20.00	ng	0.00
38) Acenaphthene-d10	9.92	164	203660	20.00	ng	0.00
63) Phenanthrene-d10	11.41	188	303070	20.00	ng	0.00
75) Chrysene-d12	14.07	240	200268	20.00	ng	0.00
86) Perylene-d12	15.59	264	158436	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.50	112	796332	95.57	ng	0.00
7) Phenol-d6	6.52	99	949828	93.15	ng	0.00
23) Nitrobenzene-d5	7.44	82	576672	64.87	ng	0.00
41) 2,4,6-Tribromophenol	10.72	330	117476	68.15	ng	0.00
44) 2-Fluorobiphenyl	9.23	172	799089	63.58	ng	0.00
78) Terphenyl-d14	12.99	244	519348	62.34	ng	0.00
Target Compounds						
15) Benzyl Alcohol	7.02	79	405673	52.799	ng	98
70) Phenanthrene	11.43	178	141435	8.480	ng	96
71) Anthracene	11.49	178	47264	2.763	ng	99
74) Fluoranthene	12.63	202	320627	19.130	ng	99
77) Pyrene	12.86	202	325381	20.839	ng	100
80) Benzo(a)anthracene	14.06	228	150525	11.584	ng	98
82) Chrysene	14.09	228	132992	10.541	ng	98
85) Indeno(1,2,3-cd)pyrene	17.13	276	50950	5.101	ng	98
87) Benzo(b)fluoranthene	15.14	252	144423m	13.864	ng	
88) Benzo(k)fluoranthene	15.16	252	49901m	5.087	ng	
89) Benzo(a)pyrene	15.52	252	100562	10.810	ng	# 87
90) Dibenzo(a,h)anthracene	17.13	278	14816	2.061	ng	# 63
91) Benzo(a,h,i)perylene	17.61	276	49482	7.043	ng	# 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SVOASRV\HPCHEM1\BNA F\DATA\BF030518\
Data File : BF103426.D
Acq On : 5 Mar 2018 23:08
Operator : SJ/JU
Sample : J1723-15
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SP-33

Manual Integrations
APPROVED

Sohil
3/6/2018 4:23:40 PM

Quant Time: Mar 06 01:16:46 2018
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF022218.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 05 13:40:47 2018
Response via : Initial Calibration

