

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF030519\
 Data File : BF112854.D
 Acq On : 5 Mar 2019 14:37
 Operator : JU/SJ
 Sample : K1727-08
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DUP-1

Quant Time: Mar 06 03:23:52 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF022719.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 01 15:40:43 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.75	152	147241	20.00	ng	0.00
21) Naphthalene-d8	8.03	136	465910	20.00	ng	0.00
39) Acenaphthene-d10	9.78	164	196533	20.00	ng	0.01
64) Phenanthrene-d10	11.26	188	342580	20.00	ng	0.01
76) Chrysene-d12	13.88	240	392169	20.00	ng	0.00
87) Perylene-d12	15.29	264	445319	20.00	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.35	112	267569	28.27	ng	0.00
7) Phenol-d6	6.37	99	213379	17.95	ng	0.00
23) Nitrobenzene-d5	7.30	82	530239	68.10	ng	0.00
42) 2,4,6-Tribromophenol	10.56	330	153336	73.49	ng	0.00
45) 2-Fluorobiphenyl	9.10	172	853027	68.28	ng	0.00
79) Terphenyl-d14	12.84	244	856875	42.36	ng	0.01
Target Compounds						
10) Phenol	6.39	94	44398	3.200	ng	89
38) 1-Methylnaphthalene	8.84	142	34160	2.361	ng	# 97
50) Dimethylphthalate	9.49	163	50425	3.480	ng	# 80
84) Bis(2-ethylhexyl)phthalate	13.87	149	75544	4.413	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF030519\
Data File : BF112854.D
Acq On : 5 Mar 2019 14:37
Operator : JU/SJ
Sample : K1727-08
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DUP-1

Quant Time: Mar 06 03:23:52 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF022719.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 01 15:40:43 2019
Response via : Initial Calibration

