

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF030519\
 Data File : BF112879.D
 Acq On : 6 Mar 2019 3:52
 Operator : JU/SJ
 Sample : K1705-18
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PST-028-B004-022719

Manual Integrations
 APPROVED

Sohil
 3/6/2019 12:24:57 PM

Quant Time: Mar 06 04:49:10 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF022719.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 01 15:40:43 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.74	152	149398	20.00	ng	0.00
21) Naphthalene-d8	8.02	136	520928	20.00	ng	0.00
39) Acenaphthene-d10	9.77	164	221709	20.00	ng	0.00
64) Phenanthrene-d10	11.25	188	420290	20.00	ng	0.00
76) Chrysene-d12	13.87	240	436399	20.00	ng	0.00
87) Perylene-d12	15.29	264	288404	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.36	112	1095063	114.02	ng	0.01
7) Phenol-d6	6.38	99	1368684	113.45	ng	0.01
23) Nitrobenzene-d5	7.30	82	829572	95.29	ng	0.00
42) 2,4,6-Tribromophenol	10.56	330	329624	140.04	ng	0.00
45) 2-Fluorobiphenyl	9.10	172	1207012	92.28	ng	0.00
79) Terphenyl-d14	12.83	244	1205781	53.56	ng	0.00
Target Compounds						
50) Dimethylphthalate	9.49	163	92733	5.673	ng	100
71) Phenanthrene	11.27	178	56740	2.713	ng	95
75) Fluoranthene	12.46	202	95463	4.261	ng	94
78) Pyrene	12.68	202	95353	2.592	ng	99
88) Benzo(b)fluoranthene	14.89	252	41969m	2.250	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF030519\
Data File : BF112879.D
Acq On : 6 Mar 2019 3:52
Operator : JU/SJ
Sample : K1705-18
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PST-028-B004-022719

Manual Integrations
APPROVED

Sohil
3/6/2019 12:24:57 PM

Quant Time: Mar 06 04:49:10 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF022719.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 01 15:40:43 2019
Response via : Initial Calibration

