

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF030615\
 Data File : BF077623.D
 Acq On : 6 Mar 2015 15:49
 Operator : TP/IZ
 Sample : G1426-01DL 5X
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DCW-WW-064-3315DL

Quant Time: Mar 06 23:49:30 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF021915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 06 22:10:32 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.20	152	63341	20.00	ng	0.00
21) Naphthalene-d8	8.78	136	251887	20.00	ng	0.00
38) Acenaphthene-d10	10.95	164	130692	20.00	ng	0.00
63) Phenanthrene-d10	12.79	188	254225	20.00	ng	0.00
75) Chrysene-d12	16.07	240	274480	20.00	ng	0.00
86) Perylene-d12	17.77	264	265654	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.49	112	43953	11.58	ng	0.00
7) Phenol-d6	6.76	99	35550	7.29	ng	0.00
23) Nitrobenzene-d5	7.89	82	80454	19.86	ng	0.00
41) 2,4,6-Tribromophenol	11.93	330	35484	28.20	ng	0.00
44) 2-Fluorobiphenyl	10.12	172	184493	22.01	ng	0.00
78) Terphenyl-d14	14.78	244	216116	18.41	ng	0.00
Target Compounds						Qvalue
13) 1,4-Dichlorobenzene	7.22	146	10152	2.05	ng	96
14) 1,2-Dichlorobenzene	7.41	146	209189	44.35	ng	97
30) 1,2,4-Trichlorobenzene	8.71	180	90918	22.54	ng	96
31) Naphthalene	8.80	128	148408	11.78	ng	98
33) 4-Chloroaniline	8.86	127	68308	12.20	ng	98
37) 2-Methylnaphthalene	9.66	142	58215	6.51	ng	99
51) Acenaphthene	10.99	154	16626	2.27	ng	98
56) 2,4-Dinitrotoluene	11.20	165	8729	3.51	ng	# 65

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF030615\
Data File : BF077623.D
Acq On : 6 Mar 2015 15:49
Operator : TP/IZ
Sample : G1426-01DL 5X
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DCW-WW-064-3315DL

Quant Time: Mar 06 23:49:30 2015
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF021915.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 06 22:10:32 2015
Response via : Initial Calibration

