

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF030615\
 Data File : BF077628.D
 Acq On : 6 Mar 2015 18:11
 Operator : TP/IZ
 Sample : G1433-01
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DCW-WW-072-3415

Quant Time: Mar 06 23:34:58 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF021915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 06 22:10:32 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.20	152	61418	20.00	ng	0.00
21) Naphthalene-d8	8.78	136	248304	20.00	ng	0.00
38) Acenaphthene-d10	10.95	164	126061	20.00	ng	0.00
63) Phenanthrene-d10	12.79	188	250456	20.00	ng	0.00
75) Chrysene-d12	16.09	240	272143	20.00	ng	0.02
86) Perylene-d12	17.89	264	254083	20.00	ng	0.13
System Monitoring Compounds						
5) 2-Fluorophenol	5.49	112	208689	56.72	ng	0.00
7) Phenol-d6	6.76	99	159259	33.67	ng	0.00
23) Nitrobenzene-d5	7.89	82	360455	90.27	ng	0.00
41) 2,4,6-Tribromophenol	11.93	330	162744	134.08	ng	0.00
44) 2-Fluorobiphenyl	10.13	172	787764	97.44	ng	0.00
78) Terphenyl-d14	14.79	244	903149	77.58	ng	0.00

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF030615\
Data File : BF077628.D
Acq On : 6 Mar 2015 18:11
Operator : TP/IZ
Sample : G1433-01
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DCW-WW-072-3415

Quant Time: Mar 06 23:34:58 2015
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF021915.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 06 22:10:32 2015
Response via : Initial Calibration

