

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF030615\
 Data File : BF077634.D
 Acq On : 6 Mar 2015 21:02
 Operator : TP/IZ
 Sample : G1424-08DL 5X
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SEP-1DL

Manual Integrations
 APPROVED

mohammad
 3/9/2015 3:25:08 PM

Quant Time: Mar 07 00:36:18 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF021915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 06 22:10:32 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.20	152	63361	20.00	ng	0.00
21) Naphthalene-d8	8.78	136	261725	20.00	ng	0.00
38) Acenaphthene-d10	10.95	164	131133	20.00	ng	0.00
63) Phenanthrene-d10	12.79	188	264297	20.00	ng	0.00
75) Chrysene-d12	16.08	240	271521	20.00	ng	0.00
86) Perylene-d12	17.85	264	275766	20.00	ng	0.09
System Monitoring Compounds						
5) 2-Fluorophenol	5.49	112	73301	19.31	ng	0.00
7) Phenol-d6	6.76	99	99164	20.32	ng	0.00
23) Nitrobenzene-d5	7.89	82	51864	12.32	ng	0.00
41) 2,4,6-Tribromophenol	11.93	330	22680	17.96	ng	0.00
44) 2-Fluorobiphenyl	10.12	172	99976	11.89	ng	0.00
78) Terphenyl-d14	14.78	244	95008	8.18	ng	0.00
Target Compounds						Qvalue
31) Naphthalene	8.80	128	79353	6.06	ng	97
37) 2-Methylnaphthalene	9.66	142	41200	4.43	ng	98
51) Acenaphthene	10.99	154	56542	7.68	ng	100
54) Dibenzofuran	11.20	168	51416	4.52	ng	94
57) Fluorene	11.63	166	102870	11.32	ng	99
70) Phenanthrene	12.82	178	889039	58.04	ng	99
71) Anthracene	12.88	178	200703	13.20	ng	98
72) Carbazole	13.09	167	44269	3.06	ng	99
74) Fluoranthene	14.29	202	701573	41.93	ng	99
77) Pyrene	14.57	202	719561	43.32	ng	# 95
80) Benzo(a)anthracene	16.07	228	285500	17.94	ng	98
82) Chrysene	16.11	228	242528	16.23	ng	98
85) Indeno(1,2,3-cd)pyrene	19.26	276	116561	6.84	ng	95
87) Benzo(b)fluoranthene	17.39	252	288414m	17.99	ng	
88) Benzo(k)fluoranthene	17.43	252	68112m	4.33	ng	
89) Benzo(a)pyrene	17.78	252	207909	13.61	ng	99
90) Dibenzo(a,h)anthracene	19.28	278	29675	2.04	ng	94
91) Benzo(g,h,i)perylene	19.68	276	122689	8.35	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF030615\
Data File : BF077634.D
Acq On : 6 Mar 2015 21:02
Operator : TP/IZ
Sample : G1424-08DL 5X
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SEP-1DL

Manual Integrations
APPROVED

mohammad
3/9/2015 3:25:08 PM

Quant Time: Mar 07 00:36:18 2015
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF021915.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 06 22:10:32 2015
Response via : Initial Calibration

