

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF030623\
 Data File : BF132649.D
 Acq On : 06 Mar 2023 15:20
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 07 00:53:59 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF022223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 24 23:17:30 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	105	0.00
2	1,4-Dioxane	40.000	40.090	-0.2	101	-0.03
3	Pyridine	40.000	39.778	0.6	101	-0.02
4	n-Nitrosodimethylamine	40.000	38.182	4.5	94	-0.02
5 S	2-Fluorophenol	80.000	78.432	2.0	102	0.00
6	Aniline	40.000	39.967	0.1	100	0.00
7 S	Phenol-d6	80.000	77.329	3.3	100	0.00
8	2-Chlorophenol	40.000	39.450	1.4	102	0.00
9	Benzaldehyde	40.000	40.463	-1.2	112	0.00
10 C	Phenol	40.000	39.474	1.3	102	0.00
11	bis(2-Chloroethyl)ether	40.000	39.261	1.8	100	0.00
12	1,3-Dichlorobenzene	40.000	39.646	0.9	103	0.00
13 C	1,4-Dichlorobenzene	40.000	39.400	1.5	102	0.00
14	1,2-Dichlorobenzene	40.000	36.689	8.3	99	0.00
15	Benzyl Alcohol	40.000	40.143	-0.4	100	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.189	4.5	97	0.00
17	2-Methylphenol	40.000	40.113	-0.3	103	0.00
18	Hexachloroethane	40.000	39.251	1.9	100	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	35.501	11.2	94	0.00
20	3+4-Methylphenols	40.000	35.667	10.8	104	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	107	0.00
22	Acetophenone	40.000	36.770	8.1	98	0.00
23 S	Nitrobenzene-d5	80.000	76.078	4.9	98	0.00
24	Nitrobenzene	40.000	38.678	3.3	99	0.00
25	Isophorone	40.000	39.763	0.6	104	0.00
26 C	2-Nitrophenol	40.000	41.994	-5.0	106	0.00
27	2,4-Dimethylphenol	40.000	39.924	0.2	103	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.954	0.1	104	0.00
29 C	2,4-Dichlorophenol	40.000	40.488	-1.2	104	0.00
30	1,2,4-Trichlorobenzene	40.000	40.272	-0.7	104	0.00
31	Naphthalene	40.000	38.968	2.6	103	0.00
32	Benzoic acid	40.000	39.710	0.7	102	0.00
33	4-Chloroaniline	40.000	41.019	-2.5	105	0.00
34 C	Hexachlorobutadiene	40.000	40.155	-0.4	103	0.00
35	Caprolactam	40.000	43.008	-7.5	107	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.444	-1.1	103	0.00
37	2-Methylnaphthalene	40.000	38.746	3.1	102	0.00
38	1-Methylnaphthalene	40.000	38.773	3.1	103	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	107	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.802	0.5	104	0.00
41 P	Hexachlorocyclopentadiene	40.000	35.820	10.4	83	0.00
42 S	2,4,6-Tribromophenol	80.000	81.828	-2.3	105	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.400	-1.0	103	0.00
44	2,4,5-Trichlorophenol	40.000	40.247	-0.6	102	0.00
45 S	2-Fluorobiphenyl	80.000	71.249	10.9	101	0.00
46	1,1'-Biphenyl	40.000	39.286	1.8	103	0.00
47	2-Chloronaphthalene	40.000	39.497	1.3	103	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF030623\
 Data File : BF132649.D
 Acq On : 06 Mar 2023 15:20
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 07 00:53:59 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF022223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 24 23:17:30 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.053	2.4	99	0.00
49	Acenaphthylene	40.000	39.165	2.1	102	0.00
50	Dimethylphthalate	40.000	40.750	-1.9	106	0.00
51	2,6-Dinitrotoluene	40.000	41.475	-3.7	105	0.00
52 C	Acenaphthene	40.000	39.471	1.3	102	0.00
53	3-Nitroaniline	40.000	42.575	-6.4	108	0.00
54 P	2,4-Dinitrophenol	40.000	38.522	3.7	94	0.00
55	Dibenzofuran	40.000	39.344	1.6	103	0.00
56 P	4-Nitrophenol	40.000	40.200	-0.5	98	0.00
57	2,4-Dinitrotoluene	40.000	42.428	-6.1	108	0.00
58	Fluorene	40.000	39.094	2.3	103	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	42.340	-5.9	106	0.00
60	Diethylphthalate	40.000	40.313	-0.8	104	0.00
61	4-Chlorophenyl-phenylether	40.000	39.978	0.1	105	0.00
62	4-Nitroaniline	40.000	42.934	-7.3	108	0.00
63	Azobenzene	40.000	39.499	1.3	101	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	106	0.00
65	4,6-Dinitro-2-methylphenol	40.000	43.508	-8.8	104	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.181	-3.0	107	0.00
67	4-Bromophenyl-phenylether	40.000	41.506	-3.8	107	0.00
68	Hexachlorobenzene	40.000	41.457	-3.6	107	0.00
69	Atrazine	40.000	40.288	-0.7	103	0.00
70 C	Pentachlorophenol	40.000	43.209	-8.0	103	0.00
71	Phenanthrene	40.000	39.029	2.4	102	0.00
72	Anthracene	40.000	39.386	1.5	103	0.00
73	Carbazole	40.000	39.379	1.6	103	0.00
74	Di-n-butylphthalate	40.000	39.949	0.1	102	0.00
75 C	Fluoranthene	40.000	40.036	-0.1	104	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	101	0.00
77	Benzydine	40.000	33.880	15.3	69	0.00
78	Pyrene	40.000	43.319	-8.3	103	0.00
79 S	Terphenyl-d14	80.000	83.141	-3.9	101	0.00
80	Butylbenzylphthalate	40.000	43.194	-8.0	100	0.00
81	Benzo(a)anthracene	40.000	40.979	-2.4	98	0.00
82	3,3'-Dichlorobenzidine	40.000	39.943	0.1	98	0.00
83	Chrysene	40.000	40.933	-2.3	101	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	42.171	-5.4	98	0.00
85 c	Di-n-octyl phthalate	40.000	42.457	-6.1	101	0.00
86 I	Perylene-d12	20.000	20.000	0.0	99	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	45.062	-12.7	102	0.00
88	Benzo(b)fluoranthene	40.000	40.124	-0.3	96	0.00
89	Benzo(k)fluoranthene	40.000	41.023	-2.6	99	0.00
90 C	Benzo(a)pyrene	40.000	41.403	-3.5	98	0.00
91	Dibenzo(a,h)anthracene	40.000	44.745	-11.9	100	0.00
92	Benzo(g,h,i)perylene	40.000	42.055	-5.1	102	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF030623\
Data File : BF132649.D
Acq On : 06 Mar 2023 15:20
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Mar 07 00:53:59 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF022223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 24 23:17:30 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0