

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF030623\
 Data File : BF132641.D
 Acq On : 06 Mar 2023 10:13
 Operator : CG\JU
 Sample : PB151188BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB151188BL

Quant Time: Mar 07 00:50:36 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF022223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 24 23:17:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.998	152	208665	20.000	ng	0.00
21) Naphthalene-d8	8.286	136	772686	20.000	ng	0.00
39) Acenaphthene-d10	10.045	164	425156	20.000	ng	0.00
64) Phenanthrene-d10	11.545	188	835380	20.000	ng	0.00
76) Chrysene-d12	14.204	240	600964	20.000	ng	0.00
86) Perylene-d12	15.757	264	498196	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.640	112	1469773	114.317	ng	0.00
7) Phenol-d6	6.628	99	1924421	114.689	ng	0.00
23) Nitrobenzene-d5	7.569	82	1247949	76.626	ng	0.00
42) 2,4,6-Tribromophenol	10.845	330	541426	117.919	ng	0.00
45) 2-Fluorobiphenyl	9.363	172	2004690	71.796	ng	0.00
79) Terphenyl-d14	13.139	244	2606389	73.328	ng	0.00


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF030623\
Data File : BF132641.D
Acq On : 06 Mar 2023 10:13
Operator : CG\JU
Sample : PB151188BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB151188BL

Quant Time: Mar 07 00:50:36 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF022223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 24 23:17:30 2023
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.998 min Scan# 710
Delta R.T. -0.009 min
Lab File: BF132641.D
Acq: 06 Mar 2023 10:13




Instrument :
BNA_F
ClientSampleId :
PB151188BL


Tgt Ion:152 Resp: 208665
Ion Ratio Lower Upper
152 100
150 157.4 125.4 188.2
115 61.5 47.8 71.6



#5
2-Fluorophenol
Concen: 114.317 ng
RT: 5.640 min Scan# 519
Delta R.T. 0.009 min
Lab File: BF132641.D
Acq: 06 Mar 2023 10:13


Tgt Ion:112 Resp: 1469773
Ion Ratio Lower Upper
112 100
64 58.6 51.6 77.4
63 30.9 27.4 41.2








#7
Phenol-d6
Concen: 114.689 ng
RT: 6.628 min Scan# 61
Delta R.T. -0.009 min
Lab File: BF132641.D
Acq: 06 Mar 2023 10:13

Instrument :
BNA_F
ClientSampleId :
PB151188BL



Tgt Ion: 99 Resp: 1924421




Ion	Ratio	Lower	Upper
99	100		
42	15.9	13.9	20.9
71	35.8	28.5	42.7




#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.286 min Scan# 969
Delta R.T. -0.003 min
Lab File: BF132641.D
Acq: 06 Mar 2023 10:13

Tgt Ion: 136 Resp: 772686

Ion	Ratio	Lower	Upper
136	100		
137	10.8	8.9	13.3
54	7.3	6.2	9.4
68	6.2	5.2	7.8





#23

Nitrobenzene-d5

Concen: 76.626 ng

RT: 7.569 min Scan# 84

Delta R.T. -0.003 min

Lab File: BF132641.D

Acq: 06 Mar 2023 10:13

Instrument :

BNA_F

ClientSampleId :

PB151188BL





Tgt Ion: 82 Resp: 1247949

Ion Ratio Lower Upper

82 100

128 43.0 32.2 48.2

54 44.8 38.4 57.6



#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 10.045 min Scan# 1268

Delta R.T. -0.009 min

Lab File: BF132641.D

Acq: 06 Mar 2023 10:13




Tgt Ion:164 Resp: 425156


Ion Ratio Lower Upper

164 100

162 100.9 79.8 119.8

160 45.7 35.8 53.6





#42

2,4,6-Tribromophenol

Concen: 117.919 ng

RT: 10.845 min Scan# 1404

Delta R.T. -0.003 min

Lab File: BF132641.D

Acq: 06 Mar 2023 10:13

Instrument :

BNA_F

ClientSampleId :

PB151188BL





Tgt Ion:330 Resp: 541426

Ion Ratio Lower Upper

330 100

332 96.5 76.2 114.4

141 33.0 28.4 42.6



#45

2-Fluorobiphenyl

Concen: 71.796 ng

RT: 9.363 min Scan# 1152

Delta R.T. -0.009 min

Lab File: BF132641.D

Acq: 06 Mar 2023 10:13




Tgt Ion:172 Resp: 2004690


Ion Ratio Lower Upper

172 100

171 35.9 28.8 43.2

170 24.5 19.7 29.5









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.545 min Scan# 11
Delta R.T. -0.003 min
Lab File: BF132641.D
Acq: 06 Mar 2023 10:13




Instrument :
BNA_F
ClientSampleId :
PB151188BL


Tgt Ion:188 Resp: 835380
Ion Ratio Lower Upper
188 100
94 8.4 8.3 12.5
80 9.0 8.7 13.1



#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.204 min Scan# 1975
Delta R.T. -0.003 min
Lab File: BF132641.D
Acq: 06 Mar 2023 10:13

Tgt Ion:240 Resp: 600964
Ion Ratio Lower Upper
240 100
120 10.6 7.2 10.8
236 26.7 21.8 32.6





#79

Terphenyl-d14

Concen: 73.328 ng

RT: 13.139 min Scan# 11

Delta R.T. -0.003 min

Lab File: BF132641.D

Acq: 06 Mar 2023 10:13

Instrument :

BNA_F

ClientSampleId :

PB151188BL





Tgt Ion:244 Resp: 2606389

Ion Ratio Lower Upper

244 100

212 8.2 6.5 9.7

122 9.2 7.5 11.3



#86

Perylene-d12

Concen: 20.000 ng

RT: 15.757 min Scan# 2239

Delta R.T. 0.003 min

Lab File: BF132641.D

Acq: 06 Mar 2023 10:13

Tgt Ion:264 Resp: 498196

Ion Ratio Lower Upper

264 100

260 24.4 18.8 28.2

265 21.5 17.1 25.7

