

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF030719\
 Data File : BF112928.D
 Acq On : 7 Mar 2019 18:22
 Operator : JU/SJ
 Sample : K1757-12
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TP-6AB-S

Quant Time: Mar 08 00:59:49 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF022719.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 01 15:40:43 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.73	152	246786	20.00	ng	0.00
21) Naphthalene-d8	8.01	136	955204	20.00	ng	-0.01
39) Acenaphthene-d10	9.76	164	471511	20.00	ng	-0.01
64) Phenanthrene-d10	11.24	188	987583	20.00	ng	0.00
76) Chrysene-d12	13.86	240	889830	20.00	ng	-0.01
87) Perylene-d12	15.26	264	776302	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.35	112	1359542	85.69	ng	0.00
7) Phenol-d6	6.37	99	1529350	76.74	ng	0.00
23) Nitrobenzene-d5	7.29	82	1244978	77.99	ng	-0.01
42) 2,4,6-Tribromophenol	10.55	330	643291	128.51	ng	0.00
45) 2-Fluorobiphenyl	9.09	172	2230607	76.29	ng	-0.01
79) Terphenyl-d14	12.82	244	2631863	57.34	ng	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF030719\
Data File : BF112928.D
Acq On : 7 Mar 2019 18:22
Operator : JU/SJ
Sample : K1757-12
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
TP-6AB-S

Quant Time: Mar 08 00:59:49 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF022719.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 01 15:40:43 2019
Response via : Initial Calibration

