

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF030722\
 Data File : BF127215.D
 Acq On : 07 Mar 2022 10:51
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 07 14:56:33 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF021622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Mar 05 02:16:58 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	74	0.00
2	1,4-Dioxane	0.592	0.675	-14.0	86	0.00
3	Pyridine	1.553	1.778	-14.5	85	0.00
4	n-Nitrosodimethylamine	0.761	0.829	-8.9	79	0.00
5 S	2-Fluorophenol	1.294	1.305	-0.9	75	0.00
6	Aniline	2.069	2.018	2.5	71	0.00
7 S	Phenol-d6	1.746	1.757	-0.6	75	0.00
8	2-Chlorophenol	1.388	1.460	-5.2	78	0.00
9	Benzaldehyde	1.063	1.135	-6.8	88	0.00
10 C	Phenol	1.764	1.738	1.5	74	0.00
11	bis(2-Chloroethyl)ether	1.384	1.372	0.9	74	0.00
12	1,3-Dichlorobenzene	1.498	1.554	-3.7	78	0.00
13 C	1,4-Dichlorobenzene	1.518	1.633	-7.6	81	0.00
14	1,2-Dichlorobenzene	1.448	1.515	-4.6	78	0.00
15	Benzyl Alcohol	1.471	1.345	8.6	67	0.00
16	2,2'-oxybis(1-Chloropropane	2.179	1.814	16.8	61	0.00
17	2-Methylphenol	1.201	1.208	-0.6	74	0.00
18	Hexachloroethane	0.582	0.610	-4.8	78	0.00
19 P	n-Nitroso-di-n-propylamine	1.125	1.063	5.5	70	0.00
20	3+4-Methylphenols	1.560	1.532	1.8	72	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	67	0.00
22	Acetophenone	0.527	0.551	-4.6	73	0.00
23 S	Nitrobenzene-d5	0.443	0.470	-6.1	74	0.00
24	Nitrobenzene	0.419	0.438	-4.5	72	0.00
25	Isophorone	0.733	0.777	-6.0	73	0.00
26 C	2-Nitrophenol	0.178	0.209	-17.4	79	0.00
27	2,4-Dimethylphenol	0.293	0.311	-6.1	74	0.00
28	bis(2-Chloroethoxy)methane	0.445	0.466	-4.7	73	0.00
29 C	2,4-Dichlorophenol	0.275	0.314	-14.2	79	0.00
30	1,2,4-Trichlorobenzene	0.308	0.346	-12.3	79	0.00
31	Naphthalene	1.044	1.134	-8.6	77	0.00
32	Benzoic acid	0.208	0.209	-0.5	64	0.00
33	4-Chloroaniline	0.411	0.444	-8.0	75	0.00
34 C	Hexachlorobutadiene	0.180	0.212	-17.8	82	0.00
35	Caprolactam	0.096	0.102	-6.2	72	0.00
36 C	4-Chloro-3-methylphenol	0.352	0.385	-9.4	74	0.00
37	2-Methylnaphthalene	0.677	0.740	-9.3	76	0.00
38	1-Methylnaphthalene	0.649	0.736	-13.4	79	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	68	0.00
40	1,2,4,5-Tetrachlorobenzene	0.526	0.606	-15.2	83	0.00
41 P	Hexachlorocyclopentadiene	0.286	0.285	0.3	67	0.00
42 S	2,4,6-Tribromophenol	0.230	0.272	-18.3	81	0.00
43 C	2,4,6-Trichlorophenol	0.386	0.418	-8.3	76	0.00
44	2,4,5-Trichlorophenol	0.406	0.447	-10.1	77	0.00
45 S	2-Fluorobiphenyl	1.359	1.545	-13.7	83	0.00
46	1,1'-Biphenyl	1.585	1.711	-7.9	78	0.00
47	2-Chloronaphthalene	1.175	1.307	-11.2	79	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF030722\
 Data File : BF127215.D
 Acq On : 07 Mar 2022 10:51
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 07 14:56:33 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF021622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Mar 05 02:16:58 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.464	0.495	-6.7	75	0.00
49	Acenaphthylene	1.909	2.058	-7.8	76	0.00
50	Dimethylphthalate	1.430	1.525	-6.6	76	0.00
51	2,6-Dinitrotoluene	0.291	0.317	-8.9	76	0.00
52 C	Acenaphthene	1.241	1.377	-11.0	78	0.00
53	3-Nitroaniline	0.356	0.383	-7.6	75	0.00
54 P	2,4-Dinitrophenol	0.154	0.163	-5.8	68	0.00
55	Dibenzofuran	1.707	1.854	-8.6	77	0.00
56 P	4-Nitrophenol	0.263	0.245	6.8	63	0.00
57	2,4-Dinitrotoluene	0.393	0.432	-9.9	76	0.00
58	Fluorene	1.330	1.417	-6.5	76	0.00
59	2,3,4,6-Tetrachlorophenol	0.322	0.362	-12.4	79	0.00
60	Diethylphthalate	1.494	1.630	-9.1	77	0.00
61	4-Chlorophenyl-phenylether	0.627	0.699	-11.5	79	0.00
62	4-Nitroaniline	0.351	0.360	-2.6	70	0.00
63	Azobenzene	1.663	1.660	0.2	70	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	67	0.00
65	4,6-Dinitro-2-methylphenol	0.118	0.139	-17.8	75	0.00
66 c	n-Nitrosodiphenylamine	0.656	0.721	-9.9	77	0.00
67	4-Bromophenyl-phenylether	0.223	0.261	-17.0	81	0.00
68	Hexachlorobenzene	0.241	0.281	-16.6	83	0.00
69	Atrazine	0.203	0.234	-15.3	81	0.00
70 C	Pentachlorophenol	0.131	0.122	6.9	60	0.00
71	Phenanthrene	1.119	1.231	-10.0	78	0.00
72	Anthracene	1.114	1.194	-7.2	75	0.00
73	Carbazole	1.022	1.080	-5.7	74	0.00
74	Di-n-butylphthalate	1.314	1.422	-8.2	75	0.00
75 C	Fluoranthene	1.063	1.204	-13.3	80	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	76	0.00
77	Benzidine	0.544	0.487	10.5	66	0.00
78	Pyrene	1.738	1.768	-1.7	77	0.00
79 S	Terphenyl-d14	1.361	1.531	-12.5	88	0.00
80	Butylbenzylphthalate	0.823	0.838	-1.8	77	0.00
81	Benzo(a)anthracene	1.348	1.402	-4.0	81	0.00
82	3,3'-Dichlorobenzidine	0.425	0.442	-4.0	79	0.00
83	Chrysene	1.268	1.385	-9.2	85	0.00
84	Bis(2-ethylhexyl)phthalate	1.080	1.060	1.9	75	0.00
85 c	Di-n-octyl phthalate	1.555	1.585	-1.9	80	0.00
86 I	Perylene-d12	1.000	1.000	0.0	75	0.00
87	Indeno(1,2,3-cd)pyrene	1.311	1.493	-13.9	82	0.00
88	Benzo(b)fluoranthene	1.343	1.553	-15.6	90	0.00
89	Benzo(k)fluoranthene	1.296	1.439	-11.0	84	0.00
90 C	Benzo(a)pyrene	1.049	1.145	-9.2	83	0.00
91	Dibenzo(a,h)anthracene	1.085	1.221	-12.5	82	0.00
92	Benzo(g,h,i)perylene	1.069	1.181	-10.5	79	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF030722\
Data File : BF127215.D
Acq On : 07 Mar 2022 10:51
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Mar 07 14:56:33 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF021622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Mar 05 02:16:58 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0