

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF030722\
 Data File : BF127224.D
 Acq On : 07 Mar 2022 16:12
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 08 00:02:58 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF021622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Mar 05 02:16:58 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	80	0.00
2	1,4-Dioxane	0.592	0.644	-8.8	89	-0.02
3	Pyridine	1.553	1.657	-6.7	85	-0.01
4	n-Nitrosodimethylamine	0.761	0.806	-5.9	83	-0.01
5 S	2-Fluorophenol	1.294	1.357	-4.9	85	0.00
6	Aniline	2.069	2.107	-1.8	80	0.00
7 S	Phenol-d6	1.746	1.819	-4.2	83	0.00
8	2-Chlorophenol	1.388	1.476	-6.3	85	0.00
9	Benzaldehyde	1.063	1.191	-12.0	100	0.00
10 C	Phenol	1.764	1.854	-5.1	85	0.00
11	bis(2-Chloroethyl)ether	1.384	1.429	-3.3	83	0.00
12	1,3-Dichlorobenzene	1.498	1.607	-7.3	87	0.00
13 C	1,4-Dichlorobenzene	1.518	1.640	-8.0	88	0.00
14	1,2-Dichlorobenzene	1.448	1.590	-9.8	89	0.00
15	Benzyl Alcohol	1.471	1.530	-4.0	82	0.00
16	2,2'-oxybis(1-Chloropropane	2.179	1.954	10.3	71	0.00
17	2-Methylphenol	1.201	1.292	-7.6	86	0.00
18	Hexachloroethane	0.582	0.646	-11.0	90	0.00
19 P	n-Nitroso-di-n-propylamine	1.125	1.192	-6.0	85	0.00
20	3+4-Methylphenols	1.560	1.707	-9.4	87	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	77	0.00
22	Acetophenone	0.527	0.567	-7.6	88	0.00
23 S	Nitrobenzene-d5	0.443	0.470	-6.1	85	0.00
24	Nitrobenzene	0.419	0.439	-4.8	84	0.00
25	Isophorone	0.733	0.760	-3.7	83	0.00
26 C	2-Nitrophenol	0.178	0.210	-18.0	92	0.00
27	2,4-Dimethylphenol	0.293	0.316	-7.8	87	0.00
28	bis(2-Chloroethoxy)methane	0.445	0.468	-5.2	85	0.00
29 C	2,4-Dichlorophenol	0.275	0.311	-13.1	90	0.00
30	1,2,4-Trichlorobenzene	0.308	0.357	-15.9	94	0.00
31	Naphthalene	1.044	1.097	-5.1	86	0.00
32	Benzoic acid	0.208	0.210	-1.0	74	0.02
33	4-Chloroaniline	0.411	0.449	-9.2	88	0.00
34 C	Hexachlorobutadiene	0.180	0.210	-16.7	94	0.00
35	Caprolactam	0.096	0.104	-8.3	85	0.01
36 C	4-Chloro-3-methylphenol	0.352	0.397	-12.8	89	0.00
37	2-Methylnaphthalene	0.677	0.758	-12.0	90	0.00
38	1-Methylnaphthalene	0.649	0.767	-18.2	95	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	85	0.00
40	1,2,4,5-Tetrachlorobenzene	0.526	0.582	-10.6	100	0.00
41 P	Hexachlorocyclopentadiene	0.286	0.283	1.0	84	0.00
42 S	2,4,6-Tribromophenol	0.230	0.276	-20.0	104	0.00
43 C	2,4,6-Trichlorophenol	0.386	0.412	-6.7	94	0.00
44	2,4,5-Trichlorophenol	0.406	0.439	-8.1	95	0.00
45 S	2-Fluorobiphenyl	1.359	1.481	-9.0	100	0.00
46	1,1'-Biphenyl	1.585	1.680	-6.0	96	0.00
47	2-Chloronaphthalene	1.175	1.268	-7.9	96	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF030722\
 Data File : BF127224.D
 Acq On : 07 Mar 2022 16:12
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 08 00:02:58 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF021622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Mar 05 02:16:58 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.464	0.465	-0.2	88	0.00
49	Acenaphthylene	1.909	1.975	-3.5	92	0.00
50	Dimethylphthalate	1.430	1.545	-8.0	96	0.00
51	2,6-Dinitrotoluene	0.291	0.322	-10.7	97	0.00
52 C	Acenaphthene	1.241	1.328	-7.0	95	0.00
53	3-Nitroaniline	0.356	0.380	-6.7	93	0.00
54 P	2,4-Dinitrophenol	0.154	0.178	-15.6	94	0.00
55	Dibenzofuran	1.707	1.795	-5.2	94	0.00
56 P	4-Nitrophenol	0.263	0.248	5.7	80	0.00
57	2,4-Dinitrotoluene	0.393	0.426	-8.4	94	0.00
58	Fluorene	1.330	1.397	-5.0	95	0.00
59	2,3,4,6-Tetrachlorophenol	0.322	0.359	-11.5	98	0.00
60	Diethylphthalate	1.494	1.630	-9.1	97	0.00
61	4-Chlorophenyl-phenylether	0.627	0.698	-11.3	99	0.00
62	4-Nitroaniline	0.351	0.367	-4.6	90	0.01
63	Azobenzene	1.663	1.604	3.5	86	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	86	0.00
65	4,6-Dinitro-2-methylphenol	0.118	0.140	-18.6	98	0.00
66 c	n-Nitrosodiphenylamine	0.656	0.689	-5.0	94	0.00
67	4-Bromophenyl-phenylether	0.223	0.261	-17.0	104	0.00
68	Hexachlorobenzene	0.241	0.288	-19.5	108	0.00
69	Atrazine	0.203	0.235	-15.8	104	0.00
70 C	Pentachlorophenol	0.131	0.135	-3.1	85	0.00
71	Phenanthrene	1.119	1.201	-7.3	98	0.00
72	Anthracene	1.114	1.181	-6.0	95	0.00
73	Carbazole	1.022	1.091	-6.8	96	0.00
74	Di-n-butylphthalate	1.314	1.448	-10.2	98	0.00
75 C	Fluoranthene	1.063	1.148	-8.0	98	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	96	0.00
77	Benzydine	0.544	0.509	6.4	88	0.00
78	Pyrene	1.738	1.740	-0.1	96	0.00
79 S	Terphenyl-d14	1.361	1.427	-4.8	104	0.00
80	Butylbenzylphthalate	0.823	0.833	-1.2	97	0.00
81	Benzo(a)anthracene	1.348	1.459	-8.2	106	0.00
82	3,3'-Dichlorobenzidine	0.425	0.458	-7.8	104	0.00
83	Chrysene	1.268	1.355	-6.9	105	0.00
84	Bis(2-ethylhexyl)phthalate	1.080	1.162	-7.6	105	0.00
85 c	Di-n-octyl phthalate	1.555	1.770	-13.8	113	0.00
86 I	Perylene-d12	1.000	1.000	0.0	91	0.00
87	Indeno(1,2,3-cd)pyrene	1.311	1.393	-6.3	93	0.02
88	Benzo(b)fluoranthene	1.343	1.485	-10.6	105	0.00
89	Benzo(k)fluoranthene	1.296	1.473	-13.7	104	0.00
90 C	Benzo(a)pyrene	1.049	1.168	-11.3	102	0.00
91	Dibenzo(a,h)anthracene	1.085	1.155	-6.5	94	0.01
92	Benzo(g,h,i)perylene	1.069	1.105	-3.4	90	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF030722\
Data File : BF127224.D
Acq On : 07 Mar 2022 16:12
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Mar 08 00:02:58 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF021622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Mar 05 02:16:58 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0