

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF030722\
 Data File : BF127224.D
 Acq On : 07 Mar 2022 16:12
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 08 00:02:58 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF021622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Mar 05 02:16:58 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	80	0.00
2	1,4-Dioxane	40.000	43.505	-8.8	89	-0.02
3	Pyridine	40.000	42.684	-6.7	85	-0.01
4	n-Nitrosodimethylamine	40.000	42.368	-5.9	83	-0.01
5 S	2-Fluorophenol	80.000	83.893	-4.9	85	0.00
6	Aniline	40.000	40.731	-1.8	80	0.00
7 S	Phenol-d6	80.000	83.338	-4.2	83	0.00
8	2-Chlorophenol	40.000	42.524	-6.3	85	0.00
9	Benzaldehyde	40.000	44.807	-12.0	100	0.00
10 C	Phenol	40.000	42.039	-5.1	85	0.00
11	bis(2-Chloroethyl)ether	40.000	41.292	-3.2	83	0.00
12	1,3-Dichlorobenzene	40.000	42.907	-7.3	87	0.00
13 C	1,4-Dichlorobenzene	40.000	43.194	-8.0	88	0.00
14	1,2-Dichlorobenzene	40.000	43.913	-9.8	89	0.00
15	Benzyl Alcohol	40.000	41.598	-4.0	82	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	35.871	10.3	71	0.00
17	2-Methylphenol	40.000	43.030	-7.6	86	0.00
18	Hexachloroethane	40.000	44.399	-11.0	90	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	42.390	-6.0	85	0.00
20	3+4-Methylphenols	40.000	43.782	-9.5	87	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	77	0.00
22	Acetophenone	40.000	43.107	-7.8	88	0.00
23 S	Nitrobenzene-d5	80.000	84.812	-6.0	85	0.00
24	Nitrobenzene	40.000	41.904	-4.8	84	0.00
25	Isophorone	40.000	41.453	-3.6	83	0.00
26 C	2-Nitrophenol	40.000	47.076	-17.7	92	0.00
27	2,4-Dimethylphenol	40.000	43.118	-7.8	87	0.00
28	bis(2-Chloroethoxy)methane	40.000	42.118	-5.3	85	0.00
29 C	2,4-Dichlorophenol	40.000	45.192	-13.0	90	0.00
30	1,2,4-Trichlorobenzene	40.000	46.431	-16.1	94	0.00
31	Naphthalene	40.000	42.011	-5.0	86	0.00
32	Benzoic acid	40.000	40.398	-1.0	74	0.02
33	4-Chloroaniline	40.000	43.701	-9.3	88	0.00
34 C	Hexachlorobutadiene	40.000	46.682	-16.7	94	0.00
35	Caprolactam	40.000	43.378	-8.4	85	0.01
36 C	4-Chloro-3-methylphenol	40.000	45.194	-13.0	89	0.00
37	2-Methylnaphthalene	40.000	44.753	-11.9	90	0.00
38	1-Methylnaphthalene	40.000	47.276	-18.2	95	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	85	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	44.217	-10.5	100	0.00
41 P	Hexachlorocyclopentadiene	40.000	39.599	1.0	84	0.00
42 S	2,4,6-Tribromophenol	80.000	95.861	-19.8	104	0.00
43 C	2,4,6-Trichlorophenol	40.000	42.699	-6.7	94	0.00
44	2,4,5-Trichlorophenol	40.000	43.175	-7.9	95	0.00
45 S	2-Fluorobiphenyl	80.000	87.206	-9.0	100	0.00
46	1,1'-Biphenyl	40.000	42.412	-6.0	96	0.00
47	2-Chloronaphthalene	40.000	43.173	-7.9	96	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF030722\
 Data File : BF127224.D
 Acq On : 07 Mar 2022 16:12
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 08 00:02:58 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF021622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Mar 05 02:16:58 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.144	-0.4	88	0.00
49	Acenaphthylene	40.000	41.387	-3.5	92	0.00
50	Dimethylphthalate	40.000	43.216	-8.0	96	0.00
51	2,6-Dinitrotoluene	40.000	44.308	-10.8	97	0.00
52 C	Acenaphthene	40.000	42.801	-7.0	95	0.00
53	3-Nitroaniline	40.000	42.790	-7.0	93	0.00
54 P	2,4-Dinitrophenol	40.000	46.280	-15.7	94	0.00
55	Dibenzofuran	40.000	42.058	-5.1	94	0.00
56 P	4-Nitrophenol	40.000	37.842	5.4	80	0.00
57	2,4-Dinitrotoluene	40.000	43.304	-8.3	94	0.00
58	Fluorene	40.000	42.021	-5.1	95	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	44.544	-11.4	98	0.00
60	Diethylphthalate	40.000	43.644	-9.1	97	0.00
61	4-Chlorophenyl-phenylether	40.000	44.542	-11.4	99	0.00
62	4-Nitroaniline	40.000	41.760	-4.4	90	0.01
63	Azobenzene	40.000	38.583	3.5	86	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	86	0.00
65	4,6-Dinitro-2-methylphenol	40.000	47.823	-19.6	98	0.00
66 c	n-Nitrosodiphenylamine	40.000	42.054	-5.1	94	0.00
67	4-Bromophenyl-phenylether	40.000	46.763	-16.9	104	0.00
68	Hexachlorobenzene	40.000	47.833	-19.6	108	0.00
69	Atrazine	40.000	46.226	-15.6	104	0.00
70 C	Pentachlorophenol	40.000	41.141	-2.9	85	0.00
71	Phenanthrene	40.000	42.929	-7.3	98	0.00
72	Anthracene	40.000	42.386	-6.0	95	0.00
73	Carbazole	40.000	42.719	-6.8	96	0.00
74	Di-n-butylphthalate	40.000	44.098	-10.2	98	0.00
75 C	Fluoranthene	40.000	43.184	-8.0	98	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	96	0.00
77	Benzydine	40.000	37.485	6.3	88	0.00
78	Pyrene	40.000	40.052	-0.1	96	0.00
79 S	Terphenyl-d14	80.000	83.916	-4.9	104	0.00
80	Butylbenzylphthalate	40.000	40.487	-1.2	97	0.00
81	Benzo(a)anthracene	40.000	43.275	-8.2	106	0.00
82	3,3'-Dichlorobenzidine	40.000	43.075	-7.7	104	0.00
83	Chrysene	40.000	42.741	-6.9	105	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	43.027	-7.6	105	0.00
85 c	Di-n-octyl phthalate	40.000	45.543	-13.9	113	0.00
86 I	Perylene-d12	20.000	20.000	0.0	91	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	42.500	-6.3	93	0.02
88	Benzo(b)fluoranthene	40.000	44.214	-10.5	105	0.00
89	Benzo(k)fluoranthene	40.000	45.472	-13.7	104	0.00
90 C	Benzo(a)pyrene	40.000	44.511	-11.3	102	0.00
91	Dibenzo(a,h)anthracene	40.000	42.603	-6.5	94	0.01
92	Benzo(g,h,i)perylene	40.000	41.320	-3.3	90	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF030722\
Data File : BF127224.D
Acq On : 07 Mar 2022 16:12
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Mar 08 00:02:58 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF021622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Mar 05 02:16:58 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0