

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF030724\
 Data File : BF137526.D
 Acq On : 07 Mar 2024 13:11
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 07 13:31:53 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF022924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 06 00:07:28 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	119	0.00
2	1,4-Dioxane	0.722	0.708	1.9	114	0.00
3	Pyridine	1.741	1.810	-4.0	113	0.00
4	n-Nitrosodimethylamine	0.652	0.701	-7.5	119	0.00
5 S	2-Fluorophenol	1.321	1.329	-0.6	115	0.00
6	Aniline	2.329	2.389	-2.6	116	0.00
7 S	Phenol-d6	1.906	1.857	2.6	115	0.00
8	2-Chlorophenol	1.393	1.381	0.9	116	0.00
9	Benzaldehyde	0.934	0.996	-6.6	119	0.00
10 C	Phenol	2.052	2.024	1.4	114	0.00
11	bis(2-Chloroethyl)ether	1.503	1.482	1.4	112	0.00
12	1,3-Dichlorobenzene	1.488	1.482	0.4	118	0.00
13 C	1,4-Dichlorobenzene	1.523	1.484	2.6	115	0.00
14	1,2-Dichlorobenzene	1.398	1.323	5.4	113	0.00
15	Benzyl Alcohol	1.187	1.247	-5.1	117	0.00
16	2,2'-oxybis(1-Chloropropane	2.584	2.468	4.5	113	0.00
17	2-Methylphenol	1.303	1.317	-1.1	119	0.00
18	Hexachloroethane	0.501	0.513	-2.4	119	0.00
19 P	n-Nitroso-di-n-propylamine	1.177	1.133	3.7	117	0.00
20	3+4-Methylphenols	1.493	1.451	2.8	115	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	119	0.00
22	Acetophenone	0.484	0.463	4.3	115	0.00
23 S	Nitrobenzene-d5	0.430	0.426	0.9	115	0.00
24	Nitrobenzene	0.432	0.438	-1.4	119	0.00
25	Isophorone	0.818	0.813	0.6	118	0.00
26 C	2-Nitrophenol	0.172	0.192	-11.6	128	0.00
27	2,4-Dimethylphenol	0.321	0.318	0.9	116	0.00
28	bis(2-Chloroethoxy)methane	0.480	0.486	-1.3	119	0.00
29 C	2,4-Dichlorophenol	0.294	0.303	-3.1	119	0.00
30	1,2,4-Trichlorobenzene	0.310	0.310	0.0	119	0.00
31	Naphthalene	1.073	1.024	4.6	115	0.00
32	Benzoic acid	0.167	0.173	-3.6	125	0.00
33	4-Chloroaniline	0.421	0.434	-3.1	116	0.00
34 C	Hexachlorobutadiene	0.183	0.182	0.5	118	0.00
35	Caprolactam	0.100	0.107	-7.0	125	0.00
36 C	4-Chloro-3-methylphenol	0.329	0.332	-0.9	117	0.00
37	2-Methylnaphthalene	0.684	0.659	3.7	115	0.00
38	1-Methylnaphthalene	0.644	0.627	2.6	117	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	120	0.00
40	1,2,4,5-Tetrachlorobenzene	0.558	0.554	0.7	118	0.00
41 P	Hexachlorocyclopentadiene	0.180	0.339	-88.3#	213#	0.00
42 S	2,4,6-Tribromophenol	0.176	0.191	-8.5	130	0.00
43 C	2,4,6-Trichlorophenol	0.364	0.377	-3.6	120	0.00
44	2,4,5-Trichlorophenol	0.419	0.430	-2.6	120	0.00
45 S	2-Fluorobiphenyl	1.278	1.164	8.9	112	0.00
46	1,1'-Biphenyl	1.464	1.410	3.7	115	0.00
47	2-Chloronaphthalene	1.116	1.089	2.4	116	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF030724\
 Data File : BF137526.D
 Acq On : 07 Mar 2024 13:11
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 07 13:31:53 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF022924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 06 00:07:28 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.379	0.419	-10.6	124	0.00
49	Acenaphthylene	1.786	1.719	3.8	114	0.00
50	Dimethylphthalate	1.392	1.399	-0.5	121	0.00
51	2,6-Dinitrotoluene	0.291	0.309	-6.2	123	0.00
52 C	Acenaphthene	1.066	1.019	4.4	115	0.00
53	3-Nitroaniline	0.308	0.334	-8.4	121	0.00
54 P	2,4-Dinitrophenol	0.111	0.136	-22.5	134	0.00
55	Dibenzofuran	1.624	1.536	5.4	114	0.00
56 P	4-Nitrophenol	0.190	0.218	-14.7	126	0.00
57	2,4-Dinitrotoluene	0.358	0.382	-6.7	121	0.00
58	Fluorene	1.247	1.162	6.8	115	0.00
59	2,3,4,6-Tetrachlorophenol	0.336	0.345	-2.7	121	0.00
60	Diethylphthalate	1.315	1.313	0.2	120	0.00
61	4-Chlorophenyl-phenylether	0.611	0.578	5.4	117	0.00
62	4-Nitroaniline	0.265	0.295	-11.3	129	0.00
63	Azobenzene	1.514	1.486	1.8	117	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	122	0.00
65	4,6-Dinitro-2-methylphenol	0.111	0.128	-15.3	133	0.00
66 c	n-Nitrosodiphenylamine	0.644	0.629	2.3	121	0.00
67	4-Bromophenyl-phenylether	0.218	0.220	-0.9	122	0.00
68	Hexachlorobenzene	0.220	0.222	-0.9	125	0.00
69	Atrazine	0.196	0.205	-4.6	127	0.00
70 C	Pentachlorophenol	0.133	0.143	-7.5	126	0.00
71	Phenanthrene	1.088	1.019	6.3	118	0.00
72	Anthracene	1.118	1.065	4.7	120	0.00
73	Carbazole	0.958	0.926	3.3	119	0.00
74	Di-n-butylphthalate	1.139	1.158	-1.7	124	0.00
75 C	Fluoranthene	1.064	1.023	3.9	121	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	124	0.00
77	Benzidine	0.489	0.532	-8.8	117	0.00
78	Pyrene	1.962	1.986	-1.2	120	0.00
79 S	Terphenyl-d14	1.454	1.478	-1.7	123	0.00
80	Butylbenzylphthalate	0.501	0.662	-32.1#	156#	0.00
81	Benzo(a)anthracene	1.402	1.364	2.7	120	0.00
82	3,3'-Dichlorobenzidine	0.373	0.433	-16.1	129	0.00
83	Chrysene	1.417	1.388	2.0	119	0.00
84	Bis(2-ethylhexyl)phthalate	0.637	0.771	-21.0	142	0.00
85 c	Di-n-octyl phthalate	1.120	1.352	-20.7#	141	0.00
86 I	Perylene-d12	1.000	1.000	0.0	127	0.00
87	Indeno(1,2,3-cd)pyrene	1.316	1.394	-5.9	127	0.00
88	Benzo(b)fluoranthene	1.183	1.166	1.4	123	0.00
89	Benzo(k)fluoranthene	1.271	1.197	5.8	123	0.00
90 C	Benzo(a)pyrene	1.166	1.168	-0.2	123	0.00
91	Dibenzo(a,h)anthracene	1.068	1.129	-5.7	127	0.00
92	Benzo(g,h,i)perylene	1.096	1.173	-7.0	128	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF030724\
Data File : BF137526.D
Acq On : 07 Mar 2024 13:11
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Mar 07 13:31:53 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF022924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 06 00:07:28 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 1