

Data Path : Z:\HPCHEM1\BNA_F\Data\BF030816\
 Data File : BF085278.D
 Acq On : 9 Mar 2016 1:51
 Operator : UM/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 09 02:55:22 2016
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF030316.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 04 00:54:58 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	70	-0.02
2	1,4-Dioxane	0.610	0.550	9.8	66	-0.03
3	Pyridine	1.526	1.303	14.6	62	-0.03
4	n-Nitrosodimethylamine	0.653	0.596	8.7	63	-0.05
5 S	2-Fluorophenol	1.192	1.088	8.7	71	-0.02
6	Aniline	2.189	2.085	4.8	66	-0.02
7 S	Phenol-d6	1.562	1.423	8.9	68	-0.02
8	2-Chlorophenol	1.435	1.386	3.4	70	-0.02
9	Benzaldehyde	0.804	0.660	17.9	66	-0.02
10 C	Phenol	1.673	1.536	8.2	69	-0.02
11	bis(2-Chloroethyl)ether	1.390	1.402	-0.9	68	-0.02
12	1,3-Dichlorobenzene	1.541	1.501	2.6	70	-0.02
13 C	1,4-Dichlorobenzene	1.545	1.499	3.0	74	-0.01
14	1,2-Dichlorobenzene	1.401	1.408	-0.5	69	-0.02
15	Benzyl Alcohol	1.147	1.025	10.6	65	-0.02
16	2,2'-oxybis(1-Chloropropane	2.003	1.783	11.0	66	-0.02
17	2-Methylphenol	1.171	1.179	-0.7	69	-0.02
18	Hexachloroethane	0.570	0.545	4.4	67	-0.02
19 P	n-Nitroso-di-n-propylamine	1.018	0.893	12.3	62	-0.03
20	3+4-Methylphenols	1.387	1.319	4.9	73	-0.02
21 I	Naphthalene-d8	1.000	1.000	0.0	71	-0.01
22	Acetophenone	0.488	0.483	1.0	70	-0.02
23 S	Nitrobenzene-d5	0.374	0.365	2.4	67	-0.02
24	Nitrobenzene	0.380	0.365	3.9	67	-0.02
25	Isophorone	0.693	0.665	4.0	67	-0.02
26 C	2-Nitrophenol	0.185	0.203	-9.7	74	-0.02
27	2,4-Dimethylphenol	0.338	0.352	-4.1	70	-0.02
28	bis(2-Chloroethoxy)methane	0.386	0.358	7.3	67	-0.02
29 C	2,4-Dichlorophenol	0.272	0.279	-2.6	71	-0.02
30	1,2,4-Trichlorobenzene	0.294	0.293	0.3	71	-0.02
31	Naphthalene	0.983	0.977	0.6	76	-0.02
32	Benzoic acid	0.224	0.263	-17.4	76	-0.03
33	4-Chloroaniline	0.398	0.397	0.3	75	-0.02
34 C	Hexachlorobutadiene	0.153	0.162	-5.9	73	-0.02
35	Caprolactam	0.089	0.097	-9.0	77	-0.02
36 C	4-Chloro-3-methylphenol	0.309	0.289	6.5	67	-0.02
37	2-Methylnaphthalene	0.631	0.648	-2.7	72	-0.02
38 I	Acenaphthene-d10	1.000	1.000	0.0	66	-0.02
39	1,2,4,5-Tetrachlorobenzene	0.572	0.620	-8.4	78	-0.01
40 P	Hexachlorocyclopentadiene	0.308	0.326	-5.8	69	-0.01
41 S	2,4,6-Tribromophenol	0.171	0.168	1.8	63	-0.02
42 C	2,4,6-Trichlorophenol	0.426	0.471	-10.6	73	-0.02
43	2,4,5-Trichlorophenol	0.404	0.426	-5.4	69	-0.02
44 S	2-Fluorobiphenyl	1.315	1.337	-1.7	72	-0.02

Data Path : Z:\HPCHEM1\BNA_F\Data\BF030816\
 Data File : BF085278.D
 Acq On : 9 Mar 2016 1:51
 Operator : UM/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 09 02:55:22 2016
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF030316.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 04 00:54:58 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.709	1.852	-8.4	81	-0.01
46	2-Chloronaphthalene	1.302	1.391	-6.8	73	-0.02
47	2-Nitroaniline	0.404	0.443	-9.7	71	-0.01
48	Acenaphthylene	2.027	1.998	1.4	69	-0.02
49	Dimethylphthalate	1.535	1.634	-6.4	70	-0.02
50	2,6-Dinitrotoluene	0.328	0.328	0.0	64	-0.02
51 C	Acenaphthene	1.231	1.220	0.9	70	-0.02
52	3-Nitroaniline	0.393	0.409	-4.1	63	-0.02
53 P	2,4-Dinitrophenol	0.137	0.201	-46.7#	93	-0.01
54	Dibenzofuran	1.613	1.545	4.2	65	-0.02
55 P	4-Nitrophenol	0.309	0.315	-1.9	61	-0.02
56	2,4-Dinitrotoluene	0.420	0.409	2.6	64	-0.02
57	Fluorene	1.267	1.197	5.5	63	-0.02
58	2,3,4,6-Tetrachlorophenol	0.284	0.300	-5.6	67	-0.02
59	Diethylphthalate	1.490	1.482	0.5	67	-0.02
60	4-Chlorophenyl-phenylether	0.616	0.577	6.3	65	-0.02
61	4-Nitroaniline	0.367	0.368	-0.3	64	-0.01
62	Azobenzene	1.468	1.316	10.4	59	-0.02
63 I	Phenanthrene-d10	1.000	1.000	0.0	65	-0.02
64	4,6-Dinitro-2-methylphenol	0.120	0.133	-10.8	64	-0.02
65 c	n-Nitrosodiphenylamine	0.622	0.641	-3.1	71	-0.01
66	4-Bromophenyl-phenylether	0.194	0.187	3.6	61	-0.02
67	Hexachlorobenzene	0.207	0.204	1.4	62	-0.02
68	Atrazine	0.173	0.190	-9.8	83	-0.01
69 C	Pentachlorophenol	0.115	0.143	-24.3#	80	-0.01
70	Phenanthrene	1.048	0.969	7.5	67	-0.02
71	Anthracene	1.113	1.128	-1.3	71	-0.02
72	Carbazole	0.976	0.867	11.2	59	-0.02
73	Di-n-butylphthalate	1.233	1.224	0.7	69	-0.01
74 C	Fluoranthene	1.052	0.989	6.0	64	-0.02
75 I	Chrysene-d12	1.000	1.000	0.0	85	-0.02
76	Benzidine	0.595	0.512	13.9	62	-0.02
77	Pyrene	1.505	1.285	14.6	65	-0.02
78 S	Terphenyl-d14	0.862	0.845	2.0	81	-0.01
79	Butylbenzylphthalate	0.683	0.674	1.3	78	-0.01
80	Benzo(a)anthracene	1.197	1.055	11.9	73	-0.01
81	3,3'-Dichlorobenzidine	0.378	0.353	6.6	73	-0.02
82	Chrysene	1.106	0.984	11.0	67	-0.02
83	Bis(2-ethylhexyl)phthalate	0.899	0.831	7.6	75	-0.01
84 c	Di-n-octyl phthalate	1.369	1.273	7.0	71	-0.02
85	Indeno(1,2,3-cd)pyrene	0.863	0.909	-5.3	75	-0.05
86 I	Perylene-d12	1.000	1.000	0.0	76	-0.02
87	Benzo(b)fluoranthene	1.333	1.389	-4.2	75	-0.02

Data Path : Z:\HPCHEM1\BNA_F\Data\BF030816\
Data File : BF085278.D
Acq On : 9 Mar 2016 1:51
Operator : UM/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Mar 09 02:55:22 2016
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF030316.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 04 00:54:58 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.075	0.932	13.3	75	-0.02
89 C	Benzo(a)pyrene	1.105	1.122	-1.5	76	-0.03
90	Dibenzo(a,h)anthracene	0.943	1.034	-9.7	75	-0.03
91	Benzo(g,h,i)perylene	0.948	1.035	-9.2	74	-0.05

(#) = Out of Range

SPCC's out = 0 CCC's out = 1