

Data Path : Z:\HPCHEM1\BNA F\DATA\BF030817\
 Data File : BF093444.D
 Acq On : 8 Mar 2017 22:29
 Operator : SJ/MA
 Sample : I2126-02
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 E2-Z7-WSW

Manual Integrations
 APPROVED

mohammad
 3/10/2017 8:12:02 AM

Quant Time: Mar 09 00:30:22 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF030717.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 07 15:55:22 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.76	152	180408m	20.00	ng	-0.01
21) Naphthalene-d8	8.05	136	720602m	20.00	ng	-0.01
38) Acenaphthene-d10	9.80	164	332155m	20.00	ng	-0.01
63) Phenanthrene-d10	11.28	188	625194	20.00	ng	0.00
75) Chrysene-d12	13.93	240	474884	20.00	ng	0.01
86) Perylene-d12	15.34	264	327359	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.36	112	1122077	103.52	ng	0.01
7) Phenol-d6	6.41	99	1409270	103.10	ng	0.00
23) Nitrobenzene-d5	7.34	82	933401	71.87	ng	0.00
41) 2,4,6-Tribromophenol	10.60	330	244472	112.97	ng	0.00
44) 2-Fluorobiphenyl	9.12	172	1341703	75.13	ng	-0.01
78) Terphenyl-d14	12.87	244	1057595	57.90	ng	0.00
Target Compounds						
10) Phenol	6.44	94	151846	9.07	ng	Qvalue # 70
17) 2-Methylphenol	7.04	107	167644	16.32	ng	93
20) 3+4-Methylphenols	7.20	107	311477m	23.38	ng	
27) 2,4-Dimethylphenol	7.73	122	243952	22.52	ng	95
31) Naphthalene	8.07	128	86637	2.40	ng	# 83
49) Dimethylphthalate	9.52	163	321589	14.60	ng	# 98
70) Phenanthrene	11.32	178	77774	2.18	ng	98
74) Fluoranthene	12.51	202	76620	2.22	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF030817\
Data File : BF093444.D
Acq On : 8 Mar 2017 22:29
Operator : SJ/MA
Sample : I2126-02
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
E2-Z7-WSW

Manual Integrations
APPROVED

mohammad
3/10/2017 8:12:02 AM

Quant Time: Mar 09 00:30:22 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF030717.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 07 15:55:22 2017
Response via : Initial Calibration

