

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF030824\
 Data File : BF137540.D
 Acq On : 08 Mar 2024 13:35
 Operator : CG\JU
 Sample : PB159511BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB159511BL

Quant Time: Mar 08 22:15:19 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF022924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 08 22:10:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.887	152	240702	20.000	ng	0.00
21) Naphthalene-d8	8.157	136	944930	20.000	ng	0.00
39) Acenaphthene-d10	9.916	164	526592	20.000	ng	0.00
64) Phenanthrene-d10	11.410	188	970195	20.000	ng	0.00
76) Chrysene-d12	14.068	240	605851	20.000	ng	0.00
86) Perylene-d12	15.615	264	491195	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.557	112	1815976	114.264	ng	0.01
7) Phenol-d6	6.534	99	2497855	108.883	ng	0.00
23) Nitrobenzene-d5	7.445	82	1556041	76.514	ng	0.00
42) 2,4,6-Tribromophenol	10.710	330	602433	130.273	ng	0.00
45) 2-Fluorobiphenyl	9.239	172	2499159	74.291	ng	0.00
79) Terphenyl-d14	13.016	244	3099564	70.351	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF030824\
Data File : BF137540.D
Acq On : 08 Mar 2024 13:35
Operator : CG\JU
Sample : PB159511BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB159511BL

Quant Time: Mar 08 22:15:19 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF022924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 08 22:10:14 2024
Response via : Initial Calibration

