

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF030916\
 Data File : BF085312.D
 Acq On : 9 Mar 2016 21:34
 Operator : UM/SJ
 Sample : H1724-05
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 UST-SW-NW

Quant Time: Mar 10 00:59:31 2016
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF030316.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 04 00:54:58 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.95	152	151414	20.00	ng	-0.03
21) Naphthalene-d8	8.23	136	657907	20.00	ng	-0.03
38) Acenaphthene-d10	9.99	164	277950	20.00	ng	-0.03
63) Phenanthrene-d10	11.48	188	554217	20.00	ng	-0.03
75) Chrysene-d12	14.11	240	347667	20.00	ng	-0.05
86) Perylene-d12	15.57	264	230249	20.00	ng	-0.05
System Monitoring Compounds						
5) 2-Fluorophenol	5.56	112	1201222	133.07	ng	-0.02
7) Phenol-d6	6.57	99	1596573	135.00	ng	-0.03
23) Nitrobenzene-d5	7.51	82	1100337	89.50	ng	-0.03
41) 2,4,6-Tribromophenol	10.78	330	375077	157.71	ng	-0.03
44) 2-Fluorobiphenyl	9.30	172	1479102	80.93	ng	-0.05
78) Terphenyl-d14	13.07	244	1530372	102.18	ng	-0.02
Target Compounds						Qvalue
49) Dimethylphthalate	9.70	163	411296	19.28	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF030916\
Data File : BF085312.D
Acq On : 9 Mar 2016 21:34
Operator : UM/SJ
Sample : H1724-05
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
UST-SW-NW

Quant Time: Mar 10 00:59:31 2016
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF030316.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 04 00:54:58 2016
Response via : Initial Calibration

