

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF030916\
 Data File : BF085321.D
 Acq On : 10 Mar 2016 1:52
 Operator : UM/SJ
 Sample : H1722-02
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 W170TH-SB-34(0.5-2.0)

Manual Integrations
 APPROVED

UMANGI
 3/10/2016 11:17:28 AM

Quant Time: Mar 10 02:41:57 2016
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF030316.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 04 00:54:58 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.94	152	168572	20.00	ng	-0.05
21) Naphthalene-d8	8.23	136	753792	20.00	ng	-0.03
38) Acenaphthene-d10	9.99	164	347683	20.00	ng	-0.03
63) Phenanthrene-d10	11.48	188	660953	20.00	ng	-0.03
75) Chrysene-d12	14.11	240	354276	20.00	ng	-0.05
86) Perylene-d12	15.57	264	257576	20.00	ng	-0.05
System Monitoring Compounds						
5) 2-Fluorophenol	5.56	112	1315677	130.92	ng	-0.02
7) Phenol-d6	6.57	99	1758089	133.53	ng	-0.03
23) Nitrobenzene-d5	7.51	82	1194199	84.78	ng	-0.03
41) 2,4,6-Tribromophenol	10.78	330	446535	150.10	ng	-0.03
44) 2-Fluorobiphenyl	9.30	172	1764754	77.19	ng	-0.05
78) Terphenyl-d14	13.07	244	1579060	103.47	ng	-0.02
Target Compounds						Qvalue
49) Dimethylphthalate	9.70	163	489902	18.36	ng	98
74) Fluoranthene	12.69	202	95322	2.74	ng	93
77) Pyrene	12.92	202	85549	3.21	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF030916\
Data File : BF085321.D
Acq On : 10 Mar 2016 1:52
Operator : UM/SJ
Sample : H1722-02
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
W170TH-SB-34(0.5-2.0)

Manual Integrations
APPROVED

UMANGI
3/10/2016 11:17:28 AM

Quant Time: Mar 10 02:41:57 2016
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF030316.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 04 00:54:58 2016
Response via : Initial Calibration

