


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF030916\
 Data File : BF085319.D
 Acq On : 10 Mar 2016 00:55
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF030316.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Mar 04 00:54:58 2016



AutoFind: Scans 825, 826, 827; Background Corrected with Scan 822

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	35.1	19260	PASS
68	69	0.00	2	1.6	315	PASS
69	198	0.00	100	36.4	19954	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	52.5	28786	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	54807	PASS
199	198	5	9	6.6	3616	PASS
275	198	10	60	23.6	12949	PASS
365	198	1	100	2.7	1504	PASS
441	442	0.01	24	15.3	6935	PASS
442	198	50	100	83.0	45468	PASS
443	442	15	24	18.6	8448	PASS