

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF030917\
 Data File : BF093483.D
 Acq On : 9 Mar 2017 18:54
 Operator : SJ/MA
 Sample : PB97463BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB97463BL

Integration Parameters: rteint.p
 Integrator: RTE
 Smoothing : ON Filtering: 5
 Sampling : 1 Min Area: 3 % of largest Peak
 Start Thrs: 0.2 Max Peaks: 100
 Stop Thrs : 0 Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >
 Peak separation: 5

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF030717.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

Signal : TIC

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	1.978	6	8	20	rVB2	41541	122148	2.23%	0.257%
2	4.916	262	265	280	rBV	808048	1527761	27.93%	3.218%
3	5.350	301	303	319	rBV	4568458	5042003	92.17%	10.621%
4	5.750	335	338	346	rVB	64469	77435	1.42%	0.163%
5	6.402	393	395	403	rBV	4390057	4745779	86.75%	9.997%
6	6.516	403	405	408	rBV	4284654	4504339	82.34%	9.488%
7	6.744	423	425	428	rBV	1345006	1227867	22.45%	2.586%
8	6.904	436	439	441	rBV	3744077	3917650	71.62%	8.252%
9	7.316	473	475	478	rBV	2685813	3183739	58.20%	6.706%
10	8.036	535	538	540	rBV	1451093	1594241	29.14%	3.358%
11	8.447	567	574	584	rVB4	12598	57510	1.05%	0.121%
12	9.110	629	632	634	rBV	5063725	5336836	97.56%	11.242%
13	9.739	678	687	689	rBV	150790	417897	7.64%	0.880%
14	9.785	689	691	693	rVB	1875715	1809182	33.07%	3.811%
15	10.573	757	760	763	rBV	2453881	2905218	53.11%	6.120%
16	11.271	818	821	823	rBV	1929717	1878341	34.34%	3.957%
17	12.859	957	960	962	rBV	4369470	5470376	100.00%	11.523%
18	13.877	1046	1049	1050	rBV	102720	117926	2.16%	0.248%
19	13.911	1050	1052	1054	rVB	1882162	1827110	33.40%	3.849%
20	15.317	1172	1175	1178	rBV	1378073	1710295	31.26%	3.603%


Sum of corrected areas: 47473653

Data Path : Z:\HPCHEM1\BNA F\DATA\BF030917\
Data File : BF093483.D
Acq On : 9 Mar 2017 18:54
Operator : SJ/MA
Sample : PB97463BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB97463BL

Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF030717.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

TIC Library : C:\DATABASE\NIST02.L
TIC Integration Parameters: LSCINT.P



Data Path : Z:\HPCHEM1\BNA_F\DATA\BF030917\
Data File : BF093483.D
Acq On : 9 Mar 2017 18:54
Operator : SJ/MA
Sample : PB97463BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB97463BL










Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF030717.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

TIC Library : C:\DATABASE\NIST02.L
TIC Integration Parameters: LSCINT.P

Peak Number 1 2-Pentanone, 4-hydroxy-4-me... Concentration Rank 3

R.T.	EstConc	Area	Relative to ISTD	R.T.
4.92	24.88 ng	1527760	1,4-Dichlorobenzene-d4	6.74

Hit#	of	Tentative ID	MW	MolForm	CAS#	Qual
1	5	2-Pentanone, 4-hydroxy-4-methyl-	116	C6H12O2	000123-42-2	50
2		Propane, 2-methyl-2-(1-methylethoxy)-	116	C7H16O	017348-59-3	36
3		3-Hexanol, 4-methyl-	116	C7H16O	000615-29-2	33
4		2-Hexanol, 2-methyl-	116	C7H16O	000625-23-0	28
5		Acetic acid, cyano-, 1,1-dimethyl-	141	C7H11NO2	001116-98-9	23



Data Path : Z:\HPCHEM1\BNA F\DATA\BF030917\
Data File : BF093483.D
Acq On : 9 Mar 2017 18:54
Operator : SJ/MA
Sample : PB97463BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB97463BL










Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF030717.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

TIC Library : C:\DATABASE\NIST02.L
TIC Integration Parameters: LSCINT.P

Peak Number 2 unknown6.52 Concentration Rank 1

R.T.	EstConc	Area	Relative to ISTD	R.T.
6.52	73.37 ng	4504340	1,4-Dichlorobenzene-d4	6.74

Hit#	of	Tentative ID	MW	MolForm	CAS#	Qual
1	5	5-Aminoindole	132	C8H8N2	005192-03-0	27
2		5-Fluoro-2-chloropyrimidine	132	C4H2ClFN2	062802-42-0	12
3		(5-METHYL-2-PYRIDYL)ACETONITRILE	132	C8H8N2	1000241-93-9	10
4		Tranylcypromine-propionyl	189	C12H15NO	1000123-86-3	10
5		1,2,3-Trifluorobenzene	132	C6H3F3	001489-53-8	9



Data Path : Z:\HPCHEM1\BNA_F\DATA\BF030917\
Data File : BF093483.D
Acq On : 9 Mar 2017 18:54
Operator : SJ/MA
Sample : PB97463BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB97463BL

Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF030717.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

TIC Library : C:\DATABASE\NIST02.L
TIC Integration Parameters: LSCINT.P

TIC Top Hit name	RT	EstConc	Units	Response	--Internal Standard--			
					#	RT	Resp	Conc
2-Pentanone, 4-hy...	4.92	24.9	ng	1527760	1	6.74	1227870	20.0
unknown6.52	6.52	73.4	ng	4504340	1	6.74	1227870	20.0