


Benzaldehyde

Response Ratio



$$\text{Response} = 6.474\text{e-}001 * \text{Amt} + 3.418\text{e-}001$$


Coef of Det (r^2) = 0.991060 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA F\Methods\8270-BF030922.M

Calibration Table Last Updated: Wed Mar 09 17:59:49 2022

Benzoic acid

Response Ratio



$$\text{Response} = 2.292\text{e-}001 * \text{Amt} - 4.697\text{e-}002$$


Coef of Det (r^2) = 0.999013 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA F\Methods\8270-BF030922.M

Calibration Table Last Updated: Wed Mar 09 17:59:49 2022

Hexachlorocyclopentadiene

Response Ratio



$$\text{Response} = 3.029\text{e-}001 * \text{Amt} - 9.811\text{e-}002$$

Coef of Det (r^2) = 0.995113 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA F\Methods\8270-BF030922.M

Calibration Table Last Updated: Wed Mar 09 17:59:49 2022