

Data Path : Z:\HPCHEM1\BNA F\DATA\BF031015\
 Data File : BF077670.D
 Acq On : 10 Mar 2015 19:35
 Operator : TP/IZ
 Sample : G1440-02
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 B2(8-9)

Manual Integrations
 APPROVED

mohammad
 3/12/2015 2:21:54 PM

Quant Time: Mar 11 01:55:58 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF021915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 11 00:54:19 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.18	152	50299	20.00	ng	0.00
21) Naphthalene-d8	8.76	136	183788	20.00	ng	0.00
38) Acenaphthene-d10	10.93	164	91075	20.00	ng	0.00
63) Phenanthrene-d10	12.77	188	172681	20.00	ng	0.00
75) Chrysene-d12	16.08	240	222409	20.00	ng	0.00
86) Perylene-d12	17.94	264	220612	20.00	ng	0.03
System Monitoring Compounds						
5) 2-Fluorophenol	5.47	112	267026	88.63	ng	0.00
7) Phenol-d6	6.75	99	351324	90.69	ng	0.00
23) Nitrobenzene-d5	7.87	82	176065	59.57	ng	0.00
41) 2,4,6-Tribromophenol	11.91	330	73997	84.38	ng	0.00
44) 2-Fluorobiphenyl	10.11	172	363498	62.23	ng	0.00
78) Terphenyl-d14	14.76	244	461264	48.48	ng	0.00
Target Compounds						
10) Phenol	6.76	94	10211	2.22	ng	94
37) 2-Methylnaphthalene	9.65	142	231560	35.48	ng	# 96
49) Dimethylphthalate	10.61	163	17964	2.81	ng	# 95
51) Acenaphthene	10.97	154	13137	2.57	ng	# 88
57) Fluorene	11.60	166	49447	7.84	ng	# 84
70) Phenanthrene	12.80	178	103904	10.38	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF031015\
Data File : BF077670.D
Acq On : 10 Mar 2015 19:35
Operator : TP/IZ
Sample : G1440-02
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
B2(8-9)

Manual Integrations
APPROVED

mohammad
3/12/2015 2:21:54 PM

Quant Time: Mar 11 01:55:58 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF021915.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 11 00:54:19 2015
Response via : Initial Calibration

