

Data Path : Z:\HPCHEM1\BNA F\DATA\BF031015\
 Data File : BF077676.D
 Acq On : 10 Mar 2015 22:26
 Operator : TP/IZ
 Sample : G1466-04
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DCW-WW-085-3915

Quant Time: Mar 11 11:21:50 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF021915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 11 00:54:19 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.18	152	60714	20.00	ng	0.00
21) Naphthalene-d8	8.76	136	239715	20.00	ng	0.00
38) Acenaphthene-d10	10.93	164	121421	20.00	ng	0.00
63) Phenanthrene-d10	12.77	188	230975	20.00	ng	0.00
75) Chrysene-d12	16.05	240	227260	20.00	ng	-0.04
86) Perylene-d12	17.75	264	229516	20.00	ng	-0.16
System Monitoring Compounds						
5) 2-Fluorophenol	5.47	112	204767	56.30	ng	0.00
7) Phenol-d6	6.75	99	151052	32.30	ng	0.00
23) Nitrobenzene-d5	7.88	82	387286	100.47	ng	0.00
41) 2,4,6-Tribromophenol	11.91	330	164209	140.45	ng	0.00
44) 2-Fluorobiphenyl	10.11	172	846480	108.70	ng	0.00
78) Terphenyl-d14	14.76	244	906687	93.27	ng	0.00
Target Compounds						Qvalue
10) Phenol	6.76	94	12433	2.24	ng	79
13) 1,4-Dichlorobenzene	7.21	146	32000	6.73	ng	99
14) 1,2-Dichlorobenzene	7.39	146	510828	112.99	ng	100
30) 1,2,4-Trichlorobenzene	8.69	180	228375	59.50	ng	99
31) Naphthalene	8.79	128	1102715	91.99	ng	99
33) 4-Chloroaniline	8.84	127	171541	32.18	ng	98
37) 2-Methylnaphthalene	9.64	142	273895	32.17	ng	98
45) 1,1'-Biphenyl	10.23	154	23035	2.50	ng	99
49) Dimethylphthalate	10.61	163	23761	2.78	ng	# 95
51) Acenaphthene	10.97	154	71342	10.47	ng	99
54) Dibenzofuran	11.18	168	33439	3.18	ng	99
56) 2,4-Dinitrotoluene	11.18	165	28518	12.33	ng	# 58
57) Fluorene	11.61	166	81570	9.70	ng	98
70) Phenanthrene	12.79	178	55605	4.15	ng	99
72) Carbazole	13.07	167	38880	3.08	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF031015\
Data File : BF077676.D
Acq On : 10 Mar 2015 22:26
Operator : TP/IZ
Sample : G1466-04
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DCW-WW-085-3915

Quant Time: Mar 11 11:21:50 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF021915.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 11 00:54:19 2015
Response via : Initial Calibration

