

Data Path : S:\HPCHEM1\BNA_F\DATA\BF031016\
 Data File : BF085360.D
 Acq On : 11 Mar 2016 4:08
 Operator : UM/SJ
 Sample : H1731-02
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-18-COMP

Quant Time: Mar 11 04:56:21 2016
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF030316.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 11 02:49:26 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.94	152	122617	20.00	ng	0.00
21) Naphthalene-d8	8.22	136	426678	20.00	ng	-0.01
38) Acenaphthene-d10	9.98	164	204526	20.00	ng	0.00
63) Phenanthrene-d10	11.46	188	354290	20.00	ng	0.00
75) Chrysene-d12	14.09	240	230352	20.00	ng	-0.01
86) Perylene-d12	15.56	264	223315	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.53	112	1189689	162.75	ng	-0.01
7) Phenol-d6	6.56	99	1503692	157.01	ng	-0.01
23) Nitrobenzene-d5	7.50	82	743747	93.28	ng	-0.01
41) 2,4,6-Tribromophenol	10.77	330	244044	139.45	ng	0.00
44) 2-Fluorobiphenyl	9.30	172	1494653	111.14	ng	0.00
78) Terphenyl-d14	13.05	244	1003972	101.18	ng	0.00
Target Compounds						Qvalue
49) Dimethylphthalate	9.69	163	290092	18.48	ng	99
74) Fluoranthene	12.68	202	55930	3.00	ng	94
77) Pyrene	12.90	202	44899	2.59	ng	95
83) Bis(2-ethylhexyl)phthalate	14.08	149	23135	2.23	ng	# 99
84) Di-n-octyl phthalate	14.72	149	52243	3.31	ng	# 64

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : S:\HPCHEM1\BNA_F\DATA\BF031016\
Data File : BF085360.D
Acq On : 11 Mar 2016 4:08
Operator : UM/SJ
Sample : H1731-02
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-18-COMP

Quant Time: Mar 11 04:56:21 2016
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF030316.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 11 02:49:26 2016
Response via : Initial Calibration

