

Data Path : S:\HPCHEM1\BNA\_F\DATA\BF031016\  
 Data File : BF085362.D  
 Acq On : 11 Mar 2016 5:05  
 Operator : UM/SJ  
 Sample : H1731-06  
 Misc :  
 ALS Vial : 32 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 SB-20-COMP

Quant Time: Mar 11 05:38:33 2016  
 Quant Method : Z:\HPCHEM1\BNA\_F\METHODS\8270-BF030316.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Fri Mar 11 02:49:26 2016  
 Response via : Initial Calibration

| Internal Standards             | R.T.  | QIon | Response | Conc   | Units | Dev(Min) |
|--------------------------------|-------|------|----------|--------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4      | 6.94  | 152  | 158119   | 20.00  | ng    | 0.00     |
| 21) Naphthalene-d8             | 8.22  | 136  | 551623   | 20.00  | ng    | -0.01    |
| 38) Acenaphthene-d10           | 9.98  | 164  | 256305   | 20.00  | ng    | 0.00     |
| 63) Phenanthrene-d10           | 11.46 | 188  | 437419   | 20.00  | ng    | 0.00     |
| 75) Chrysene-d12               | 14.09 | 240  | 268720   | 20.00  | ng    | -0.01    |
| 86) Perylene-d12               | 15.56 | 264  | 268741   | 20.00  | ng    | -0.01    |
| System Monitoring Compounds    |       |      |          |        |       |          |
| 5) 2-Fluorophenol              | 5.54  | 112  | 1218484  | 129.26 | ng    | 0.00     |
| 7) Phenol-d6                   | 6.57  | 99   | 1794452  | 145.30 | ng    | 0.00     |
| 23) Nitrobenzene-d5            | 7.50  | 82   | 968385   | 93.94  | ng    | -0.01    |
| 41) 2,4,6-Tribromophenol       | 10.77 | 330  | 286200   | 130.50 | ng    | 0.00     |
| 44) 2-Fluorobiphenyl           | 9.30  | 172  | 1498609  | 88.92  | ng    | 0.00     |
| 78) Terphenyl-d14              | 13.05 | 244  | 1048602  | 90.59  | ng    | 0.00     |
| Target Compounds               |       |      |          |        |       | Qvalue   |
| 49) Dimethylphthalate          | 9.69  | 163  | 349236   | 17.75  | ng    | 99       |
| 83) Bis(2-ethylhexyl)phthalate | 14.08 | 149  | 26302    | 2.18   | ng    | # 99     |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : S:\HPCHEM1\BNA\_F\DATA\BF031016\  
Data File : BF085362.D  
Acq On : 11 Mar 2016 5:05  
Operator : UM/SJ  
Sample : H1731-06  
Misc :  
ALS Vial : 32 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
SB-20-COMP

Quant Time: Mar 11 05:38:33 2016  
Quant Method : Z:\HPCHEM1\BNA\_F\METHODS\8270-BF030316.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Fri Mar 11 02:49:26 2016  
Response via : Initial Calibration

