

Data Path : Z:\HPCHEM1\BNA F\Data\BF031017\
 Data File : BF093561.D
 Acq On : 11 Mar 2017 9:28
 Operator : SJ/MA
 Sample : I2211-10
 Misc :
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :

Quant Time: Mar 13 14:22:21 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF030717.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 07 15:55:22 2017
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.74	152	172896	20.00	ng	-0.02
21) Naphthalene-d8	8.02	136	725865	20.00	ng	-0.03
38) Acenaphthene-d10	9.77	164	325216	20.00	ng	-0.03
63) Phenanthrene-d10	11.26	188	628602	20.00	ng	-0.02
75) Chrysene-d12	13.90	240	455902	20.00	ng	-0.02
86) Perylene-d12	15.31	264	305897	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.36	112	1489446	143.38	ng	0.01
7) Phenol-d6	6.41	99	1789922	136.64	ng	0.00
23) Nitrobenzene-d5	7.32	82	1217118	93.04	ng	-0.02
41) 2,4,6-Tribromophenol	10.57	330	342123	161.47	ng	-0.02
44) 2-Fluorobiphenyl	9.10	172	1687731	96.53	ng	-0.03
78) Terphenyl-d14	12.85	244	1664953	94.95	ng	-0.02
Target Compounds						
9) Benzaldehyde	6.41	77	2190	Below Cal	#	1
19) n-Nitroso-di-n-propylamine	7.32	70	165573	18.38	ng	# 83
25) Isophorone	7.32	82	1220207	46.25	ng	# 65
50) 2,6-Dinitrotoluene	9.77	165	41052	8.68	ng	# 31
56) 2,4-Dinitrotoluene	9.77	165	41055	7.06	ng	# 20
66) 4-Bromophenyl-phenylether	10.57	248	25396	3.82	ng	# 1

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\Data\BF031017\
Data File : BF093561.D
Acq On : 11 Mar 2017 9:28
Operator : SJ/MA
Sample : I2211-10
Misc :
ALS Vial : 41 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :

Quant Time: Mar 13 14:22:21 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF030717.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 07 15:55:22 2017
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 20.00 ng

RT: 6.74 min Scan# 425

Delta R.T. -0.02 min

Lab File: BF093561.D

Acq: 11 Mar 2017 9:28

Instrument :

BNA_F

ClientSampleId :


Tot Ion:152 Resp: 172896

Ion Ratio Lower Upper

152 100

150 156.5 124.9 187.3




115 59.2 46.0 69.0



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#5

2-Fluorophenol

Concen: 143.38 ng

RT: 5.36 min Scan# 304

Delta R.T. 0.01 min

Lab File: BF093561.D

Acq: 11 Mar 2017 9:28


Tgt Ion:112 Resp: 1489446

Ion Ratio Lower Upper

112 100

64 60.8 46.1 69.1



63 30.8 23.8 35.6






Abundance Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BF0

Ion 63.00 (62.70 to 63.70): BF0





#7

Phenol-d6

Concen: 136.64 ng

RT: 6.41 min Scan# 396

Delta R.T. -0.00 min

Lab File: BF093561.D

Acq: 11 Mar 2017 9:28

Instrument :

BNA_F

ClientSampled :





Tot Ion: 99 Resp: 1789922

Ion Ratio Lower Upper

99 100

42 18.1 15.6 23.4

71 30.5 27.5 41.3



#19

n-Nitroso-di-n-propylamine

Concen: 18.38 ng

RT: 7.32 min Scan# 475

Delta R.T. 0.13 min

Lab File: BF093561.D

Acq: 11 Mar 2017 9:28

Tgt Ion: 70 Resp: 165573


Ion Ratio Lower Upper


70 100

42 53.5 49.4 74.0

101 0.0 7.4 11.2#

130 1.9 14.2 21.4#









#21
Naphthalene-d8
Concen: 20.00 ng
RT: 8.02 min Scan# 537
Delta R.T. -0.03 min
Lab File: BF093561.D
Acq: 11 Mar 2017 9:28

Instrument :
BNA_F
ClientSampleId :

Tot Ion:	136	Resp:	725865
Ion	Ratio	Lower	Upper
136	100		
137	11.0	8.4	12.6
54	11.7	4.5	6.7#
68	8.3	3.6	5.4#




Abundance Ion 136.00 (135.70 to 136.70): E
Ion 137.00 (136.70 to 137.70): E
Ion 54.00 (53.70 to 54.70): BF0
Ion 68.00 (67.70 to 68.70): BF0






#23
Nitrobenzene-d5
Concen: 93.04 ng
RT: 7.32 min Scan# 475
Delta R.T. -0.02 min
Lab File: BF093561.D
Acq: 11 Mar 2017 9:28

Tgt Ion:	82	Resp:	1217118
Ion	Ratio	Lower	Upper
82	100		
128	40.2	34.6	52.0
54	59.3	39.5	59.3#



Abundance Ion 82.00 (81.70 to 82.70): BF0
Ion 128.00 (127.70 to 128.70): E
Ion 54.00 (53.70 to 54.70): BF0





#25

Isophorone

Concen: 46.25 ng

RT: 7.32 min Scan# 475

Delta R.T. -0.29 min

Lab File: BF093561.D

Acq: 11 Mar 2017 9:28

Instrument :

BNA_F

ClientSampleId :





Tot Ion: 82 Resp: 1220207

Ion Ratio Lower Upper

82 100

95 0.0 5.6 8.4#

138 0.0 14.6 22.0#



#38

Acenaphthene-d10

Concen: 20.00 ng

RT: 9.77 min Scan# 690

Delta R.T. -0.03 min

Lab File: BF093561.D

Acq: 11 Mar 2017 9:28




Tgt Ion: 164 Resp: 325216


Ion Ratio Lower Upper

164 100

162 101.0 80.2 120.2

160 44.7 36.9 55.3









#41
 2,4,6-Tribromophenol
 Concen: 161.47 ng
 RT: 10.57 min Scan# 760
 Delta R.T. -0.02 min
 Lab File: BF093561.D
 Acq: 11 Mar 2017 9:28

Instrument :
 BNA_F
 ClientSampleId :

Tot Ion:330 Resp: 342123
 Ion Ratio Lower Upper
 330 100
 332 97.3 77.4 116.2
 141 43.7 34.1 51.1




Abundance Ion 329.80 (329.50 to 330.50): E
 Ion 331.80 (331.50 to 332.50): E
 Ion 141.00 (140.70 to 141.70): E






#44
 2-Fluorobiphenyl
 Concen: 96.53 ng
 RT: 9.10 min Scan# 631
 Delta R.T. -0.03 min
 Lab File: BF093561.D
 Acq: 11 Mar 2017 9:28

Tgt Ion:172 Resp: 1687731
 Ion Ratio Lower Upper
 172 100
 171 36.8 29.4 44.0
 170 25.0 20.2 30.4



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E









#50
 2,6-Dinitrotoluene
 Concen: 8.68 ng
 RT: 9.77 min Scan# 690
 Delta R.T. 0.17 min
 Lab File: BF093561.D
 Acq: 11 Mar 2017 9:28




Instrument :
 BNA_F
 ClientSampleId :


Tot Ion	Ratio	Lower	Upper
165	100		
63	1.3	36.2	54.4#
89	2.0	40.2	60.4#



#56
 2,4-Dinitrotoluene
 Concen: 7.06 ng
 RT: 9.77 min Scan# 690
 Delta R.T. -0.24 min
 Lab File: BF093561.D
 Acq: 11 Mar 2017 9:28

Tgt Ion	Ratio	Lower	Upper
165	100		
63	1.3	40.2	60.4#
89	2.0	62.5	93.7#
182	0.0	1.8	2.8#









#63
 Phenanthrene-d10
 Concen: 20.00 ng
 RT: 11.26 min Scan# 820
 Delta R.T. -0.02 min
 Lab File: BF093561.D
 Acq: 11 Mar 2017 9:28

Instrument :
 BNA_F
 ClientSampleId :

Tot Ion:188 Resp: 628602
 Ion Ratio Lower Upper
 188 100
 94 11.8 10.0 15.0
 80 12.7 11.2 16.8




Abundance Ion 188.00 (187.70 to 188.70): E
 Ion 94.00 (93.70 to 94.70): BF0
 Ion 80.00 (79.70 to 80.70): BF0






#66
 4-Bromophenyl-phenylether
 Concen: 3.82 ng
 RT: 10.57 min Scan# 760
 Delta R.T. -0.26 min
 Lab File: BF093561.D
 Acq: 11 Mar 2017 9:28

Tgt Ion:248 Resp: 25396
 Ion Ratio Lower Upper
 248 100
 250 208.9 76.9 115.3#
 141 419.9 68.2 102.4#



Abundance Ion 248.00 (247.70 to 248.70): E
 Ion 250.00 (249.70 to 250.70): E
 Ion 141.00 (140.70 to 141.70): E









#75
Chrysene-d12
Concen: 20.00 ng
RT: 13.90 min Scan# 1051
Delta R.T. -0.02 min
Lab File: BF093561.D
Acq: 11 Mar 2017 9:28




Instrument :
BNA_F
ClientSampled :


Tot Ion:240 Resp: 455902
Ion Ratio Lower Upper
240 100
120 15.3 12.9 19.3
236 25.7 20.9 31.3



#78
Terphenyl-d14
Concen: 94.95 ng
RT: 12.85 min Scan# 959
Delta R.T. -0.02 min
Lab File: BF093561.D
Acq: 11 Mar 2017 9:28

Tgt Ion:244 Resp: 1664953
Ion Ratio Lower Upper
244 100
212 7.9 6.2 9.2
122 11.2 11.4 17.2#






#86
Pervlene-d12
Concen: 20.00 ng
RT: 15.31 min Scan# 1174
Delta R.T. -0.02 min
Lab File: BF093561.D
Acq: 11 Mar 2017 9:28

Instrument :
BNA_F
ClientSampleId :

Tot Ion: 264 Resp: 305897
Ion Ratio Lower Upper
264 100
260 24.3 19.2 28.8
265 21.4 17.4 26.0



Abundance Ion 264.00 (263.70 to 264.70): E
Ion 260.00 (259.70 to 260.70): E
Ion 265.00 (264.70 to 265.70): E

