

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031023\
 Data File : BF132748.D
 Acq On : 10 Mar 2023 18:43
 Operator : CG\JU
 Sample : 01846-04MS
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TP-4MS

Quant Time: Mar 11 03:05:05 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF022223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 09 16:26:19 2023
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Christian
 Giraldo

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.981	152	221586	20.000	ng	0.00
21) Naphthalene-d8	8.269	136	774449	20.000	ng	0.00
39) Acenaphthene-d10	10.028	164	405578	20.000	ng	0.00
64) Phenanthrene-d10	11.527	188	686697	20.000	ng	0.00
76) Chrysene-d12	14.186	240	361410	20.000	ng	# 0.00
86) Perylene-d12	15.727	264	476900	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.616	112	1314574	96.284	ng	0.00
7) Phenol-d6	6.616	99	1666549	93.529	ng	0.00
23) Nitrobenzene-d5	7.545	82	1053977	64.569	ng	-0.01
42) 2,4,6-Tribromophenol	10.827	330	413589	94.425	ng	0.00
45) 2-Fluorobiphenyl	9.345	172	1679178	63.041	ng	0.00
79) Terphenyl-d14	13.110	244	1393462	65.189	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	2.904	88	287145	38.021	ng	95
3) Pyridine	3.663	79	737232	36.775	ng	95
4) n-Nitrosodimethylamine	3.634	42	291901	38.549	ng	85
6) Aniline	6.645	93	722377	30.125	ng	96
8) 2-Chlorophenol	6.769	128	654721	46.204	ng	96
9) Benzaldehyde	6.534	77	308457	32.081	ng	95
10) Phenol	6.628	94	840442	41.141	ng	99
11) bis(2-Chloroethyl)ether	6.710	93	678852	43.046	ng	97
12) 1,3-Dichlorobenzene	6.922	146	677558	44.557	ng	97
13) 1,4-Dichlorobenzene	6.998	146	669539	44.095	ng	99
14) 1,2-Dichlorobenzene	7.151	146	646314	44.353	ng	98
15) Benzyl Alcohol	7.122	79	593139	43.223	ng	98
16) 2,2'-oxybis(1-Chloropr...	7.251	45	817212	40.636	ng	95
17) 2-Methylphenol	7.234	107	528917	39.994	ng	97
18) Hexachloroethane	7.492	117	260982	43.995	ng	97
19) n-Nitroso-di-n-propyla...	7.392	70	414994	37.843	ng	96
20) 3+4-Methylphenols	7.387	107	595482	34.853	ng	95
22) Acetophenone	7.392	105	801905	41.064	ng	# 90
24) Nitrobenzene	7.569	77	704624	40.362	ng	100
25) Isophorone	7.804	82	1307471	41.778	ng	100
26) 2-Nitrophenol	7.881	139	339665	44.098	ng	97
27) 2,4-Dimethylphenol	7.916	122	524804	43.169	ng	98
28) bis(2-Chloroethoxy)met...	8.010	93	788454	43.067	ng	99
29) 2,4-Dichlorophenol	8.128	162	522817	43.233	ng	100
30) 1,2,4-Trichlorobenzene	8.204	180	584814	44.536	ng	100
31) Naphthalene	8.292	128	1877915	47.366	ng	99
32) Benzoic acid	8.045	122	333051m	35.823	ng	
33) 4-Chloroaniline	8.334	127	212622	12.515	ng	96
34) Hexachlorobutadiene	8.398	225	378646	45.378	ng	100
35) Caprolactam	8.722	113	171692m	43.065	ng	
36) 4-Chloro-3-methylphenol	8.828	107	563738	42.448	ng	94
37) 2-Methylnaphthalene	8.981	142	1216404	45.020	ng	98
38) 1-Methylnaphthalene	9.081	142	1099800	43.556	ng	99
40) 1,2,4,5-Tetrachloroben...	9.151	216	599589	45.185	ng	99
41) Hexachlorocyclopentadiene	9.133	237	487645	67.754	ng	100
43) 2,4,6-Trichlorophenol	9.263	196	386283	42.955	ng	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031023\
 Data File : BF132748.D
 Acq On : 10 Mar 2023 18:43
 Operator : CG\JU
 Sample : 01846-04MS
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TP-4MS


Quant Time: Mar 11 03:05:05 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF022223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 09 16:26:19 2023
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Christian
 Giraldo


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	9.310	196	406901	38.768	ng	99
46) 1,1'-Biphenyl	9.445	154	1358585	44.257	ng	98
47) 2-Chloronaphthalene	9.475	162	1070748	43.994	ng	100
48) 2-Nitroaniline	9.575	65	338882	37.805	ng	87
49) Acenaphthylene	9.892	152	1621772	41.868	ng	100
50) Dimethylphthalate	9.745	163	1282828	43.440	ng	100
51) 2,6-Dinitrotoluene	9.810	165	284701	42.332	ng	99
52) Acenaphthene	10.063	154	1291649m	52.514	ng	03/13/2023
53) 3-Nitroaniline	9.986	138	193667	25.743	ng	91
54) 2,4-Dinitrophenol	10.098	184	249909	59.452	ng	94
55) Dibenzofuran	10.239	168	1768891	49.331	ng	99
56) 4-Nitrophenol	10.151	139	359282	64.037	ng	99
57) 2,4-Dinitrotoluene	10.222	165	348588	38.960	ng	97
58) Fluorene	10.586	166	1385934	50.531	ng	98
59) 2,3,4,6-Tetrachlorophenol	10.357	232	319429	40.970	ng	98
60) Diethylphthalate	10.445	149	1212449	42.038	ng	100
61) 4-Chlorophenyl-phenyle...	10.569	204	598718	42.079	ng	97
62) 4-Nitroaniline	10.604	138	249357	33.766	ng	99
63) Azobenzene	10.727	77	1239593	40.582	ng	98
65) 4,6-Dinitro-2-methylph...	10.633	198	181395	39.021	ng	98
66) n-Nitrosodiphenylamine	10.692	169	994540	46.456	ng	99
67) 4-Bromophenyl-phenylether	11.063	248	372719	47.984	ng	98
68) Hexachlorobenzene	11.133	284	397984	48.693	ng	99
69) Atrazine	11.216	200	328091	49.090	ng	98
70) Pentachlorophenol	11.333	266	324177	66.664	ng	99
71) Phenanthrene	11.563	178	4116944	115.728	ng	99
72) Anthracene	11.610	178	2298386	62.740	ng	99
73) Carbazole	11.763	167	1583203	47.934	ng	100
74) Di-n-butylphthalate	12.074	149	1713419	44.225	ng	99
75) Fluoranthene	12.757	202	3591225	92.343	ng	97
77) Benzidine	12.869	184	528617	60.085	ng	99
78) Pyrene	12.986	202	3178205	96.724	ng	98
80) Butylbenzylphthalate	13.586	149	602488	46.464	ng	97
81) Benzo(a)anthracene	14.180	228	2174312	80.406	ng	97
82) 3,3'-Dichlorobenzidine	14.133	252	263296	33.026	ng	98
83) Chrysene	14.216	228	1932642	76.428	ng	99
84) Bis(2-ethylhexyl)phtha...	14.139	149	784646	49.259	ng	100
85) Di-n-octyl phthalate	14.768	149	1335142	57.359	ng	98
87) Indeno(1,2,3-cd)pyrene	17.327	276	1508742	48.282	ng	98
88) Benzo(b)fluoranthene	15.274	252	2355094	73.690	ng	98
89) Benzo(k)fluoranthene	15.304	252	1430141	45.723	ng	98
90) Benzo(a)pyrene	15.668	252	1846539	61.734	ng	98
91) Dibenzo(a,h)anthracene	17.345	278	1091195	42.859	ng	97
92) Benzo(g,h,i)perylene	17.815	276	1224017	42.724	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.981 min Scan# 710
Delta R.T. -0.000 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43



Instrument :
BNA_F
ClientSampleId :
TP-4MS



Tgt Ion:152 Resp: 221580
Ion Ratio Lower Upper
152 100
150 153.6 125.4 188.2
115 58.5 47.8 71.6


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






03/13/2023
Supervised By :Jagrut
Upadhyay


03/13/2023



#2
1,4-Dioxane
Concen: 38.021 ng
RT: 2.904 min Scan# 54
Delta R.T. 0.065 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43

Tgt Ion: 88 Resp: 287145
Ion Ratio Lower Upper
88 100
58 68.5 58.3 87.5
43 23.2 20.7 31.1







#3
Pyridine
Concen: 36.775 ng
RT: 3.663 min Scan# 11
Delta R.T. 0.041 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43

Instrument :
BNA_F
ClientSampleId :
TP-4MS

Tgt Ion: 79 Resp: 73723
Ion Ratio Lower Upper
79 100
52 62.3 53.7 80.5
51 31.0 26.6 40.0



Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






03/13/2023
Supervised By :Jagrut
Upadhyay


03/13/2023



#4
n-Nitrosodimethylamine
Concen: 38.549 ng
RT: 3.634 min Scan# 178
Delta R.T. 0.047 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43


Tgt Ion: 42 Resp: 291901
Ion Ratio Lower Upper
42 100
74 167.7 118.2 177.2
44 6.4 5.3 7.9





#5
2-Fluorophenol
Concen: 96.284 ng
RT: 5.616 min Scan# 515
Delta R.T. 0.006 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43



Instrument :
BNA_F
ClientSampleId :
TP-4MS



Tgt Ion: 112 Resp: 1314574
Ion Ratio Lower Upper
112 100
64 60.1 51.6 77.4
63 32.1 27.4 41.2


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






03/13/2023
Supervised By :Jagrut
Upadhyay


03/13/2023



#6
Aniline
Concen: 30.125 ng
RT: 6.645 min Scan# 690
Delta R.T. -0.006 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43

Tgt Ion: 93 Resp: 722377
Ion Ratio Lower Upper
93 100
66 37.5 32.5 48.7
65 19.3 16.6 24.8







#7
Phenol-d6
Concen: 93.529 ng
RT: 6.616 min Scan# 609
Delta R.T. -0.006 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43

Instrument :
BNA_F
ClientSampleId :
TP-4MS

Tgt Ion: 99 Resp: 1666549
Ion Ratio Lower Upper
99 100
42 16.0 13.9 20.9
71 34.8 28.5 42.7



Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






03/13/2023
Supervised By :Jagrut
Upadhyay


03/13/2023



#8
2-Chlorophenol
Concen: 46.204 ng
RT: 6.769 min Scan# 711
Delta R.T. -0.006 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43


Tgt Ion:128 Resp: 654721
Ion Ratio Lower Upper
128 100
130 33.3 12.1 52.1
64 48.6 32.0 72.0





#9
Benzaldehyde
Concen: 32.081 ng
RT: 6.534 min Scan# 61
Delta R.T. -0.006 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43



Instrument :
BNA_F
ClientSampleId :
TP-4MS



Tgt Ion: 77 Resp: 30845
Ion Ratio Lower Upper
77 100
105 102.5 77.9 117.9
106 99.3 73.4 113.4


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






03/13/2023
Supervised By :Jagrut
Upadhyay


03/13/2023



#10
Phenol
Concen: 41.141 ng
RT: 6.628 min Scan# 687
Delta R.T. -0.012 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43

Tgt Ion: 94 Resp: 840442
Ion Ratio Lower Upper
94 100
65 28.9 9.4 49.4
66 38.9 19.8 59.8






#11
bis(2-Chloroethyl)ether
Concen: 43.046 ng
RT: 6.710 min Scan# 701
Delta R.T. -0.012 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43

Instrument :

BNA_F

ClientSampleId :



TP-4MS



Tgt Ion: 93 Resp: 678852
Ion Ratio Lower Upper
93 100
63 76.2 59.6 99.6
95 33.1 12.8 52.8


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




03/13/2023
Supervised By :Jagrut
Upadhyay



03/13/2023




#12
1,3-Dichlorobenzene
Concen: 44.557 ng
RT: 6.922 min Scan# 737
Delta R.T. -0.006 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43




Tgt Ion:146 Resp: 677558
Ion Ratio Lower Upper
146 100
148 63.5 51.9 77.9
75 33.9 28.9 43.3





#13
1,4-Dichlorobenzene
Concen: 44.095 ng
RT: 6.998 min Scan# 713
Delta R.T. 0.000 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43



Instrument :
BNA_F
ClientSampleId :
TP-4MS



Tgt Ion:146 Resp: 669539
Ion Ratio Lower Upper
146 100
148 64.9 51.1 76.7
111 43.4 34.8 52.2


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




03/13/2023
Supervised By :Jagrut
Upadhyay



03/13/2023




#14
1,2-Dichlorobenzene
Concen: 44.353 ng
RT: 7.151 min Scan# 776
Delta R.T. -0.006 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43




Tgt Ion:146 Resp: 646314
Ion Ratio Lower Upper
146 100
148 65.1 50.6 75.8
111 45.2 37.2 55.8





#15
Benzyl Alcohol
Concen: 43.223 ng
RT: 7.122 min Scan# 711
Delta R.T. -0.006 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43



Instrument :
BNA_F
ClientSampleId :
TP-4MS



Tgt Ion: 79 Resp: 593139
Ion Ratio Lower Upper
79 100
108 77.9 60.2 90.4
77 63.2 49.9 74.9


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




03/13/2023
Supervised By :Jagrut
Upadhyay



03/13/2023




#16
2,2'-oxybis(1-Chloropropane)
Concen: 40.636 ng
RT: 7.251 min Scan# 793
Delta R.T. -0.006 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43



Tgt Ion: 45 Resp: 817212
Ion Ratio Lower Upper
45 100
77 18.6 0.0 36.4
79 14.9 0.0 33.4





#17
2-Methylphenol
Concen: 39.994 ng
RT: 7.234 min Scan# 790
Delta R.T. -0.006 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43

Instrument :

BNA_F



ClientSampleId :

TP-4MS

Tgt Ion:107 Resp: 52891
Ion Ratio Lower Upper
107 100
108 114.4 93.0 139.4
77 49.5 42.1 63.1
79 47.8 39.6 59.4



Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






03/13/2023
Supervised By :Jagrut
Upadhyay


03/13/2023



#18
Hexachloroethane
Concen: 43.995 ng
RT: 7.492 min Scan# 834
Delta R.T. 0.000 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43


Tgt Ion:117 Resp: 260982
Ion Ratio Lower Upper
117 100
119 96.0 78.0 117.0
201 103.5 79.0 118.4





#19
n-Nitroso-di-n-propylamine
Concen: 37.843 ng
RT: 7.392 min Scan# 816
Delta R.T. -0.006 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43



Instrument :
BNA_F
ClientSampleId :
TP-4MS



Tgt Ion: 70 Resp: 414994
Ion Ratio Lower Upper
70 100
42 51.0 44.1 66.1
101 9.2 7.4 11.2
130 18.6 14.2 21.4


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




03/13/2023
Supervised By :Jagrut
Upadhyay



03/13/2023




#20
3+4-Methylphenols
Concen: 34.853 ng
RT: 7.387 min Scan# 816
Delta R.T. -0.006 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43




Tgt Ion:107 Resp: 595482
Ion Ratio Lower Upper
107 100
108 90.2 72.6 112.6
77 116.9 104.0 144.0
79 30.5 14.1 54.1





#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.269 min Scan# 90
Delta R.T. 0.000 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43



Instrument :
BNA_F
ClientSampleId :
TP-4MS



Tgt Ion:136 Resp: 774449
Ion Ratio Lower Upper
136 100
137 11.2 8.9 13.3
54 7.1 6.2 9.4
68 6.2 5.2 7.8


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






03/13/2023
Supervised By :Jagrut
Upadhyay


03/13/2023



#22
Acetophenone
Concen: 41.064 ng
RT: 7.392 min Scan# 817
Delta R.T. -0.006 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43

Tgt Ion:105 Resp: 801905
Ion Ratio Lower Upper
105 100
71 9.0 5.0 7.4#
51 30.5 31.5 47.3#
120 21.8 17.0 25.4





#23

Nitrobenzene-d5

Concen: 64.569 ng

RT: 7.545 min Scan# 84

Delta R.T. -0.012 min

Lab File: BF132748.D

Acq: 10 Mar 2023 18:43

Instrument :

BNA_F

ClientSampleId :

TP-4MS

Tgt Ion: 82 Resp: 105397

Ion Ratio Lower Upper

82 100



128 40.6 32.2 48.2

54 46.0 38.4 57.6

Manual Integrations



APPROVED

Reviewed By :Christian
Giraldo



03/13/2023
Supervised By :Jagrut
Upadhyay

03/13/2023



#24

Nitrobenzene

Concen: 40.362 ng

RT: 7.569 min Scan# 847

Delta R.T. -0.006 min

Lab File: BF132748.D

Acq: 10 Mar 2023 18:43




Tgt Ion: 77 Resp: 704624


Ion Ratio Lower Upper

77 100

123 44.4 35.8 53.6

65 13.6 10.9 16.3







#25
Isophorone
Concen: 41.778 ng
RT: 7.804 min Scan# 80
Delta R.T. -0.006 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43

Instrument :
BNA_F
ClientSampleId :
TP-4MS

Tgt Ion: 82 Resp: 130747
Ion Ratio Lower Upper
82 100
95 6.9 5.7 8.5
138 16.4 13.3 19.9



Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






03/13/2023
Supervised By :Jagrut
Upadhyay


03/13/2023



#26
2-Nitrophenol
Concen: 44.098 ng
RT: 7.881 min Scan# 900
Delta R.T. -0.006 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43

Tgt Ion:139 Resp: 339665
Ion Ratio Lower Upper
139 100
109 31.1 26.2 39.2
65 56.9 43.8 65.8





#27
2,4-Dimethylphenol
Concen: 43.169 ng
RT: 7.916 min Scan# 906
Delta R.T. -0.006 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43

Instrument :

BNA_F



ClientSampleId :

TP-4MS

Tgt Ion:122 Resp: 524804
Ion Ratio Lower Upper
122 100
107 126.2 98.7 148.1
121 57.9 45.8 68.6



Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






03/13/2023
Supervised By :Jagrut
Upadhyay


03/13/2023



#28
bis(2-Chloroethoxy)methane
Concen: 43.067 ng
RT: 8.010 min Scan# 922
Delta R.T. 0.000 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43

Tgt Ion: 93 Resp: 788454
Ion Ratio Lower Upper
93 100
95 33.5 26.3 39.5
123 10.9 8.3 12.5





#29

2,4-Dichlorophenol

Concen: 43.233 ng

RT: 8.128 min Scan# 942

Delta R.T. -0.006 min

Lab File: BF132748.D

Acq: 10 Mar 2023 18:43

Instrument :

BNA_F



ClientSampleId :

TP-4MS

Tgt Ion:	162	Resp:	52281
Ion Ratio	Lower	Upper	
162	100		
164	65.6	45.3	85.3
98	38.1	17.9	57.9



Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo



03/13/2023
Supervised By :Jagrut
Upadhyay

03/13/2023



#30

1,2,4-Trichlorobenzene

Concen: 44.536 ng




RT: 8.204 min Scan# 955


Delta R.T. -0.006 min

Lab File: BF132748.D

Acq: 10 Mar 2023 18:43


Tgt Ion:	180	Resp:	584814
Ion Ratio	Lower	Upper	
180	100		
182	94.3	75.1	112.7
145	29.5	23.8	35.6





#31
Naphthalene
Concen: 47.366 ng
RT: 8.292 min Scan# 91
Delta R.T. 0.000 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43



Instrument :
BNA_F
ClientSampleId :
TP-4MS



Tgt Ion:128 Resp: 187791
Ion Ratio Lower Upper
128 100
129 11.9 9.1 13.7
127 13.7 11.0 16.6


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




03/13/2023
Supervised By :Jagrut
Upadhyay



03/13/2023




#32
Benzoic acid
Concen: 35.823 ng m
RT: 8.045 min Scan# 928
Delta R.T. -0.006 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43



Tgt Ion:122 Resp: 333051
Ion Ratio Lower Upper
122 100
105 120.5 102.6 142.6
77 90.5 70.2 110.2





#33
4-Chloroaniline
Concen: 12.515 ng
RT: 8.334 min Scan# 91
Delta R.T. -0.006 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43

Instrument :

BNA_F



ClientSampleId :

TP-4MS

Tgt Ion:127 Resp: 21262
Ion Ratio Lower Upper
127 100
129 32.4 26.4 39.6
65 31.9 28.8 43.2
92 20.1 17.1 25.7



Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






03/13/2023
Supervised By :Jagrut
Upadhyay

03/13/2023



#34
Hexachlorobutadiene
Concen: 45.378 ng
RT: 8.398 min Scan# 988
Delta R.T. -0.006 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43

Tgt Ion:225 Resp: 378646
Ion Ratio Lower Upper
225 100
223 63.3 51.0 76.4
227 65.0 51.7 77.5





#35

Caprolactam

Concen: 43.065 ng m

RT: 8.722 min Scan# 1043

Delta R.T. -0.012 min

Lab File: BF132748.D

Acq: 10 Mar 2023 18:43

Instrument :

BNA_F

ClientSampleId :

TP-4MS

Tgt Ion:113 Resp: 171692

Ion Ratio Lower Upper

113 100



55 171.1 149.4 189.4

56 125.3 105.0 145.0

Manual Integrations



APPROVED

Reviewed By :Christian
Giraldo



03/13/2023
Supervised By :Jagrut
Upadhyay

03/13/2023



#36

4-Chloro-3-methylphenol

Concen: 42.448 ng

RT: 8.828 min Scan# 1061

Delta R.T. 0.006 min

Lab File: BF132748.D

Acq: 10 Mar 2023 18:43




Tgt Ion:107 Resp: 563738


Ion Ratio Lower Upper

107 100

144 25.3 18.9 28.3

142 78.3 57.9 86.9





#37
2-Methylnaphthalene
Concen: 45.020 ng
RT: 8.981 min Scan# 1104
Delta R.T. -0.006 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43

Instrument :

BNA_F



ClientSampleId :

TP-4MS

Tgt Ion:142 Resp: 1216404
Ion Ratio Lower Upper
142 100
141 86.2 70.2 105.2



Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






03/13/2023
Supervised By :Jagrut
Upadhyay


03/13/2023



#38
1-Methylnaphthalene
Concen: 43.556 ng
RT: 9.081 min Scan# 1104
Delta R.T. -0.006 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43


Tgt Ion:142 Resp: 1099800
Ion Ratio Lower Upper
142 100
141 90.0 73.0 109.6





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 10.028 min Scan# 1116
Delta R.T. -0.006 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43



Instrument :
BNA_F
ClientSampleId :
TP-4MS



Tgt Ion:164 Resp: 405578
Ion Ratio Lower Upper
164 100
162 102.5 79.8 119.8
160 45.6 35.8 53.6


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




03/13/2023
Supervised By :Jagrut
Upadhyay



03/13/2023




#40
1,2,4,5-Tetrachlorobenzene
Concen: 45.185 ng
RT: 9.151 min Scan# 1116
Delta R.T. 0.000 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43




Tgt Ion:216 Resp: 599589
Ion Ratio Lower Upper
216 100
214 79.7 63.0 94.6
179 20.4 16.6 24.8
108 19.1 15.4 23.0





#41
Hexachlorocyclopentadiene
Concen: 67.754 ng
RT: 9.133 min Scan# 1113
Delta R.T. 0.000 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43



Instrument :
BNA_F
ClientSampleId :
TP-4MS



Tgt Ion:237 Resp: 487649
Ion Ratio Lower Upper
237 100
235 62.6 42.4 82.4
272 12.9 0.0 32.7


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




03/13/2023
Supervised By :Jagrut
Upadhyay



03/13/2023




#42
2,4,6-Tribromophenol
Concen: 94.425 ng
RT: 10.827 min Scan# 1401
Delta R.T. 0.000 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43




Tgt Ion:330 Resp: 413589
Ion Ratio Lower Upper
330 100
332 96.9 76.2 114.4
141 33.9 28.4 42.6





#43
2,4,6-Trichlorophenol
Concen: 42.955 ng
RT: 9.263 min Scan# 1135
Delta R.T. 0.000 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43



Instrument :
BNA_F
ClientSampleId :
TP-4MS



Tgt Ion:196 Resp: 38628
Ion Ratio Lower Upper
196 100
198 95.6 77.1 115.7
200 31.1 25.0 37.4


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




03/13/2023
Supervised By :Jagrut
Upadhyay



03/13/2023




#44
2,4,5-Trichlorophenol
Concen: 38.768 ng
RT: 9.310 min Scan# 1143
Delta R.T. 0.000 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43




Tgt Ion:196 Resp: 406901
Ion Ratio Lower Upper
196 100
198 96.7 78.3 117.5
97 46.0 38.1 57.1
132 25.0 20.2 30.4





#45
2-Fluorobiphenyl
Concen: 63.041 ng
RT: 9.345 min Scan# 1112
Delta R.T. 0.000 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43



Instrument :
BNA_F
ClientSampleId :
TP-4MS



Tgt Ion:172 Resp: 1679178
Ion Ratio Lower Upper
172 100
171 35.5 28.8 43.2
170 24.2 19.7 29.5


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






03/13/2023
Supervised By :Jagrut
Upadhyay


03/13/2023



#46
1,1'-Biphenyl
Concen: 44.257 ng
RT: 9.445 min Scan# 1166
Delta R.T. -0.006 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43

Tgt Ion:154 Resp: 1358585
Ion Ratio Lower Upper
154 100
153 40.0 20.9 60.9
76 13.7 0.0 34.7





#47

2-Chloronaphthalene

Concen: 43.994 ng

RT: 9.475 min Scan# 1171

Delta R.T. -0.006 min

Lab File: BF132748.D

Acq: 10 Mar 2023 18:43

Instrument :

BNA_F

ClientSampleId :

TP-4MS

Tgt Ion:162 Resp: 1070748

Ion Ratio Lower Upper

162 100



127 40.6 32.3 48.5

164 33.3 26.9 40.3

Manual Integrations



APPROVED

Reviewed By :Christian
Giraldo



03/13/2023
Supervised By :Jagrut
Upadhyay

03/13/2023



#48

2-Nitroaniline

Concen: 37.805 ng

RT: 9.575 min Scan# 1188

Delta R.T. 0.000 min

Lab File: BF132748.D

Acq: 10 Mar 2023 18:43




Tgt Ion: 65 Resp: 338882


Ion Ratio Lower Upper

65 100

92 75.3 54.5 81.7


138 112.7 77.2 115.8





#49
Acenaphthylene
Concen: 41.868 ng
RT: 9.892 min Scan# 1242
Delta R.T. -0.006 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43



Instrument :
BNA_F
ClientSampleId :
TP-4MS



Tgt Ion:152 Resp: 1621772
Ion Ratio Lower Upper
152 100
151 20.5 16.4 24.6
153 13.5 11.0 16.6


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




03/13/2023
Supervised By :Jagrut
Upadhyay



03/13/2023




#50
Dimethylphthalate
Concen: 43.440 ng
RT: 9.745 min Scan# 1217
Delta R.T. -0.006 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43




Tgt Ion:163 Resp: 1282828
Ion Ratio Lower Upper
163 100
194 4.1 3.3 4.9
164 10.3 8.3 12.5





#51
2,6-Dinitrotoluene
Concen: 42.332 ng
RT: 9.810 min Scan# 1191
Delta R.T. -0.006 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43



Instrument :
BNA_F
ClientSampleId :
TP-4MS



Tgt Ion:165 Resp: 28470
Ion Ratio Lower Upper
165 100
63 58.8 47.3 70.9
89 61.8 48.9 73.3


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






03/13/2023
Supervised By :Jagrut
Upadhyay


03/13/2023



#52
Acenaphthene
Concen: 52.514 ng m
RT: 10.063 min Scan# 1271
Delta R.T. -0.006 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43


Tgt Ion:154 Resp: 1291649
Ion Ratio Lower Upper
154 100
153 110.2 89.8 134.8
152 54.0 43.3 64.9





#53
3-Nitroaniline
Concen: 25.743 ng
RT: 9.986 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43



Instrument :
BNA_F
ClientSampleId :
TP-4MS



Tgt Ion:138 Resp: 19366
Ion Ratio Lower Upper
138 100
108 9.2 8.1 12.1
92 110.8 97.2 145.8


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




03/13/2023
Supervised By :Jagrut
Upadhyay



03/13/2023




#54
2,4-Dinitrophenol
Concen: 59.452 ng
RT: 10.098 min Scan# 1277
Delta R.T. 0.000 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43




Tgt Ion:184 Resp: 249909
Ion Ratio Lower Upper
184 100
63 54.9 51.4 77.2
154 60.1 48.3 72.5





#55
Dibenzenofuran
Concen: 49.331 ng
RT: 10.239 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43



Instrument :
BNA_F
ClientSampleId :
TP-4MS



Tgt Ion:168 Resp: 176889
Ion Ratio Lower Upper
168 100
139 38.0 29.8 44.6
169 13.9 10.9 16.3


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






03/13/2023
Supervised By :Jagrut
Upadhyay


03/13/2023



#56
4-Nitrophenol
Concen: 64.037 ng
RT: 10.151 min Scan# 1286
Delta R.T. 0.000 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43


Tgt Ion:139 Resp: 359282
Ion Ratio Lower Upper
139 100
109 77.5 56.7 96.7
65 102.3 83.5 123.5





#57
2,4-Dinitrotoluene
Concen: 38.960 ng
RT: 10.222 min Scan# 1360
Delta R.T. 0.000 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43



Instrument :
BNA_F
ClientSampleId :
TP-4MS



Tgt Ion:165 Resp: 348588
Ion Ratio Lower Upper
165 100
63 43.0 36.5 54.7
89 75.8 62.4 93.6
182 2.6 2.1 3.1


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




03/13/2023
Supervised By :Jagrut
Upadhyay



03/13/2023




#58
Fluorene
Concen: 50.531 ng
RT: 10.586 min Scan# 1360
Delta R.T. 0.000 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43




Tgt Ion:166 Resp: 1385934
Ion Ratio Lower Upper
166 100
165 94.6 74.2 111.4
167 13.9 11.1 16.7





#59
2,3,4,6-Tetrachlorophenol
Concen: 40.970 ng
RT: 10.357 min Scan# 1321
Delta R.T. 0.000 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43



Instrument :
BNA_F
ClientSampleId :
TP-4MS



Tgt Ion:232 Resp: 319429
Ion Ratio Lower Upper
232 100
131 41.3 33.8 50.8
130 2.4 1.7 2.5
166 28.4 23.4 35.0


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






03/13/2023
Supervised By :Jagrut
Upadhyay


03/13/2023



#60
Diethylphthalate
Concen: 42.038 ng
RT: 10.445 min Scan# 1336
Delta R.T. 0.000 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43


Tgt Ion:149 Resp: 1212449
Ion Ratio Lower Upper
149 100
177 22.9 18.1 27.1
150 12.7 10.2 15.2





#61
4-Chlorophenyl-phenylether
Concen: 42.079 ng
RT: 10.569 min Scan# 1357
Delta R.T. 0.000 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43



Instrument :
BNA_F
ClientSampleId :
TP-4MS



Tgt Ion:204 Resp: 598718
Ion Ratio Lower Upper
204 100
206 32.5 26.3 39.5
141 56.8 47.8 71.8


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






03/13/2023
Supervised By :Jagrut
Upadhyay


03/13/2023



#62
4-Nitroaniline
Concen: 33.766 ng
RT: 10.604 min Scan# 1363
Delta R.T. -0.006 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43


Tgt Ion:138 Resp: 249357
Ion Ratio Lower Upper
138 100
92 53.7 33.8 73.8
108 82.3 63.7 103.7





#63
Azobenzene
Concen: 40.582 ng
RT: 10.727 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43


Instrument :
BNA_F
ClientSampleId :
TP-4MS



Tgt Ion: 77 Resp: 123959
Ion Ratio Lower Upper
77 100
182 26.9 5.5 45.5
105 21.0 1.8 41.8
51 26.5 6.4 46.4


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






03/13/2023
Supervised By :Jagrut
Upadhyay


03/13/2023



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.527 min Scan# 1520
Delta R.T. 0.000 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43


Tgt Ion:188 Resp: 686697
Ion Ratio Lower Upper
188 100
94 9.0 8.3 12.5
80 9.8 8.7 13.1





#65
4,6-Dinitro-2-methylphenol
Concen: 39.021 ng
RT: 10.633 min Scan# 1368
Delta R.T. 0.000 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43



Instrument :
BNA_F
ClientSampleId :
TP-4MS



Tgt Ion:198 Resp: 18139
Ion Ratio Lower Upper
198 100
51 39.0 20.6 60.6
105 43.7 24.3 64.3


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




03/13/2023
Supervised By :Jagrut
Upadhyay



03/13/2023




#66
n-Nitrosodiphenylamine
Concen: 46.456 ng
RT: 10.692 min Scan# 1378
Delta R.T. 0.000 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43




Tgt Ion:169 Resp: 994540
Ion Ratio Lower Upper
169 100
168 68.3 54.2 81.4
167 37.4 29.6 44.4





#67
4-Bromophenyl-phenylether
Concen: 47.984 ng
RT: 11.063 min Scan# 1404
Delta R.T. 0.000 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43



Instrument :
BNA_F
ClientSampleId :
TP-4MS



Tgt Ion:248 Resp: 372719
Ion Ratio Lower Upper
248 100
250 98.0 77.0 115.4
141 65.0 54.2 81.2


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




03/13/2023
Supervised By :Jagrut
Upadhyay



03/13/2023




#68
Hexachlorobenzene
Concen: 48.693 ng
RT: 11.133 min Scan# 1453
Delta R.T. 0.000 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43




Tgt Ion:284 Resp: 397984
Ion Ratio Lower Upper
284 100
142 28.2 22.2 33.4
249 30.4 24.9 37.3





#69
Atrazine
Concen: 49.090 ng
RT: 11.216 min Scan# 1467
Delta R.T. -0.006 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43



Instrument :
BNA_F
ClientSampleId :
TP-4MS



Tgt Ion:200 Resp: 32809
Ion Ratio Lower Upper
200 100
173 26.1 6.4 46.4
215 49.4 31.0 71.0


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




03/13/2023
Supervised By :Jagrut
Upadhyay



03/13/2023




#70
Pentachlorophenol
Concen: 66.664 ng
RT: 11.333 min Scan# 1487
Delta R.T. 0.000 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43



Tgt Ion:266 Resp: 324177
Ion Ratio Lower Upper
266 100
268 63.0 49.8 74.6
264 63.5 50.9 76.3







#71
Phenanthrene
Concen: 115.728 ng
RT: 11.563 min Scan# 1153
Delta R.T. 0.006 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43

Instrument :
BNA_F
ClientSampleId :
TP-4MS

Tgt Ion:178 Resp: 411694
Ion Ratio Lower Upper
178 100
176 21.1 16.5 24.7
179 16.7 13.0 19.6



Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






03/13/2023
Supervised By :Jagrut
Upadhyay


03/13/2023



#72
Anthracene
Concen: 62.740 ng
RT: 11.610 min Scan# 1534
Delta R.T. 0.006 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43


Tgt Ion:178 Resp: 2298386
Ion Ratio Lower Upper
178 100
176 20.0 15.6 23.4
179 16.2 12.5 18.7





#73
 Carbazole
 Concen: 47.934 ng
 RT: 11.763 min Scan# 1612
 Delta R.T. 0.000 min
 Lab File: BF132748.D
 Acq: 10 Mar 2023 18:43



Instrument :
 BNA_F
 ClientSampleId :
 TP-4MS



Tgt Ion:167 Resp: 1583200
 Ion Ratio Lower Upper
 167 100
 166 22.1 17.7 26.5
 139 12.7 10.5 15.7


Manual Integrations
 APPROVED

Reviewed By :Christian
 Giraldo




03/13/2023
 Supervised By :Jagrut
 Upadhyay



03/13/2023




#74
 Di-n-butylphthalate
 Concen: 44.225 ng
 RT: 12.074 min Scan# 1613
 Delta R.T. 0.000 min
 Lab File: BF132748.D
 Acq: 10 Mar 2023 18:43




Tgt Ion:149 Resp: 1713419
 Ion Ratio Lower Upper
 149 100
 150 9.6 7.8 11.8
 104 5.1 4.3 6.5





#75
Fluoranthene
Concen: 92.343 ng
RT: 12.757 min Scan# 11
Delta R.T. 0.006 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43



Instrument :
BNA_F
ClientSampleId :
TP-4MS



Tgt Ion:202 Resp: 359122
Ion Ratio Lower Upper
202 100
101 9.7 0.0 31.4
203 19.1 0.0 38.0


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






03/13/2023
Supervised By :Jagrut
Upadhyay


03/13/2023



#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.186 min Scan# 1972
Delta R.T. 0.000 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43

Tgt Ion:240 Resp: 361410
Ion Ratio Lower Upper
240 100
120 11.1 7.2 10.8#
236 26.2 21.8 32.6





#77

Benzidine

Concen: 60.085 ng

RT: 12.869 min Scan# 1710

Delta R.T. 0.000 min

Lab File: BF132748.D

Acq: 10 Mar 2023 18:43

Instrument :

BNA_F

ClientSampleId :



TP-4MS

Tgt Ion:184 Resp: 52861

Ion	Ratio	Lower	Upper
184	100		
185	14.2	11.7	17.5
183	11.0	8.7	13.1



Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo



03/13/2023
Supervised By :Jagrut
Upadhyay

03/13/2023



#78

Pyrene

Concen: 96.724 ng

RT: 12.986 min Scan# 1768




Delta R.T. 0.006 min


Lab File: BF132748.D

Acq: 10 Mar 2023 18:43

Tgt Ion:202 Resp: 3178205


Ion	Ratio	Lower	Upper
202	100		
200	22.0	17.1	25.7
203	19.1	14.6	21.8





#79
 Terphenyl-d14
 Concen: 65.189 ng
 RT: 13.110 min Scan# 1753
 Delta R.T. -0.006 min
 Lab File: BF132748.D
 Acq: 10 Mar 2023 18:43



Instrument :
 BNA_F
 ClientSampleId :
 TP-4MS



Tgt Ion:244 Resp: 1393462
 Ion Ratio Lower Upper
 244 100
 212 7.6 6.5 9.7
 122 9.9 7.5 11.3


Manual Integrations
 APPROVED

Reviewed By :Christian
 Giraldo




03/13/2023
 Supervised By :Jagrut
 Upadhyay



03/13/2023




#80
 Butylbenzylphthalate
 Concen: 46.464 ng
 RT: 13.586 min Scan# 1870
 Delta R.T. 0.000 min
 Lab File: BF132748.D
 Acq: 10 Mar 2023 18:43



Tgt Ion:149 Resp: 602488
 Ion Ratio Lower Upper
 149 100
 91 70.0 58.4 87.6
 206 22.6 16.8 25.2







#81
Benzo(a)anthracene
Concen: 80.406 ng
RT: 14.180 min Scan# 1933
Delta R.T. 0.006 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43

Instrument :
BNA_F
ClientSampleId :
TP-4MS

Tgt Ion:228 Resp: 217431
Ion Ratio Lower Upper
228 100
226 28.2 21.7 32.5
229 20.8 15.4 23.2



Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






03/13/2023
Supervised By :Jagrut
Upadhyay


03/13/2023



#82
3,3'-Dichlorobenzidine
Concen: 33.026 ng
RT: 14.133 min Scan# 1963
Delta R.T. 0.000 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43

Tgt Ion:252 Resp: 263296
Ion Ratio Lower Upper
252 100
254 62.7 51.4 77.2
126 9.2 8.9 13.3







#83
Chrysene
Concen: 76.428 ng
RT: 14.216 min Scan# 1939
Delta R.T. 0.000 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43

Instrument :
BNA_F
ClientSampleId :
TP-4MS

Tgt Ion:228 Resp: 193264
Ion Ratio Lower Upper
228 100
226 31.5 25.0 37.6
229 21.9 16.4 24.6



Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






03/13/2023
Supervised By :Jagrut
Upadhyay


03/13/2023



#84
Bis(2-ethylhexyl)phthalate
Concen: 49.259 ng
RT: 14.139 min Scan# 1964
Delta R.T. -0.006 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43


Tgt Ion:149 Resp: 784646
Ion Ratio Lower Upper
149 100
167 27.2 21.8 32.6
279 3.7 3.0 4.6





#85
Di-n-octyl phthalate
Concen: 57.359 ng
RT: 14.768 min Scan# 2034
Delta R.T. 0.000 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43


Instrument :
BNA_F
ClientSampleId :
TP-4MS



Tgt Ion:149 Resp: 133514
Ion Ratio Lower Upper
149 100
167 1.7 1.3 1.9
43 8.0 7.2 10.8


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






03/13/2023
Supervised By :Jagrut
Upadhyay


03/13/2023



#86
Perylene-d12
Concen: 20.000 ng
RT: 15.727 min Scan# 2234
Delta R.T. 0.000 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43


Tgt Ion:264 Resp: 476900
Ion Ratio Lower Upper
264 100
260 24.5 18.8 28.2
265 23.0 17.1 25.7





#87
Indeno(1,2,3-cd)pyrene
Concen: 48.282 ng
RT: 17.327 min Scan# 2157
Delta R.T. 0.000 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43



Instrument :
BNA_F
ClientSampleId :
TP-4MS



Tgt Ion:276 Resp: 150874
Ion Ratio Lower Upper
276 100
138 21.7 19.0 28.4
277 24.9 20.0 30.0


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






03/13/2023
Supervised By :Jagrut
Upadhyay


03/13/2023



#88
Benzo(b)fluoranthene
Concen: 73.690 ng
RT: 15.274 min Scan# 2157
Delta R.T. 0.006 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43


Tgt Ion:252 Resp: 2355094
Ion Ratio Lower Upper
252 100
253 23.2 17.4 26.0
125 9.0 7.0 10.6





#89
Benzo(k)fluoranthene
Concen: 45.723 ng
RT: 15.304 min Scan# 2125
Delta R.T. 0.000 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43



Instrument :
BNA_F
ClientSampleId :
TP-4MS



Tgt Ion:252 Resp: 143014
Ion Ratio Lower Upper
252 100
253 22.9 17.5 26.3
125 8.9 7.2 10.8


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




03/13/2023
Supervised By :Jagrut
Upadhyay



03/13/2023




#90
Benzo(a)pyrene
Concen: 61.734 ng
RT: 15.668 min Scan# 2224
Delta R.T. 0.006 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43



Tgt Ion:252 Resp: 1846539
Ion Ratio Lower Upper
252 100
253 23.1 17.2 25.8
125 9.9 8.0 12.0







#91
Dibenzo(a,h)anthracene
Concen: 42.859 ng
RT: 17.345 min Scan# 2509
Delta R.T. 0.000 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43

Instrument :
BNA_F
ClientSampleId :
TP-4MS

Tgt Ion:278 Resp: 109119
Ion Ratio Lower Upper
278 100
139 13.3 13.0 19.4
279 24.2 19.0 28.4



Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo



03/13/2023
Supervised By :Jagrut
Upadhyay

03/13/2023



#92
Benzo(g,h,i)perylene
Concen: 42.724 ng
RT: 17.815 min Scan# 2589
Delta R.T. 0.000 min
Lab File: BF132748.D
Acq: 10 Mar 2023 18:43

Tgt Ion:276 Resp: 1224017
Ion Ratio Lower Upper
276 100
277 24.3 18.9 28.3
138 19.5 16.2 24.4

