

Data Path : Z:\HPCHEM1\BNA F\DATA\BF031115\
 Data File : BF077707.D
 Acq On : 12 Mar 2015 1:45
 Operator : TP/IZ
 Sample : G1474-03
 Misc : W
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 MW-5

Quant Time: Mar 12 06:06:27 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF031115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 12 02:24:12 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.17	152	58847	20.00	ng	0.00
21) Naphthalene-d8	8.74	136	238049	20.00	ng	-0.01
38) Acenaphthene-d10	10.91	164	121939	20.00	ng	0.00
63) Phenanthrene-d10	12.75	188	241279	20.00	ng	0.00
75) Chrysene-d12	16.03	240	246214	20.00	ng	0.00
86) Perylene-d12	17.71	264	236335	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.46	112	209427	62.12	ng	0.00
7) Phenol-d6	6.73	99	164212	39.42	ng	-0.01
23) Nitrobenzene-d5	7.86	82	304350	83.81	ng	0.00
41) 2,4,6-Tribromophenol	11.89	330	138646	118.84	ng	0.00
44) 2-Fluorobiphenyl	10.09	172	699598	84.32	ng	0.00
78) Terphenyl-d14	14.75	244	842852	83.62	ng	0.00
Target Compounds						
49) Dimethylphthalate	10.59	163	20815	2.30	ng	98
59) Diethylphthalate	11.47	149	303614	34.45	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF031115\
Data File : BF077707.D
Acq On : 12 Mar 2015 1:45
Operator : TP/IZ
Sample : G1474-03
Misc : W
ALS Vial : 25 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
MW-5

Quant Time: Mar 12 06:06:27 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF031115.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 12 02:24:12 2015
Response via : Initial Calibration

