

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF031116\
 Data File : BF085394.D
 Acq On : 12 Mar 2016 2:56
 Operator : UM/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 14 02:22:37 2016
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF030316.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 04 00:54:58 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	72	-0.05
2	1,4-Dioxane	0.610	0.583	4.4	73	-0.08
3	Pyridine	1.526	1.333	12.6	66	-0.09
4	n-Nitrosodimethylamine	0.653	0.620	5.1	68	-0.10
5 S	2-Fluorophenol	1.192	1.107	7.1	74	-0.05
6	Aniline	2.189	2.057	6.0	68	-0.05
7 S	Phenol-d6	1.562	1.561	0.1	77	-0.05
8	2-Chlorophenol	1.435	1.427	0.6	74	-0.05
9	Benzaldehyde	0.804	0.720	10.4	74	-0.05
10 C	Phenol	1.673	1.800	-7.6	83	-0.03
11	bis(2-Chloroethyl)ether	1.390	1.407	-1.2	70	-0.05
12	1,3-Dichlorobenzene	1.541	1.536	0.3	74	-0.05
13 C	1,4-Dichlorobenzene	1.545	1.446	6.4	74	-0.05
14	1,2-Dichlorobenzene	1.401	1.386	1.1	71	-0.05
15	Benzyl Alcohol	1.147	1.018	11.2	67	-0.05
16	2,2'-oxybis(1-Chloropropane	2.003	1.714	14.4	66	-0.05
17	2-Methylphenol	1.171	1.070	8.6	65	-0.05
18	Hexachloroethane	0.570	0.549	3.7	70	-0.05
19 P	n-Nitroso-di-n-propylamine	1.018	0.946	7.1	68	-0.05
20	3+4-Methylphenols	1.387	1.340	3.4	77	-0.03
21 I	Naphthalene-d8	1.000	1.000	0.0	73	-0.03
22	Acetophenone	0.488	0.424	13.1	63	-0.05
23 S	Nitrobenzene-d5	0.374	0.358	4.3	68	-0.03
24	Nitrobenzene	0.380	0.334	12.1	63	-0.05
25	Isophorone	0.693	0.670	3.3	70	-0.05
26 C	2-Nitrophenol	0.185	0.205	-10.8	77	-0.05
27	2,4-Dimethylphenol	0.338	0.316	6.5	65	-0.05
28	bis(2-Chloroethoxy)methane	0.386	0.403	-4.4	77	-0.03
29 C	2,4-Dichlorophenol	0.272	0.269	1.1	71	-0.05
30	1,2,4-Trichlorobenzene	0.294	0.295	-0.3	73	-0.05
31	Naphthalene	0.983	0.986	-0.3	78	-0.03
32	Benzoic acid	0.224	0.255	-13.8	76	-0.05
33	4-Chloroaniline	0.398	0.429	-7.8	83	-0.03
34 C	Hexachlorobutadiene	0.153	0.165	-7.8	77	-0.05
35	Caprolactam	0.089	0.086	3.4	70	-0.05
36 C	4-Chloro-3-methylphenol	0.309	0.328	-6.1	78	-0.03
37	2-Methylnaphthalene	0.631	0.580	8.1	66	-0.05
38 I	Acenaphthene-d10	1.000	1.000	0.0	65	-0.05
39	1,2,4,5-Tetrachlorobenzene	0.572	0.667	-16.6	82	-0.03
40 P	Hexachlorocyclopentadiene	0.308	0.187	39.3#	39#	-0.03
41 S	2,4,6-Tribromophenol	0.171	0.176	-2.9	65	-0.03
42 C	2,4,6-Trichlorophenol	0.426	0.418	1.9	64	-0.05
43	2,4,5-Trichlorophenol	0.404	0.466	-15.3	73	-0.03
44 S	2-Fluorobiphenyl	1.315	1.253	4.7	66	-0.05

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF031116\
 Data File : BF085394.D
 Acq On : 12 Mar 2016 2:56
 Operator : UM/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 14 02:22:37 2016
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF030316.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 04 00:54:58 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.709	1.875	-9.7	80	-0.03
46	2-Chloronaphthalene	1.302	1.316	-1.1	67	-0.05
47	2-Nitroaniline	0.404	0.400	1.0	62	-0.03
48	Acenaphthylene	2.027	1.921	5.2	65	-0.05
49	Dimethylphthalate	1.535	1.501	2.2	63	-0.05
50	2,6-Dinitrotoluene	0.328	0.327	0.3	62	-0.05
51 C	Acenaphthene	1.231	1.221	0.8	68	-0.05
52	3-Nitroaniline	0.393	0.363	7.6	54	-0.05
53 P	2,4-Dinitrophenol	0.137	0.115	16.1	52	-0.03
54	Dibenzofuran	1.613	1.513	6.2	62	-0.05
55 P	4-Nitrophenol	0.309	0.319	-3.2	61	-0.03
56	2,4-Dinitrotoluene	0.420	0.470	-11.9	72	-0.03
57	Fluorene	1.267	1.211	4.4	62	-0.05
58	2,3,4,6-Tetrachlorophenol	0.284	0.291	-2.5	63	-0.03
59	Diethylphthalate	1.490	1.521	-2.1	67	-0.05
60	4-Chlorophenyl-phenylether	0.616	0.645	-4.7	71	-0.03
61	4-Nitroaniline	0.367	0.405	-10.4	69	-0.03
62	Azobenzene	1.468	1.435	2.2	63	-0.05
63 I	Phenanthrene-d10	1.000	1.000	0.0	58	-0.05
64	4,6-Dinitro-2-methylphenol	0.120	0.096	20.0	41#	-0.04
65 c	n-Nitrosodiphenylamine	0.622	0.716	-15.1	71	-0.03
66	4-Bromophenyl-phenylether	0.194	0.212	-9.3	62	-0.05
67	Hexachlorobenzene	0.207	0.209	-1.0	57	-0.05
68	Atrazine	0.173	0.206	-19.1	81	-0.03
69 C	Pentachlorophenol	0.115	0.128	-11.3	64	-0.03
70	Phenanthrene	1.048	1.063	-1.4	65	-0.05
71	Anthracene	1.113	1.098	1.3	62	-0.05
72	Carbazole	0.976	0.893	8.5	55	-0.05
73	Di-n-butylphthalate	1.233	1.198	2.8	61	-0.03
74 C	Fluoranthene	1.052	0.932	11.4	54	-0.05
75 I	Chrysene-d12	1.000	1.000	0.0	63	-0.05
76	Benzidine	0.595	0.745	-25.2#	68	-0.03
77	Pyrene	1.505	1.462	2.9	55	-0.05
78 S	Terphenyl-d14	0.862	0.888	-3.0	64	-0.03
79	Butylbenzylphthalate	0.683	0.698	-2.2	60	-0.03
80	Benzo(a)anthracene	1.197	1.201	-0.3	62	-0.03
81	3,3'-Dichlorobenzidine	0.378	0.438	-15.9	68	-0.03
82	Chrysene	1.106	1.249	-12.9	64	-0.03
83	Bis(2-ethylhexyl)phthalate	0.899	0.956	-6.3	64	-0.03
84 c	Di-n-octyl phthalate	1.369	1.606	-17.3	67	-0.03
85	Indeno(1,2,3-cd)pyrene	0.863	1.216	-40.9#	75	-0.06
86 I	Perylene-d12	1.000	1.000	0.0	80	-0.05
87	Benzo(b)fluoranthene	1.333	1.102	17.3	62	-0.03

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF031116\
Data File : BF085394.D
Acq On : 12 Mar 2016 2:56
Operator : UM/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Mar 14 02:22:37 2016
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF030316.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 04 00:54:58 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.075	1.186	-10.3	101	-0.03
89 C	Benzo(a)pyrene	1.105	1.116	-1.0	79	-0.05
90	Dibenzo(a,h)anthracene	0.943	1.005	-6.6	77	-0.06
91	Benzo(g,h,i)perylene	0.948	0.933	1.6	70	-0.06

(#) = Out of Range

SPCC's out = 0 CCC's out = 0