

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF031117\
 Data File : BF093601.D
 Acq On : 12 Mar 2017 9:55
 Operator : SJ/MA
 Sample : I2198-02
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 W-23

Quant Time: Mar 13 18:52:38 2017
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF030717.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 13 15:05:08 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.72	152	174222	20.00	ng	0.00
21) Naphthalene-d8	8.01	136	671000	20.00	ng	0.00
38) Acenaphthene-d10	9.76	164	289835	20.00	ng	0.00
63) Phenanthrene-d10	11.25	188	425777	20.00	ng	0.00
75) Chrysene-d12	13.88	240	354905	20.00	ng	-0.01
86) Perylene-d12	15.28	264	293930	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.32	112	631248	60.30	ng	0.00
7) Phenol-d6	6.38	99	518117	39.25	ng	0.00
23) Nitrobenzene-d5	7.29	82	1003445	82.97	ng	-0.01
41) 2,4,6-Tribromophenol	10.55	330	238796	126.46	ng	0.00
44) 2-Fluorobiphenyl	9.09	172	1739241	111.62	ng	0.00
78) Terphenyl-d14	12.83	244	1072376	78.56	ng	0.00
Target Compounds						
49) Dimethylphthalate	9.49	163	49018	2.55	ng	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF031117\
Data File : BF093601.D
Acq On : 12 Mar 2017 9:55
Operator : SJ/MA
Sample : I2198-02
Misc :
ALS Vial : 36 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
W-23

Quant Time: Mar 13 18:52:38 2017
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF030717.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 13 15:05:08 2017
Response via : Initial Calibration

