

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031124\
 Data File : BF137553.D
 Acq On : 11 Mar 2024 13:10
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 12 02:01:30 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF022924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 12 02:00:03 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	122	0.00
2	1,4-Dioxane	0.722	0.688	4.7	114	0.00
3	Pyridine	1.741	1.791	-2.9	115	0.00
4	n-Nitrosodimethylamine	0.652	0.701	-7.5	122	0.00
5 S	2-Fluorophenol	1.321	1.301	1.5	115	0.00
6	Aniline	2.329	2.379	-2.1	118	0.00
7 S	Phenol-d6	1.906	1.835	3.7	117	0.00
8	2-Chlorophenol	1.393	1.361	2.3	118	0.00
9	Benzaldehyde	0.934	0.975	-4.4	119	0.00
10 C	Phenol	2.052	2.001	2.5	115	0.00
11	bis(2-Chloroethyl)ether	1.503	1.534	-2.1	119	0.00
12	1,3-Dichlorobenzene	1.488	1.464	1.6	119	0.00
13 C	1,4-Dichlorobenzene	1.523	1.486	2.4	118	0.00
14	1,2-Dichlorobenzene	1.398	1.321	5.5	116	0.00
15	Benzyl Alcohol	1.187	1.218	-2.6	117	0.00
16	2,2'-oxybis(1-Chloropropane	2.584	2.401	7.1	113	0.00
17	2-Methylphenol	1.303	1.312	-0.7	121	0.00
18	Hexachloroethane	0.501	0.503	-0.4	120	0.00
19 P	n-Nitroso-di-n-propylamine	1.177	1.096	6.9	116	0.00
20	3+4-Methylphenols	1.493	1.411	5.5	114	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	121	0.00
22	Acetophenone	0.484	0.462	4.5	116	0.00
23 S	Nitrobenzene-d5	0.430	0.423	1.6	116	0.00
24	Nitrobenzene	0.432	0.428	0.9	118	0.00
25	Isophorone	0.818	0.813	0.6	120	0.00
26 C	2-Nitrophenol	0.172	0.187	-8.7	127	0.00
27	2,4-Dimethylphenol	0.321	0.323	-0.6	120	0.00
28	bis(2-Chloroethoxy)methane	0.480	0.481	-0.2	119	0.00
29 C	2,4-Dichlorophenol	0.294	0.308	-4.8	123	0.00
30	1,2,4-Trichlorobenzene	0.310	0.313	-1.0	122	0.00
31	Naphthalene	1.073	1.037	3.4	119	0.00
32	Benzoic acid	0.167	0.181	-8.4	133	0.00
33	4-Chloroaniline	0.421	0.420	0.2	115	0.00
34 C	Hexachlorobutadiene	0.183	0.184	-0.5	121	0.00
35	Caprolactam	0.100	0.106	-6.0	125	0.00
36 C	4-Chloro-3-methylphenol	0.329	0.336	-2.1	121	0.00
37	2-Methylnaphthalene	0.684	0.653	4.5	115	0.00
38	1-Methylnaphthalene	0.644	0.622	3.4	118	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	123	0.00
40	1,2,4,5-Tetrachlorobenzene	0.558	0.553	0.9	121	0.00
41 P	Hexachlorocyclopentadiene	0.180	0.329	-82.8#	213#	0.00
42 S	2,4,6-Tribromophenol	0.176	0.187	-6.3	132	0.00
43 C	2,4,6-Trichlorophenol	0.364	0.372	-2.2	122	0.00
44	2,4,5-Trichlorophenol	0.419	0.428	-2.1	123	0.00
45 S	2-Fluorobiphenyl	1.278	1.149	10.1	114	0.00
46	1,1'-Biphenyl	1.464	1.391	5.0	117	0.00
47	2-Chloronaphthalene	1.116	1.093	2.1	120	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031124\
 Data File : BF137553.D
 Acq On : 11 Mar 2024 13:10
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 12 02:01:30 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF022924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 12 02:00:03 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.379	0.408	-7.7	125	0.00
49	Acenaphthylene	1.786	1.668	6.6	114	0.00
50	Dimethylphthalate	1.392	1.368	1.7	122	0.00
51	2,6-Dinitrotoluene	0.291	0.303	-4.1	125	0.00
52 C	Acenaphthene	1.066	1.009	5.3	118	0.00
53	3-Nitroaniline	0.308	0.333	-8.1	125	0.00
54 P	2,4-Dinitrophenol	0.111	0.129	-16.2	130	0.00
55	Dibenzofuran	1.624	1.501	7.6	115	0.00
56 P	4-Nitrophenol	0.190	0.212	-11.6	127	0.00
57	2,4-Dinitrotoluene	0.358	0.366	-2.2	120	0.00
58	Fluorene	1.247	1.131	9.3	115	0.00
59	2,3,4,6-Tetrachlorophenol	0.336	0.340	-1.2	123	0.00
60	Diethylphthalate	1.315	1.283	2.4	121	0.00
61	4-Chlorophenyl-phenylether	0.611	0.565	7.5	118	0.00
62	4-Nitroaniline	0.265	0.289	-9.1	131	0.00
63	Azobenzene	1.514	1.442	4.8	117	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	123	0.00
65	4,6-Dinitro-2-methylphenol	0.111	0.122	-9.9	128	0.00
66 c	n-Nitrosodiphenylamine	0.644	0.626	2.8	121	0.00
67	4-Bromophenyl-phenylether	0.218	0.223	-2.3	124	0.00
68	Hexachlorobenzene	0.220	0.229	-4.1	130	0.00
69	Atrazine	0.196	0.200	-2.0	125	0.00
70 C	Pentachlorophenol	0.133	0.145	-9.0	128	0.00
71	Phenanthrene	1.088	1.029	5.4	119	0.00
72	Anthracene	1.118	1.059	5.3	119	0.00
73	Carbazole	0.958	0.935	2.4	121	0.00
74	Di-n-butylphthalate	1.139	1.106	2.9	119	0.00
75 C	Fluoranthene	1.064	1.007	5.4	120	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	126	0.00
77	Benzydine	0.489	0.514	-5.1	114	0.00
78	Pyrene	1.962	1.971	-0.5	120	0.00
79 S	Terphenyl-d14	1.454	1.418	2.5	119	0.00
80	Butylbenzylphthalate	0.501	0.622	-24.2	148	0.00
81	Benzo(a)anthracene	1.402	1.364	2.7	121	0.00
82	3,3'-Dichlorobenzidine	0.373	0.450	-20.6	135	0.00
83	Chrysene	1.417	1.399	1.3	121	0.00
84	Bis(2-ethylhexyl)phthalate	0.637	0.726	-14.0	135	0.00
85 c	Di-n-octyl phthalate	1.120	1.271	-13.5	133	0.00
86 I	Perylene-d12	1.000	1.000	0.0	131	0.00
87	Indeno(1,2,3-cd)pyrene	1.316	1.382	-5.0	131	0.00
88	Benzo(b)fluoranthene	1.183	1.146	3.1	126	0.00
89	Benzo(k)fluoranthene	1.271	1.221	3.9	130	0.00
90 C	Benzo(a)pyrene	1.166	1.164	0.2	127	0.00
91	Dibenzo(a,h)anthracene	1.068	1.115	-4.4	129	0.00
92	Benzo(g,h,i)perylene	1.096	1.147	-4.7	130	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031124\
Data File : BF137553.D
Acq On : 11 Mar 2024 13:10
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Mar 12 02:01:30 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF022924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 12 02:00:03 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0