

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031124\
 Data File : BF137553.D
 Acq On : 11 Mar 2024 13:10
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 12 02:01:30 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF022924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 12 02:00:03 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	122	0.00
2	1,4-Dioxane	40.000	38.129	4.7	114	0.00
3	Pyridine	40.000	41.150	-2.9	115	0.00
4	n-Nitrosodimethylamine	40.000	40.414	-1.0	122	0.00
5 S	2-Fluorophenol	80.000	78.835	1.5	115	0.00
6	Aniline	40.000	40.858	-2.1	118	0.00
7 S	Phenol-d6	80.000	77.018	3.7	117	0.00
8	2-Chlorophenol	40.000	39.095	2.3	118	0.00
9	Benzaldehyde	40.000	41.744	-4.4	119	0.00
10 C	Phenol	40.000	39.011	2.5	115	0.00
11	bis(2-Chloroethyl)ether	40.000	40.803	-2.0	119	0.00
12	1,3-Dichlorobenzene	40.000	39.342	1.6	119	0.00
13 C	1,4-Dichlorobenzene	40.000	39.033	2.4	118	0.00
14	1,2-Dichlorobenzene	40.000	37.813	5.5	116	0.00
15	Benzyl Alcohol	40.000	41.054	-2.6	117	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	37.165	7.1	113	0.00
17	2-Methylphenol	40.000	40.257	-0.6	121	0.00
18	Hexachloroethane	40.000	40.156	-0.4	120	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	37.254	6.9	116	0.00
20	3+4-Methylphenols	40.000	37.806	5.5	114	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	121	0.00
22	Acetophenone	40.000	38.149	4.6	116	0.00
23 S	Nitrobenzene-d5	80.000	78.563	1.8	116	0.00
24	Nitrobenzene	40.000	39.625	0.9	118	0.00
25	Isophorone	40.000	39.766	0.6	120	0.00
26 C	2-Nitrophenol	40.000	43.499	-8.7	127	0.00
27	2,4-Dimethylphenol	40.000	40.266	-0.7	120	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.109	-0.3	119	0.00
29 C	2,4-Dichlorophenol	40.000	41.800	-4.5	123	0.00
30	1,2,4-Trichlorobenzene	40.000	40.405	-1.0	122	0.00
31	Naphthalene	40.000	38.661	3.3	119	0.00
32	Benzoic acid	40.000	40.428	-1.1	133	0.00
33	4-Chloroaniline	40.000	39.972	0.1	115	0.00
34 C	Hexachlorobutadiene	40.000	40.255	-0.6	121	0.00
35	Caprolactam	40.000	42.326	-5.8	125	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.852	-2.1	121	0.00
37	2-Methylnaphthalene	40.000	38.187	4.5	115	0.00
38	1-Methylnaphthalene	40.000	38.599	3.5	118	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	123	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.623	0.9	121	0.00
41 P	Hexachlorocyclopentadiene	40.000	60.058	-50.1#	213	0.00
42 S	2,4,6-Tribromophenol	80.000	85.254	-6.6	132	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.903	-2.3	122	0.00
44	2,4,5-Trichlorophenol	40.000	40.919	-2.3	123	0.00
45 S	2-Fluorobiphenyl	80.000	71.968	10.0	114	0.00
46	1,1'-Biphenyl	40.000	38.010	5.0	117	0.00
47	2-Chloronaphthalene	40.000	39.164	2.1	120	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031124\
 Data File : BF137553.D
 Acq On : 11 Mar 2024 13:10
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 12 02:01:30 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF022924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 12 02:00:03 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	43.026	-7.6	125	0.00
49	Acenaphthylene	40.000	37.354	6.6	114	0.00
50	Dimethylphthalate	40.000	39.291	1.8	122	0.00
51	2,6-Dinitrotoluene	40.000	41.620	-4.0	125	0.00
52 C	Acenaphthene	40.000	37.847	5.4	118	0.00
53	3-Nitroaniline	40.000	43.223	-8.1	125	0.00
54 P	2,4-Dinitrophenol	40.000	40.768	-1.9	130	0.00
55	Dibenzofuran	40.000	36.981	7.5	115	0.00
56 P	4-Nitrophenol	40.000	41.083	-2.7	127	0.00
57	2,4-Dinitrotoluene	40.000	40.952	-2.4	120	0.00
58	Fluorene	40.000	36.262	9.3	115	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.548	-1.4	123	0.00
60	Diethylphthalate	40.000	39.035	2.4	121	0.00
61	4-Chlorophenyl-phenylether	40.000	36.991	7.5	118	0.00
62	4-Nitroaniline	40.000	43.709	-9.3	131	0.00
63	Azobenzene	40.000	38.102	4.7	117	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	123	0.00
65	4,6-Dinitro-2-methylphenol	40.000	43.721	-9.3	128	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.856	2.9	121	0.00
67	4-Bromophenyl-phenylether	40.000	40.956	-2.4	124	0.00
68	Hexachlorobenzene	40.000	41.676	-4.2	130	0.00
69	Atrazine	40.000	40.866	-2.2	125	0.00
70 C	Pentachlorophenol	40.000	43.633	-9.1	128	0.00
71	Phenanthrene	40.000	37.828	5.4	119	0.00
72	Anthracene	40.000	37.876	5.3	119	0.00
73	Carbazole	40.000	39.064	2.3	121	0.00
74	Di-n-butylphthalate	40.000	38.847	2.9	119	0.00
75 C	Fluoranthene	40.000	37.864	5.3	120	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	126	0.00
77	Benzydine	40.000	41.974	-4.9	114	0.00
78	Pyrene	40.000	40.177	-0.4	120	0.00
79 S	Terphenyl-d14	80.000	77.995	2.5	119	0.00
80	Butylbenzylphthalate	40.000	49.654	-24.1	148	0.00
81	Benzo(a)anthracene	40.000	38.919	2.7	121	0.00
82	3,3'-Dichlorobenzidine	40.000	48.172	-20.4	135	0.00
83	Chrysene	40.000	39.510	1.2	121	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	45.585	-14.0	135	0.00
85 c	Di-n-octyl phthalate	40.000	42.192	-5.5	133	0.00
86 I	Perylene-d12	20.000	20.000	0.0	131	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	42.002	-5.0	131	0.00
88	Benzo(b)fluoranthene	40.000	38.731	3.2	126	0.00
89	Benzo(k)fluoranthene	40.000	38.430	3.9	130	0.00
90 C	Benzo(a)pyrene	40.000	39.953	0.1	127	0.00
91	Dibenzo(a,h)anthracene	40.000	41.787	-4.5	129	0.00
92	Benzo(g,h,i)perylene	40.000	41.859	-4.6	130	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031124\
Data File : BF137553.D
Acq On : 11 Mar 2024 13:10
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Mar 12 02:01:30 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF022924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 12 02:00:03 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0