

Data Path : Z:\HPCHEM1\BNA_F\Data\BF031215\
 Data File : BF077726.D
 Acq On : 12 Mar 2015 13:30
 Operator : TP/IZ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 13 01:06:19 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF031115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 12 14:10:27 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
2	1,4-Dioxane	0.520	0.473	9.0	93	0.00
3	Pyridine	1.398	1.175	16.0	89	0.00
4	n-Nitrosodimethylamine	0.609	0.587	3.6	93	0.00
5 S	2-Fluorophenol	1.146	1.116	2.6	104	0.00
6	Aniline	2.025	2.025	0.0	107	0.00
7 S	Phenol-d6	1.416	1.407	0.6	104	0.00
8	2-Chlorophenol	1.364	1.332	2.3	106	0.00
9	Benzaldehyde	0.580	0.599	-3.3	108	0.00
10 C	Phenol	1.673	1.617	3.3	103	0.00
11	bis(2-Chloroethyl)ether	1.253	1.272	-1.5	106	0.00
12	1,3-Dichlorobenzene	1.517	1.496	1.4	106	0.00
13 C	1,4-Dichlorobenzene	1.576	1.511	4.1	105	0.00
14	1,2-Dichlorobenzene	1.445	1.386	4.1	104	0.00
15	Benzyl Alcohol	1.105	1.093	1.1	104	0.00
16	2,2'-oxybis(1-Chloropropane	1.689	1.719	-1.8	108	0.00
17	2-Methylphenol	1.110	1.096	1.3	103	0.00
18	Hexachloroethane	0.517	0.512	1.0	109	0.00
19 P	n-Nitroso-di-n-propylamine	0.979	0.938	4.2	103	0.00
20	3+4-Methylphenols	1.457	1.413	3.0	104	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	103	0.00
22	Acetophenone	0.443	0.442	0.2	104	0.00
23 S	Nitrobenzene-d5	0.305	0.316	-3.6	110	0.00
24	Nitrobenzene	0.329	0.326	0.9	108	0.00
25	Isophorone	0.616	0.629	-2.1	105	0.00
26 C	2-Nitrophenol	0.161	0.167	-3.7	101	0.00
27	2,4-Dimethylphenol	0.293	0.303	-3.4	106	0.00
28	bis(2-Chloroethoxy)methane	0.374	0.382	-2.1	108	0.00
29 C	2,4-Dichlorophenol	0.279	0.278	0.4	102	0.00
30	1,2,4-Trichlorobenzene	0.313	0.303	3.2	101	0.00
31	Naphthalene	1.027	1.019	0.8	104	0.00
32	Benzoic acid	0.141	0.157	-11.3	104	0.00
33	4-Chloroaniline	0.417	0.416	0.2	101	0.00
34 C	Hexachlorobutadiene	0.177	0.180	-1.7	107	0.00
35	Caprolactam	0.082	0.082	0.0	103	0.00
36 C	4-Chloro-3-methylphenol	0.281	0.291	-3.6	109	0.00
37	2-Methylnaphthalene	0.690	0.689	0.1	102	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	104	0.00
39	1,2,4,5-Tetrachlorobenzene	0.597	0.583	2.3	103	0.00
40 P	Hexachlorocyclopentadiene	0.254	0.257	-1.2	97	0.00
41 S	2,4,6-Tribromophenol	0.191	0.172	9.9	92	0.00
42 C	2,4,6-Trichlorophenol	0.413	0.406	1.7	98	0.00
43	2,4,5-Trichlorophenol	0.446	0.443	0.7	104	0.00
44 S	2-Fluorobiphenyl	1.361	1.328	2.4	106	0.00

Data Path : Z:\HPCHEM1\BNA_F\Data\BF031215\
 Data File : BF077726.D
 Acq On : 12 Mar 2015 13:30
 Operator : TP/IZ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 13 01:06:19 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF031115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 12 14:10:27 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.599	1.565	2.1	101	0.00
46	2-Chloronaphthalene	1.299	1.286	1.0	105	0.00
47	2-Nitroaniline	0.323	0.311	3.7	95	0.00
48	Acenaphthylene	2.161	2.009	7.0	99	0.00
49	Dimethylphthalate	1.483	1.371	7.6	98	0.00
50	2,6-Dinitrotoluene	0.307	0.304	1.0	102	0.00
51 C	Acenaphthene	1.239	1.178	4.9	99	0.00
52	3-Nitroaniline	0.357	0.370	-3.6	105	0.00
53 P	2,4-Dinitrophenol	0.093	0.079	15.1	87	0.00
54	Dibenzofuran	1.837	1.787	2.7	102	0.00
55 P	4-Nitrophenol	0.265	0.254	4.2	93	0.00
56	2,4-Dinitrotoluene	0.401	0.411	-2.5	107	0.00
57	Fluorene	1.526	1.440	5.6	98	0.00
58	2,3,4,6-Tetrachlorophenol	0.344	0.318	7.6	95	0.00
59	Diethylphthalate	1.445	1.435	0.7	104	0.00
60	4-Chlorophenyl-phenylether	0.696	0.648	6.9	98	0.00
61	4-Nitroaniline	0.356	0.356	0.0	103	0.00
62	Azobenzene	1.381	1.366	1.1	95	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	103	0.00
64	4,6-Dinitro-2-methylphenol	0.078	0.076	2.6	88	0.00
65 c	n-Nitrosodiphenylamine	0.666	0.656	1.5	99	0.00
66	4-Bromophenyl-phenylether	0.213	0.207	2.8	98	0.00
67	Hexachlorobenzene	0.235	0.227	3.4	99	0.00
68	Atrazine	0.187	0.182	2.7	96	0.00
69 C	Pentachlorophenol	0.104	0.110	-5.8	99	0.00
70	Phenanthrene	1.148	1.109	3.4	99	0.00
71	Anthracene	1.134	1.142	-0.7	107	0.00
72	Carbazole	1.079	1.088	-0.8	108	0.00
73	Di-n-butylphthalate	1.210	1.233	-1.9	105	0.00
74 C	Fluoranthene	1.266	1.178	7.0	90	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	101	0.00
76	Benzidine	0.546	0.545	0.2	108	0.00
77	Pyrene	1.258	1.216	3.3	90	0.00
78 S	Terphenyl-d14	0.819	0.774	5.5	90	0.00
79	Butylbenzylphthalate	0.496	0.498	-0.4	100	0.00
80	Benzo(a)anthracene	1.186	1.179	0.6	104	0.00
81	3,3'-Dichlorobenzidine	0.391	0.373	4.6	101	0.00
82	Chrysene	1.066	1.054	1.1	100	0.00
83	Bis(2-ethylhexyl)phthalate	0.751	0.748	0.4	103	0.00
84 c	Di-n-octyl phthalate	1.284	1.263	1.6	102	0.00
85	Indeno(1,2,3-cd)pyrene	1.331	1.294	2.8	106	0.00
86 I	Perylene-d12	1.000	1.000	0.0	103	0.00
87	Benzo(b)fluoranthene	1.306	1.167	10.6	93	0.00

Data Path : Z:\HPCHEM1\BNA_F\Data\BF031215\
Data File : BF077726.D
Acq On : 12 Mar 2015 13:30
Operator : TP/IZ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Mar 13 01:06:19 2015
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF031115.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 12 14:10:27 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.020	1.107	-8.5	118	0.00
89 C	Benzo(a)pyrene	1.089	1.087	0.2	105	0.00
90	Dibenzo(a,h)anthracene	1.081	1.084	-0.3	106	0.00
91	Benzo(g,h,i)perylene	1.100	1.072	2.5	101	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0