


Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF031219\
 Data File : BF113022.D
 Acq On : 13 Mar 2019 00:48
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF022719.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Mar 13 03:42:55 2019



AutoFind: Scans 1601, 1602, 1603; Background Corrected with Scan 1596

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	35.5	46317	PASS
68	69	0.00	2	1.6	665	PASS
69	198	0.00	100	32.5	42309	PASS
70	69	0.00	2	0.6	236	PASS
127	198	10	80	45.6	59461	PASS
197	198	0.00	2	0.3	364	PASS
198	198	100	100	100.0	130349	PASS
199	198	5	9	6.7	8680	PASS
275	198	10	60	28.2	36789	PASS
365	198	1	100	3.3	4317	PASS
441	198	0.01	100	13.5	17619	PASS
442	198	50	100	83.2	108394	PASS
443	442	15	24	19.6	21251	PASS