


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF031317\
 Data File : BF093602.D
 Acq On : 13 Mar 2017 11:07
 Operator : SJ/MA
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF030717.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Mar 13 15:05:08 2017



AutoFind: Scans 841, 842, 843; Background Corrected with Scan 838

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	47.2	51071	PASS
68	69	0.00	2	1.8	791	PASS
69	198	0.00	100	40.2	43461	PASS
70	69	0.00	2	0.4	184	PASS
127	198	10	80	49.8	53864	PASS
197	198	0.00	2	0.1	130	PASS
198	198	100	100	100.0	108130	PASS
199	198	5	9	7.0	7568	PASS
275	198	10	60	22.3	24136	PASS
365	198	1	100	2.2	2365	PASS
441	442	0.01	24	15.7	14716	PASS
442	198	50	100	86.9	93971	PASS
443	442	15	24	17.8	16765	PASS