

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031422\
 Data File : BF127318.D
 Acq On : 14 Mar 2022 14:01
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 14 15:29:42 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031022.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 14 12:19:36 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	150	0.00
2	1,4-Dioxane	40.000	35.626	10.9	140	0.04
3	Pyridine	40.000	36.459	8.9	135	0.03
4	n-Nitrosodimethylamine	40.000	34.555	13.6	125	0.04
5 S	2-Fluorophenol	80.000	73.996	7.5	144	0.00
6	Aniline	40.000	38.543	3.6	141	0.00
7 S	Phenol-d6	80.000	74.223	7.2	140	0.00
8	2-Chlorophenol	40.000	36.848	7.9	137	0.00
9	Benzaldehyde	40.000	32.015	20.0	131	0.00
10 C	Phenol	40.000	37.342	6.6	138	0.00
11	bis(2-Chloroethyl)ether	40.000	37.836	5.4	137	0.00
12	1,3-Dichlorobenzene	40.000	37.018	7.5	139	0.00
13 C	1,4-Dichlorobenzene	40.000	37.603	6.0	140	0.00
14	1,2-Dichlorobenzene	40.000	36.621	8.4	137	0.00
15	Benzyl Alcohol	40.000	36.009	10.0	129	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	36.969	7.6	134	0.00
17	2-Methylphenol	40.000	37.501	6.2	140	0.00
18	Hexachloroethane	40.000	37.332	6.7	141	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	35.314	11.7	128	0.00
20	3+4-Methylphenols	40.000	34.506	13.7	131	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	150	0.00
22	Acetophenone	40.000	34.717	13.2	130	0.00
23 S	Nitrobenzene-d5	80.000	73.780	7.8	138	0.00
24	Nitrobenzene	40.000	37.591	6.0	139	0.00
25	Isophorone	40.000	37.914	5.2	143	0.00
26 C	2-Nitrophenol	40.000	38.060	4.8	144	0.00
27	2,4-Dimethylphenol	40.000	38.816	3.0	146	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.225	4.4	140	0.00
29 C	2,4-Dichlorophenol	40.000	37.933	5.2	139	0.00
30	1,2,4-Trichlorobenzene	40.000	38.061	4.8	142	0.00
31	Naphthalene	40.000	37.908	5.2	145	0.00
32	Benzoic acid	40.000	40.401	-1.0	148	0.00
33	4-Chloroaniline	40.000	37.425	6.4	142	0.00
34 C	Hexachlorobutadiene	40.000	37.207	7.0	137	0.00
35	Caprolactam	40.000	38.826	2.9	135	0.00
36 C	4-Chloro-3-methylphenol	40.000	38.204	4.5	139	0.00
37	2-Methylnaphthalene	40.000	37.217	7.0	140	0.00
38	1-Methylnaphthalene	40.000	37.041	7.4	136	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	143	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.138	4.7	137	0.00
41 P	Hexachlorocyclopentadiene	40.000	44.011	-10.0	139	0.00
42 S	2,4,6-Tribromophenol	80.000	73.175	8.5	129	0.00
43 C	2,4,6-Trichlorophenol	40.000	38.620	3.5	134	0.00
44	2,4,5-Trichlorophenol	40.000	38.658	3.4	137	0.00
45 S	2-Fluorobiphenyl	80.000	71.632	10.5	132	0.00
46	1,1'-Biphenyl	40.000	37.208	7.0	136	0.00
47	2-Chloronaphthalene	40.000	38.224	4.4	141	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031422\
 Data File : BF127318.D
 Acq On : 14 Mar 2022 14:01
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 14 15:29:42 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031022.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 14 12:19:36 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	38.375	4.1	135	0.00
49	Acenaphthylene	40.000	37.313	6.7	132	0.00
50	Dimethylphthalate	40.000	37.574	6.1	138	0.00
51	2,6-Dinitrotoluene	40.000	37.804	5.5	133	0.00
52 C	Acenaphthene	40.000	37.201	7.0	135	0.00
53	3-Nitroaniline	40.000	37.800	5.5	135	0.00
54 P	2,4-Dinitrophenol	40.000	37.360	6.6	121	0.00
55	Dibenzofuran	40.000	36.304	9.2	131	0.00
56 P	4-Nitrophenol	40.000	39.405	1.5	128	0.00
57	2,4-Dinitrotoluene	40.000	37.340	6.6	129	0.00
58	Fluorene	40.000	36.238	9.4	130	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.089	2.3	137	0.00
60	Diethylphthalate	40.000	36.680	8.3	133	0.00
61	4-Chlorophenyl-phenylether	40.000	36.342	9.1	128	0.00
62	4-Nitroaniline	40.000	37.702	5.7	132	0.00
63	Azobenzene	40.000	37.257	6.9	128	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	139	0.00
65	4,6-Dinitro-2-methylphenol	40.000	41.364	-3.4	129	0.00
66 c	n-Nitrosodiphenylamine	40.000	37.499	6.3	132	0.00
67	4-Bromophenyl-phenylether	40.000	37.939	5.2	128	0.00
68	Hexachlorobenzene	40.000	37.866	5.3	126	0.00
69	Atrazine	40.000	38.925	2.7	128	0.00
70 C	Pentachlorophenol	40.000	41.312	-3.3	130	0.00
71	Phenanthrene	40.000	36.763	8.1	131	0.00
72	Anthracene	40.000	37.019	7.5	131	0.00
73	Carbazole	40.000	36.610	8.5	128	0.00
74	Di-n-butylphthalate	40.000	36.916	7.7	129	0.00
75 C	Fluoranthene	40.000	36.329	9.2	127	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	131	0.00
77	Benzydine	40.000	33.216	17.0	128	0.00
78	Pyrene	40.000	38.789	3.0	129	0.00
79 S	Terphenyl-d14	80.000	76.080	4.9	128	0.00
80	Butylbenzylphthalate	40.000	39.665	0.8	127	0.00
81	Benzo(a)anthracene	40.000	38.091	4.8	124	0.00
82	3,3'-Dichlorobenzidine	40.000	35.997	10.0	122	0.00
83	Chrysene	40.000	38.134	4.7	127	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	38.792	3.0	124	0.00
85 c	Di-n-octyl phthalate	40.000	39.555	1.1	130	0.00
86 I	Perylene-d12	20.000	20.000	0.0	139	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	37.624	5.9	127	0.00
88	Benzo(b)fluoranthene	40.000	37.554	6.1	130	0.00
89	Benzo(k)fluoranthene	40.000	40.303	-0.8	142	0.00
90 C	Benzo(a)pyrene	40.000	38.551	3.6	130	0.00
91	Dibenzo(a,h)anthracene	40.000	37.725	5.7	127	0.00
92	Benzo(g,h,i)perylene	40.000	37.644	5.9	126	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031422\
Data File : BF127318.D
Acq On : 14 Mar 2022 14:01
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Mar 14 15:29:42 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031022.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 14 12:19:36 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0