

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031424\
 Data File : BF137589.D
 Acq On : 14 Mar 2024 14:02
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 14 14:50:21 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF022924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 14 14:45:20 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	89	0.00
2	1,4-Dioxane	40.000	40.583	-1.5	88	0.00
3	Pyridine	40.000	41.448	-3.6	84	0.00
4	n-Nitrosodimethylamine	40.000	39.295	1.8	87	0.00
5 S	2-Fluorophenol	80.000	83.361	-4.2	89	0.00
6	Aniline	40.000	39.699	0.8	84	0.00
7 S	Phenol-d6	80.000	78.375	2.0	87	0.00
8	2-Chlorophenol	40.000	40.886	-2.2	90	0.00
9	Benzaldehyde	40.000	39.749	0.6	83	0.00
10 C	Phenol	40.000	39.534	1.2	85	0.00
11	bis(2-Chloroethyl)ether	40.000	40.343	-0.9	86	0.00
12	1,3-Dichlorobenzene	40.000	39.920	0.2	88	0.00
13 C	1,4-Dichlorobenzene	40.000	39.564	1.1	87	0.00
14	1,2-Dichlorobenzene	40.000	39.497	1.3	88	0.00
15	Benzyl Alcohol	40.000	39.517	1.2	82	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	36.130	9.7	80	0.00
17	2-Methylphenol	40.000	39.412	1.5	86	0.00
18	Hexachloroethane	40.000	41.269	-3.2	90	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	35.534	11.2	80	0.00
20	3+4-Methylphenols	40.000	38.840	2.9	86	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	85	0.00
22	Acetophenone	40.000	39.513	1.2	84	0.00
23 S	Nitrobenzene-d5	80.000	80.519	-0.6	83	0.00
24	Nitrobenzene	40.000	40.372	-0.9	84	0.00
25	Isophorone	40.000	37.573	6.1	79	0.00
26 C	2-Nitrophenol	40.000	40.775	-1.9	83	0.00
27	2,4-Dimethylphenol	40.000	40.287	-0.7	84	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.817	0.5	83	0.00
29 C	2,4-Dichlorophenol	40.000	41.060	-2.7	85	0.00
30	1,2,4-Trichlorobenzene	40.000	40.631	-1.6	86	0.00
31	Naphthalene	40.000	39.691	0.8	85	0.00
32	Benzoic acid	40.000	29.869	25.3#	64	0.00
33	4-Chloroaniline	40.000	40.333	-0.8	81	0.00
34 C	Hexachlorobutadiene	40.000	40.616	-1.5	86	0.00
35	Caprolactam	40.000	39.095	2.3	81	0.00
36 C	4-Chloro-3-methylphenol	40.000	38.347	4.1	80	0.00
37	2-Methylnaphthalene	40.000	39.063	2.3	83	0.00
38	1-Methylnaphthalene	40.000	38.766	3.1	83	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	79	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	42.214	-5.5	83	0.00
41 P	Hexachlorocyclopentadiene	40.000	38.975	2.6	82	0.00
42 S	2,4,6-Tribromophenol	80.000	79.710	0.4	79	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.300	-3.2	80	0.00
44	2,4,5-Trichlorophenol	40.000	40.586	-1.5	79	0.00
45 S	2-Fluorobiphenyl	80.000	78.754	1.6	81	0.00
46	1,1'-Biphenyl	40.000	40.590	-1.5	81	0.00
47	2-Chloronaphthalene	40.000	41.064	-2.7	81	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031424\
 Data File : BF137589.D
 Acq On : 14 Mar 2024 14:02
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 14 14:50:21 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF022924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 14 14:45:20 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.378	-3.4	77	0.00
49	Acenaphthylene	40.000	40.243	-0.6	79	0.00
50	Dimethylphthalate	40.000	39.053	2.4	78	0.00
51	2,6-Dinitrotoluene	40.000	40.795	-2.0	79	0.00
52 C	Acenaphthene	40.000	39.361	1.6	79	0.00
53	3-Nitroaniline	40.000	41.523	-3.8	77	0.00
54 P	2,4-Dinitrophenol	40.000	31.695	20.8	60	0.00
55	Dibenzofuran	40.000	39.182	2.0	78	0.00
56 P	4-Nitrophenol	40.000	35.744	10.6	70	0.00
57	2,4-Dinitrotoluene	40.000	41.306	-3.3	78	0.00
58	Fluorene	40.000	38.067	4.8	78	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	38.961	2.6	76	0.00
60	Diethylphthalate	40.000	37.828	5.4	75	0.00
61	4-Chlorophenyl-phenylether	40.000	37.856	5.4	78	0.00
62	4-Nitroaniline	40.000	37.387	6.5	72	0.00
63	Azobenzene	40.000	36.846	7.9	73	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	74	0.00
65	4,6-Dinitro-2-methylphenol	40.000	41.127	-2.8	73	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.954	-2.4	77	0.00
67	4-Bromophenyl-phenylether	40.000	41.896	-4.7	77	0.00
68	Hexachlorobenzene	40.000	40.996	-2.5	77	0.00
69	Atrazine	40.000	40.657	-1.6	75	0.00
70 C	Pentachlorophenol	40.000	39.477	1.3	70	0.00
71	Phenanthrene	40.000	39.907	0.2	76	0.00
72	Anthracene	40.000	40.393	-1.0	77	0.00
73	Carbazole	40.000	40.571	-1.4	76	0.00
74	Di-n-butylphthalate	40.000	39.308	1.7	73	0.00
75 C	Fluoranthene	40.000	39.554	1.1	76	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	93	0.00
77	Benzydine	40.000	19.266	51.8#	39	0.00
78	Pyrene	40.000	34.894	12.8	77	0.00
79 S	Terphenyl-d14	80.000	68.003	15.0	77	0.00
80	Butylbenzylphthalate	40.000	44.261	-10.7	98	0.00
81	Benzo(a)anthracene	40.000	39.050	2.4	90	0.00
82	3,3'-Dichlorobenzidine	40.000	46.486	-16.2	97	0.00
83	Chrysene	40.000	39.434	1.4	90	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	47.179	-17.9	103	0.00
85 c	Di-n-octyl phthalate	40.000	42.941	-7.4	101	0.00
86 I	Perylene-d12	20.000	20.000	0.0	98	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	36.970	7.6	86	0.00
88	Benzo(b)fluoranthene	40.000	39.697	0.8	96	0.00
89	Benzo(k)fluoranthene	40.000	39.010	2.5	99	0.00
90 C	Benzo(a)pyrene	40.000	40.095	-0.2	95	0.00
91	Dibenzo(a,h)anthracene	40.000	36.756	8.1	85	0.00
92	Benzo(g,h,i)perylene	40.000	37.113	7.2	86	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031424\
Data File : BF137589.D
Acq On : 14 Mar 2024 14:02
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Mar 14 14:50:21 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF022924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 14 14:45:20 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0