




Quant Time: Mar 14 11:49:11 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:46:22 2025
Response via : Initial Calibration



TIC: BF141954.D\data.ms

(70) **Pentachlorophenol (C)**




11.198min (-0.006) 97299.40 ng

response **229558**

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.80	63.38
264.00	62.70	63.36
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031425\
Data File : BF141954.D
Acq On : 14 Mar 2025 09:08
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Mar 14 11:49:11 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:46:22 2025
Response via : Initial Calibration



TIC: BF141954.D\data.ms

(77) Benzidine

12.733min (+ 0.000) 228340.27 ng

response 946176

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.40	14.40
183.00	12.10	11.99
0.00	0.00	0.00